

LAB MONTHLY SUMMARY

Lab Name **SARDANA LABS**Lab No **2314**Month **March**Year **2021**Constituent Group **Chemistry II**Date of Result Entered : **20/03/2021**Date of Report Published : **01/04/2021**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Any Analyser (Automation / Semi Automation)	1002	157.22	10.81	17.00	176 mg/dl	1.10	1.07
2	UREA II	Urease UV / GLDH II	Any Analyser (Automation / Semi Automation)	580	42.43	12.42	5.27	53 mg/dl	2.01	0.44
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Any Analyser (Automation / Semi Automation)	729	1.64	13.37	0.22	1.9 mg/dl	1.18	0.02
4	T-PROTEIN II	Biuret - Colorimetric II	Any Analyser (Automation / Semi Automation)	711	5.71	8.56	0.49	5.4 g/dl	-0.63	0.04
5	URIC ACID II	Enzymatic / Uricase Colorimetric II	Any Analyser (Automation / Semi Automation)	799	7.29	15.23	1.11	6.5 mg/dl	-0.71	0.08
6	CHOLESTROL II	CHOD-PAP II	Any Analyser (Automation / Semi Automation)	946	123.69	12.29	15.20	124 mg/dl	0.02	0.99
7	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Any Analyser (Automation / Semi Automation)	788	119.29	14.60	17.42	137 mg/dl	1.02	1.24

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ± 1.0 to ± 2.0	Good.
Between ± 2.0 to ± 3.0	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Dr. Pamela Christudoss
CMC EQAS Co-Ordinator
Christian Medical College, Vellore

***** End of Report *****