

Qc results

From: 01/01/2021 To: 12/03/2021 Concentration: Normal

No.	Test name	QC lot	Result	Run date	QC range
1	CAL	26471	9.27	01/05/2021::17:19:59	8.25-10.01
2	ALP	26471	102.31	01/05/2021::17:20:10	96.26-111.74
3	SGPT	26471	31.97	01/05/2021::17:20:31	31.10-35.86
4	ALB	26471	4.12	01/05/2021::17:21:34	3.75-4.35
5	GGT	26471	62.27	01/05/2021::17:22:10	59.00-64.60
6	CREAT	26471	1.98	01/05/2021::17:22:24	1.60-2.48
7	DBIL	26471	0.46	01/05/2021::17:22:24	0.33-0.49
8	UREA	26471	36.35	01/05/2021::17:23:42	29.38-36.30
9	SGOT	26471	41.91	01/05/2021::17:24:23	38.48-43.96
10	CAL	26471	9.28	01/05/2021::17:48:30	8.25-10.01
11	ALP	26471	100.68	01/05/2021::17:48:40	96.26-111.74
12	SGPT	26471	33.64	01/05/2021::17:49:02	31.10-35.86
13	ALB	26471	4.44	01/05/2021::17:50:16	3.75-4.35
14	CREAT	26471	1.96	01/05/2021::17:51:06	1.60-2.48
15	GGT	26471	61.34	01/05/2021::17:51:06	59.00-64.60
16	DBIL	26471	0.43	01/05/2021::17:51:19	0.33-0.49
17	UREA	26471	34.05	01/05/2021::17:52:24	29.38-36.30
18	SGOT	26471	41.85	01/05/2021::17:53:05	38.48-43.96
19	UA	26471	4.37	01/05/2021::17:54:21	3.82-5.06
20	TBIL	26471	1.16	01/05/2021::17:55:12	0.78-1.46
21	TP	26471	6.46	01/05/2021::17:55:28	5.93-6.89
22	GLUC	26471	89.12	01/05/2021::17:55:41	81.69-90.21
23	CHOL	26471	260.82	01/05/2021::17:58:17	247.81-266.89
24	HDLC	26471	66.12	01/05/2021::18:00:28	45.37-71.97
25	LDLC	26471	134.87	01/05/2021::18:00:28	125.50-144.62
26	TG	26471	199.53	01/05/2021::18:20:52	192.95-201.43
27	CAL	26471	9.41	01/07/2021::11:18:38	8.25-10.01
28	ALP	26471	103.01	01/07/2021::11:19:22	96.26-111.74
29	SGPT	26471	35.32	01/07/2021::11:19:38	31.10-35.86
30	ALB	26471	4.25	01/07/2021::11:19:53	3.75-4.35
31	DBIL	26471	0.43	01/07/2021::11:20:26	0.33-0.49
32	GGT	26471	60.17	01/07/2021::11:20:47	59.00-64.60
33	CREAT	26471	1.77	01/07/2021::11:21:27	1.60-2.48
34	UREA	26471	33.36	01/07/2021::11:21:27	29.38-36.30
35	SGOT	26471	41.91	01/07/2021::11:22:19	38.48-43.96
36	UA	26471	4.37	01/07/2021::11:24:25	3.82-5.06
37	TBIL	26471	1.16	01/07/2021::11:26:08	0.78-1.46
38	TP	26471	6.55	01/07/2021::11:26:08	5.93-6.89
39	GLUC	26471	87.92	01/07/2021::11:26:21	81.69-90.21
40	CHOL	26471	260.82	01/07/2021::11:27:15	247.81-266.89
41	TG	26471	199.53	01/07/2021::11:28:59	192.95-201.43
42	HDLC	26471	66.12	01/07/2021::11:30:04	45.37-71.97
43	LDLC	26471	147.03	01/07/2021::12:53:05	125.50-144.62
44	DBIL	26471	0.43	01/07/2021::15:20:50	0.33-0.49
45	CAL	26471	9.45	01/09/2021::10:56:44	8.25-10.01
46	ALP	26471	105.14	01/09/2021::10:57:25	96.26-111.74
47	SGPT	26471	34.73	01/09/2021::10:57:38	31.10-35.86
48	ALB	26471	4.29	01/09/2021::10:58:01	3.75-4.35
49	DBIL	26471	0.41	01/09/2021::10:58:23	0.33-0.49
50	GGT	26471	60.79	01/09/2021::10:58:36	59.00-64.60
51	UREA	26471	33.43	01/09/2021::10:59:42	29.38-36.30
52	SGOT	26471	41.11	01/09/2021::11:00:21	38.48-43.96
53	UA	26471	4.44	01/09/2021::11:02:38	3.82-5.06
54	TBIL	26471	1.18	01/09/2021::11:03:57	0.78-1.46

Qc results

From: 01/01/2021 To: 12/03/2021 Concentration: Normal

No.	Test name	QC lot	Result	Run date	QC range
55	TP	26471	6.64	01/09/2021::11:04:10	5.93-6.89
56	GLUC	26471	88.36	01/09/2021::11:04:23	81.69-90.21
57	CHOL	26471	259.31	01/09/2021::11:05:17	247.81-266.89
58	TG	26471	198.09	01/09/2021::11:07:02	192.95-201.43
59	LDLC	26471	141.04	01/09/2021::11:07:54	125.50-144.62
60	HDLC	26471	63.96	01/09/2021::11:08:07	45.37-71.97
61	ALP	26471	99.81	01/12/2021::10:53:11	96.26-111.74
62	SGPT	26471	32.37	01/12/2021::10:53:24	31.10-35.86
63	ALB	26471	4.12	01/12/2021::10:53:47	3.75-4.35
64	DBIL	26471	0.43	01/12/2021::10:54:10	0.33-0.49
65	GGT	26471	62.03	01/12/2021::10:54:23	59.00-64.60
66	CREAT	26471	2.47	01/12/2021::10:55:15	1.60-2.48
67	UREA	26471	30.67	01/12/2021::10:55:15	29.38-36.30
68	SGOT	26471	39.52	01/12/2021::10:56:20	38.48-43.96
69	UA	26471	4.28	01/12/2021::10:58:25	3.82-5.06
70	TBIL	26471	1.17	01/12/2021::10:59:44	0.78-1.46
71	TP	26471	6.29	01/12/2021::10:59:57	5.93-6.89
72	GLUC	26471	84.73	01/12/2021::11:00:10	81.69-90.21
73	CHOL	26471	265.50	01/12/2021::11:01:04	247.81-266.89
74	TG	26471	195.07	01/12/2021::11:02:48	192.95-201.43
75	HDLC	26471	60.04	01/12/2021::11:03:40	45.37-71.97
76	LDLC	26471	133.53	01/12/2021::11:03:41	125.50-144.62
77	CAL	26471	8.53	01/14/2021::11:27:21	8.25-10.01
78	ALP	26471	100.87	01/14/2021::11:28:04	96.26-111.74
79	SGPT	26471	32.08	01/14/2021::11:28:17	31.10-35.86
80	ALB	26471	3.91	01/14/2021::11:28:40	3.75-4.35
81	DBIL	26471	0.44	01/14/2021::11:29:02	0.33-0.49
82	GGT	26471	62.34	01/14/2021::11:29:15	59.00-64.60
83	CREAT	26471	2.07	01/14/2021::11:30:08	1.60-2.48
84	UREA	26471	34.48	01/14/2021::11:30:08	29.38-36.30
85	SGOT	26471	41.39	01/14/2021::11:31:13	38.48-43.96
86	UA	26471	4.36	01/14/2021::11:33:18	3.82-5.06
87	TBIL	26471	1.16	01/14/2021::11:34:37	0.78-1.46
88	TP	26471	6.29	01/14/2021::11:34:50	5.93-6.89
89	GLUC	26471	85.53	01/14/2021::11:35:03	81.69-90.21
90	CHOL	26471	258.31	01/14/2021::11:35:57	247.81-266.89
91	TG	26471	196.64	01/14/2021::11:37:42	192.95-201.43
92	HDLC	26471	68.23	01/14/2021::11:38:33	45.37-71.97
93	LDLC	26471	137.40	01/14/2021::11:38:34	125.50-144.62
94	ALP	26471	105.14	01/16/2021::11:16:59	96.26-111.74
95	ALB	26471	3.92	01/16/2021::11:17:25	3.75-4.35
96	SGPT	26471	33.55	01/16/2021::11:17:25	31.10-35.86
97	GGT	26471	62.66	01/16/2021::11:18:20	59.00-64.60
98	DBIL	26471	0.38	01/16/2021::11:18:57	0.33-0.49
99	CREAT	26471	1.94	01/16/2021::11:19:10	1.60-2.48
100	UREA	26471	34.05	01/16/2021::11:20:12	29.38-36.30
101	SGOT	26471	42.78	01/16/2021::11:21:04	38.48-43.96
102	CAL	26471	8.62	01/16/2021::11:21:56	8.25-10.01
103	UA	26471	4.32	01/16/2021::11:23:10	3.82-5.06
104	TBIL	26471	1.19	01/16/2021::11:23:36	0.78-1.46
105	TP	26471	6.42	01/16/2021::11:23:49	5.93-6.89
106	GLUC	26471	84.58	01/16/2021::11:24:02	81.69-90.21
107	CHOL	26471	248.76	01/16/2021::11:25:10	247.81-266.89
108	TG	26471	197.60	01/16/2021::11:27:33	192.95-201.43

Qc results

From: 01/01/2021

To: 12/03/2021

Concentration: Normal

No.	Test name	QC lot	Result	Run date	QC range
109	LDLC	26471	136.12	01/16/2021::11:28:12	125.50-144.62
110	HDLC	26471	54.00	01/16/2021::11:28:25	45.37-71.97
111	CAL	26471	9.42	01/23/2021::11:07:27	8.25-10.01
112	ALP	26471	104.43	01/23/2021::11:08:09	96.26-111.74
113	SGPT	26471	34.43	01/23/2021::11:08:22	31.10-35.86
114	ALB	26471	3.97	01/23/2021::11:08:45	3.75-4.35
115	DBIL	26471	0.40	01/23/2021::11:09:07	0.33-0.49
116	GGT	26471	63.59	01/23/2021::11:09:21	59.00-64.60
117	CREAT	26471	2.30	01/23/2021::11:10:13	1.60-2.48
118	UREA	26471	30.52	01/23/2021::11:10:13	29.38-36.30
119	SGOT	26471	40.00	01/23/2021::11:11:05	38.48-43.96
120	UA	26471	4.41	01/23/2021::11:13:23	3.82-5.06
121	TBIL	26471	1.17	01/23/2021::11:14:42	0.78-1.46
122	GLUC	26471	86.87	01/23/2021::11:15:08	81.69-90.21
123	CHOL	26471	259.16	01/23/2021::11:16:02	247.81-266.89
124	TG	26471	194.48	01/23/2021::11:17:46	192.95-201.43
125	HDLC	26471	52.04	01/23/2021::11:18:38	45.37-71.97
126	LDLC	26471	133.38	01/23/2021::11:18:39	125.50-144.62
127	CAL	26471	9.58	01/27/2021::10:51:05	8.25-10.01
128	ALP	26471	104.43	01/27/2021::10:51:48	96.26-111.74
129	SGPT	26471	32.08	01/27/2021::10:52:01	31.10-35.86
130	ALB	26471	3.97	01/27/2021::10:52:24	3.75-4.35
131	DBIL	26471	0.37	01/27/2021::10:52:46	0.33-0.49
132	GGT	26471	59.55	01/27/2021::10:52:59	59.00-64.60
133	CREAT	26471	1.97	01/27/2021::10:53:52	1.60-2.48
134	UREA	26471	32.05	01/27/2021::10:53:52	29.38-36.30
135	SGOT	26471	41.11	01/27/2021::10:54:44	38.48-43.96
136	UA	26471	4.27	01/27/2021::10:57:03	3.82-5.06
137	TBIL	26471	1.13	01/27/2021::10:58:21	0.78-1.46
138	GLUC	26471	86.48	01/27/2021::10:58:47	81.69-90.21
139	CHOL	26471	256.10	01/27/2021::10:59:41	247.81-266.89
140	TG	26471	198.57	01/27/2021::11:01:26	192.95-201.43
141	HDLC	26471	54.12	01/27/2021::11:02:18	45.37-71.97
142	LDLC	26471	131.36	01/27/2021::11:02:18	125.50-144.62
143	CAL	26471	9.23	01/29/2021::10:48:33	8.25-10.01
144	ALP	26471	113.66	01/29/2021::10:49:16	96.26-111.74
145	SGPT	26471	34.43	01/29/2021::10:49:29	31.10-35.86
146	ALB	26471	3.90	01/29/2021::10:49:51	3.75-4.35
147	DBIL	26471	0.49	01/29/2021::10:50:14	0.33-0.49
148	GGT	26471	63.90	01/29/2021::10:50:27	59.00-64.60
149	CREAT	26471	2.24	01/29/2021::10:51:19	1.60-2.48
150	UREA	26471	31.62	01/29/2021::10:51:20	29.38-36.30
151	SGOT	26471	43.79	01/29/2021::10:52:25	38.48-43.96
152	UA	26471	4.34	01/29/2021::10:54:30	3.82-5.06
153	TBIL	26471	1.19	01/29/2021::10:55:48	0.78-1.46
154	TP	26471	6.22	01/29/2021::10:56:01	5.93-6.89
155	GLUC	26471	85.92	01/29/2021::10:56:15	81.69-90.21
156	CHOL	26471	257.45	01/29/2021::10:57:09	247.81-266.89
157	TG	26471	197.30	01/29/2021::10:58:53	192.95-201.43
158	HDLC	26471	56.65	01/29/2021::10:59:45	45.37-71.97
159	LDLC	26471	133.79	01/29/2021::10:59:45	125.50-144.62
160	CAL	26471	9.36	02/01/2021::11:21:30	8.25-10.01
161	ALP	26471	107.98	02/01/2021::11:22:12	96.26-111.74
162	SGPT	26471	34.14	02/01/2021::11:22:25	31.10-35.86

Qc results

From: 01/01/2021 To: 12/03/2021 Concentration: Normal

No.	Test name	QC lot	Result	Run date	QC range
163	ALB	26471	3.97	02/01/2021::11:22:48	3.75-4.35
164	DBIL	26471	0.41	02/01/2021::11:23:11	0.33-0.49
165	GGT	26471	64.21	02/01/2021::11:23:24	59.00-64.60
166	CREAT	26471	1.98	02/01/2021::11:24:17	1.60-2.48
167	UREA	26471	34.40	02/01/2021::11:24:17	29.38-36.30
168	SGOT	26471	44.35	02/01/2021::11:25:09	38.48-43.96
169	UA	26471	4.42	02/01/2021::11:27:27	3.82-5.06
170	TBIL	26471	1.19	02/01/2021::11:28:46	0.78-1.46
171	TP	26471	6.26	02/01/2021::11:28:59	5.93-6.89
172	GLUC	26471	86.84	02/01/2021::11:29:12	81.69-90.21
173	CHOL	26471	257.60	02/01/2021::11:30:06	247.81-266.89
174	TG	26471	192.66	02/01/2021::11:31:50	192.95-201.43
175	HDLC	26471	58.00	02/01/2021::11:32:42	45.37-71.97
176	LDLC	26471	129.31	02/01/2021::11:32:42	125.50-144.62
177	CAL	26471	9.54	02/03/2021::10:50:42	8.25-10.01
178	ALP	26471	107.27	02/03/2021::10:51:24	96.26-111.74
179	SGPT	26471	33.55	02/03/2021::10:51:37	31.10-35.86
180	ALB	26471	3.96	02/03/2021::10:52:00	3.75-4.35
181	DBIL	26471	0.44	02/03/2021::10:52:22	0.33-0.49
182	GGT	26471	62.97	02/03/2021::10:52:35	59.00-64.60
183	CREAT	26471	1.85	02/03/2021::10:53:28	1.60-2.48
184	UREA	26471	33.12	02/03/2021::10:53:41	29.38-36.30
185	SGOT	26471	44.07	02/03/2021::10:54:20	38.48-43.96
186	UA	26471	4.35	02/03/2021::10:56:38	3.82-5.06
187	TBIL	26471	1.19	02/03/2021::10:57:57	0.78-1.46
188	TP	26471	6.36	02/03/2021::10:58:10	5.93-6.89
189	GLUC	26471	86.04	02/03/2021::10:58:23	81.69-90.21
190	CHOL	26471	255.54	02/03/2021::10:59:17	247.81-266.89
191	TG	26471	199.11	02/03/2021::11:01:01	192.95-201.43
192	HDLC	26471	63.39	02/03/2021::11:01:54	45.37-71.97
193	LDLC	26471	132.88	02/03/2021::11:01:54	125.50-144.62
194	CAL	26471	8.88	02/05/2021::11:18:58	8.25-10.01
195	ALP	26471	99.81	02/05/2021::11:19:40	96.26-111.74
196	SGPT	26471	34.14	02/05/2021::11:19:53	31.10-35.86
197	ALB	26471	3.93	02/05/2021::11:20:16	3.75-4.35
198	DBIL	26471	0.46	02/05/2021::11:20:39	0.33-0.49
199	GGT	26471	63.59	02/05/2021::11:20:52	59.00-64.60
200	CREAT	26471	1.93	02/05/2021::11:21:44	1.60-2.48
201	UREA	26471	33.05	02/05/2021::11:21:57	29.38-36.30
202	SGOT	26471	42.66	02/05/2021::11:22:36	38.48-43.96
203	UA	26471	4.45	02/05/2021::11:24:55	3.82-5.06
204	TBIL	26471	1.20	02/05/2021::11:26:01	0.78-1.46
205	TP	26471	6.54	02/05/2021::11:26:27	5.93-6.89
206	GLUC	26471	85.30	02/05/2021::11:26:40	81.69-90.21
207	CHOL	26471	257.23	02/05/2021::11:27:34	247.81-266.89
208	TG	26471	195.68	02/05/2021::11:29:19	192.95-201.43
209	HDLC	26471	56.73	02/05/2021::11:30:11	45.37-71.97
210	LDLC	26471	132.00	02/05/2021::11:30:11	125.50-144.62
211	CAL	26471	9.09	02/08/2021::11:10:04	8.25-10.01
212	ALP	26471	96.97	02/08/2021::11:10:42	96.26-111.74
213	SGPT	26471	32.08	02/08/2021::11:11:01	31.10-35.86
214	GGT	26471	62.34	02/08/2021::11:12:05	59.00-64.60
215	DBIL	26471	0.42	02/08/2021::11:13:19	0.33-0.49
216	UREA	26471	31.75	02/08/2021::11:14:12	29.38-36.30

Qc results

From: 01/01/2021

To: 12/03/2021

Concentration: Normal

No.	Test name	QC lot	Result	Run date	QC range
217	SGOT	26471	41.53	02/08/2021::11:15:04	
218	UA	26471	4.57	02/08/2021::11:15:58	38.48-43.96
219	TP	26471	6.23	02/08/2021::11:17:38	3.82-5.06
220	GLUC	26471	83.65	02/08/2021::11:17:51	5.93-6.89
221	TBIL	26471	1.16	02/08/2021::11:18:04	81.69-90.21
222	CHOL	26471	258.47	02/08/2021::11:18:43	0.78-1.46
223	UREA	26471	30.75	02/08/2021::11:20:14	247.81-266.89
224	TG	26471	193.08	02/08/2021::11:20:27	29.38-36.30
225	LDLC	26471	142.43	02/08/2021::11:22:24	192.95-201.43
226	HDLC	26471	57.08	02/08/2021::11:22:37	125.50-144.62
227	ALB	26471	3.77	02/08/2021::11:36:43	45.37-71.97
228	UREA	26471	30.56	02/08/2021::11:37:09	3.75-4.35
229	CREAT	26471	2.78	02/08/2021::11:39:19	29.38-36.30
230	CAL	26471	9.05	02/10/2021::11:16:00	1.60-2.48
231	ALP	26471	98.39	02/10/2021::11:16:43	8.25-10.01
232	SGPT	26471	32.96	02/10/2021::11:16:56	96.26-111.74
233	ALB	26471	3.90	02/10/2021::11:17:19	31.10-35.86
234	DBIL	26471	0.48	02/10/2021::11:17:41	3.75-4.35
235	GGT	26471	62.34	02/10/2021::11:17:54	0.33-0.49
236	CREAT	26471	2.06	02/10/2021::11:18:47	59.00-64.60
237	UREA	26471	31.31	02/10/2021::11:18:47	1.60-2.48
238	SGOT	26471	40.96	02/10/2021::11:19:39	29.38-36.30
239	UA	26471	4.62	02/10/2021::11:21:57	38.48-43.96
240	TBIL	26471	1.18	02/10/2021::11:23:16	3.82-5.06
241	TP	26471	6.51	02/10/2021::11:23:29	0.78-1.46
242	GLUC	26471	82.57	02/10/2021::11:23:42	5.93-6.89
243	CHOL	26471	259.2	02/10/2021::11:24:36	81.69-90.21
244	TG	26471	197.79	02/10/2021::11:26:20	247.81-266.89
245	HDLC	26471	57.23	02/10/2021::11:27:13	192.95-201.43
246	LDLC	26471	140.04	02/10/2021::11:27:13	45.37-71.97
247	CAL	26471	8.94	02/12/2021::11:05:54	125.50-144.62
248	ALP	26471	103.01	02/12/2021::11:06:37	8.25-10.01
249	SGPT	26471	34.43	02/12/2021::11:06:50	96.26-111.74
250	ALB	26471	3.88	02/12/2021::11:07:13	31.10-35.86
251	DBIL	26471	0.48	02/12/2021::11:07:35	3.75-4.35
252	GGT	26471	65.14	02/12/2021::11:07:48	0.33-0.49
253	CREAT	26471	2.12	02/12/2021::11:08:41	59.00-64.60
254	UREA	26471	32.15	02/12/2021::11:08:54	1.60-2.48
255	SGOT	26471	40.96	02/12/2021::11:09:33	29.38-36.30
256	UA	26471	4.35	02/12/2021::11:11:51	38.48-43.96
257	TBIL	26471	1.22	02/12/2021::11:13:10	3.82-5.06
258	TP	26471	6.69	02/12/2021::11:13:23	0.78-1.46
259	GLUC	26471	86.12	02/12/2021::11:13:36	5.93-6.89
260	CHOL	26471	253.03	02/12/2021::11:14:30	81.69-90.21
261	TG	26471	199.96	02/12/2021::11:16:15	247.81-266.89
262	HDLC	26471	57.54	02/12/2021::11:17:07	192.95-201.43
263	CAL	26471	8.67	02/15/2021::10:53:54	45.37-71.97
264	ALP	26471	109.04	02/15/2021::10:54:36	8.25-10.01
265	SGPT	26471	34.43	02/15/2021::10:54:49	96.26-111.74
266	ALB	26471	3.94	02/15/2021::10:55:12	31.10-35.86
267	DBIL	26471	0.39	02/15/2021::10:55:34	3.75-4.35
268	GGT	26471	63.28	02/15/2021::10:55:47	0.33-0.49
269	CREAT	26471	2.21	02/15/2021::10:56:40	59.00-64.60
270	UREA	26471	31.95	02/15/2021::10:56:53	1.60-2.48
					29.38-36.30

Qc results

From: 01/01/2021

To: 12/03/2021

Concentration: Normal

No.	Test name	QC lot	Result	Run date	QC range
271	SGOT	26471	42.94	02/15/2021::10:57:32	38.48-43.96
272	UA	26471	4.29	02/15/2021::10:59:50	3.82-5.06
273	TBIL	26471	1.20	02/15/2021::11:01:08	0.78-1.46
274	TP	26471	6.79	02/15/2021::11:01:22	5.93-6.89
275	GLUC	26471	84.99	02/15/2021::11:01:35	81.69-90.21
276	CHOL	26471	250.82	02/15/2021::11:02:29	247.81-266.89
277	TG	26471	194.52	02/15/2021::11:04:13	192.95-201.43
278	HDLC	26471	56.85	02/15/2021::11:05:05	45.37-71.97
279	CAL	26471	8.94	02/17/2021::13:00:50	8.25-10.01
280	ALP	26471	112.60	02/17/2021::13:01:32	96.26-111.74
281	SGPT	26471	35.53	02/17/2021::13:01:45	31.10-35.86
282	ALB	26471	3.86	02/17/2021::13:02:08	3.75-4.35
283	DBIL	26471	0.45	02/17/2021::13:02:31	0.33-0.49
284	GGT	26471	63.59	02/17/2021::13:02:44	59.00-64.60
285	CREAT	26471	2.15	02/17/2021::13:03:36	1.60-2.48
286	UREA	26471	32.74	02/17/2021::13:03:49	29.38-36.30
287	SGOT	26471	46.33	02/17/2021::13:04:28	38.48-43.96
288	UA	26471	4.50	02/17/2021::13:06:46	3.82-5.06
289	TBIL	26471	1.19	02/17/2021::13:08:05	0.78-1.46
290	TP	26471	6.69	02/17/2021::13:08:19	5.93-6.89
291	GLUC	26471	85.32	02/17/2021::13:08:31	81.69-90.21
292	CHOL	26471	251.68	02/17/2021::13:09:25	247.81-266.89
293	TG	26471	197.23	02/17/2021::13:11:24	192.95-201.43
294	HDLC	26471	49.77	02/17/2021::13:12:03	45.37-71.97
295	CAL	26471	8.91	02/19/2021::11:05:34	8.25-10.01
296	ALP	26471	106.91	02/19/2021::11:06:16	96.26-111.74
297	SGPT	26471	33.38	02/19/2021::11:06:29	31.10-35.86
298	ALB	26471	3.87	02/19/2021::11:06:52	3.75-4.35
299	DBIL	26471	0.43	02/19/2021::11:07:15	0.33-0.49
300	GGT	26471	60.79	02/19/2021::11:07:28	59.00-64.60
301	CREAT	26471	1.77	02/19/2021::11:08:20	1.60-2.48
302	UREA	26471	30.21	02/19/2021::11:08:33	29.38-36.30
303	SGOT	26471	44.35	02/19/2021::11:09:12	38.48-43.96
304	UA	26471	4.48	02/19/2021::11:11:30	3.82-5.06
305	TBIL	26471	1.15	02/19/2021::11:12:49	0.78-1.46
306	TP	26471	6.45	02/19/2021::11:13:02	5.93-6.89
307	GLUC	26471	85.90	02/19/2021::11:13:15	81.69-90.21
308	CHOL	26471	257.21	02/19/2021::11:14:09	247.81-266.89
309	TG	26471	198.62	02/19/2021::11:15:53	192.95-201.43
310	HDLC	26471	53.31	02/19/2021::11:16:46	45.37-71.97
311	CAL	26471	8.87	02/22/2021::10:35:05	8.25-10.01
312	ALP	26471	101.23	02/22/2021::10:35:47	96.26-111.74
313	SGPT	26471	34.61	02/22/2021::10:36:00	31.10-35.86
314	ALB	26471	3.89	02/22/2021::10:36:23	3.75-4.35
315	GGT	26471	62.66	02/22/2021::10:36:59	59.00-64.60
316	DBIL	26471	0.41	02/22/2021::10:37:12	0.33-0.49
317	CREAT	26471	2.16	02/22/2021::10:37:51	1.60-2.48
318	UREA	26471	32.89	02/22/2021::10:38:04	29.38-36.30
319	SGOT	26471	42.66	02/22/2021::10:38:44	38.48-43.96
320	UA	26471	4.29	02/22/2021::10:41:01	3.82-5.06
321	TBIL	26471	1.15	02/22/2021::10:42:20	0.78-1.46
322	TP	26471	6.29	02/22/2021::10:42:33	5.93-6.89
323	GLUC	26471	83.78	02/22/2021::10:42:47	81.69-90.21
324	CHOL	26471	258.41	02/22/2021::10:43:40	247.81-266.89

Qc results

From: 01/01/2021 To: 12/03/2021 Concentration: Normal

No.	Test name	QC lot	Result	Run date	QC range
325	TG	26471	199.82	02/22/2021::10:45:12	192.95-201.43
326	HDLC	26471	48.62	02/22/2021::10:46:17	45.37-71.97
327	LDLC	26471	139.32	02/22/2021::10:46:17	125.50-144.62
328	CAL	26471	8.94	02/24/2021::11:16:41	8.25-10.01
329	ALP	26471	97.68	02/24/2021::11:17:24	96.26-111.74
330	SGPT	26471	34.91	02/24/2021::11:17:37	31.10-35.86
331	ALB	26471	3.75	02/24/2021::11:18:00	3.75-4.35
332	DBIL	26471	0.43	02/24/2021::11:18:22	0.33-0.49
333	GGT	26471	63.59	02/24/2021::11:18:35	59.00-64.60
334	CREAT	26471	2.11	02/24/2021::11:19:28	1.60-2.48
335	UREA	26471	32.41	02/24/2021::11:19:41	29.38-36.30
336	SGOT	26471	39.83	02/24/2021::11:20:20	38.48-43.96
337	UA	26471	4.39	02/24/2021::11:22:37	3.82-5.06
338	TBIL	26471	1.19	02/24/2021::11:23:56	0.78-1.46
339	TP	26471	6.58	02/24/2021::11:24:09	5.93-6.89
340	GLUC	26471	87.20	02/24/2021::11:24:22	81.69-90.21
341	CHOL	26471	259.69	02/24/2021::11:25:16	247.81-266.89
342	TG	26471	198.54	02/24/2021::11:27:01	192.95-201.43
343	HDLC	26471	63.08	02/24/2021::11:27:53	45.37-71.97
344	LDLC	26471	140.85	02/24/2021::11:27:53	125.50-144.62
345	CAL	26471	9.02	02/26/2021::10:53:45	8.25-10.01
346	ALP	26471	94.13	02/26/2021::10:54:28	96.26-111.74
347	SGPT	26471	32.46	02/26/2021::10:54:40	31.10-35.86
348	ALB	26471	3.69	02/26/2021::10:55:04	3.75-4.35
349	DBIL	26471	0.44	02/26/2021::10:55:26	0.33-0.49
350	GGT	26471	61.69	02/26/2021::10:55:39	59.00-64.60
351	CREAT	26471	1.99	02/26/2021::10:56:31	1.60-2.48
352	UREA	26471	30.42	02/26/2021::10:56:32	29.38-36.30
353	SGOT	26471	40.96	02/26/2021::10:57:37	38.48-43.96
354	UA	26471	4.16	02/26/2021::10:59:42	3.82-5.06
355	TBIL	26471	1.16	02/26/2021::11:01:00	0.78-1.46
356	TP	26471	6.39	02/26/2021::11:01:14	5.93-6.89
357	GLUC	26471	86.23	02/26/2021::11:01:27	81.69-90.21
358	CHOL	26471	265.23	02/26/2021::11:02:21	247.81-266.89
359	TG	26471	198.23	02/26/2021::11:04:05	192.95-201.43
360	HDLC	26471	54.47	02/26/2021::11:04:57	45.37-71.97
361	LDLC	26471	142.70	02/26/2021::11:04:57	125.50-144.62
362	CAL	26471	9.43	03/02/2021::10:18:28	8.25-10.01
363	ALP	26471	108.69	03/02/2021::10:19:10	96.26-111.74
364	SGPT	26471	32.50	03/02/2021::10:19:23	31.10-35.86
365	ALB	26471	4.44	03/02/2021::10:19:46	3.75-4.35
366	DBIL	26471	0.41	03/02/2021::10:20:08	0.33-0.49
367	GGT	26471	64.52	03/02/2021::10:20:21	59.00-64.60
368	CREAT	26471	2.18	03/02/2021::10:21:14	1.60-2.48
369	UREA	26471	34.54	03/02/2021::10:21:27	29.38-36.30
370	SGOT	26471	45.65	03/02/2021::10:22:06	38.48-43.96
371	UA	26471	4.91	03/02/2021::10:24:24	3.82-5.06
372	TBIL	26471	1.15	03/02/2021::10:25:43	0.78-1.46
373	TP	26471	6.90	03/02/2021::10:25:56	5.93-6.89
374	GLUC	26471	90.45	03/02/2021::10:26:09	81.69-90.21
375	CHOL	26471	268.13	03/02/2021::10:27:03	247.81-266.89
376	TG	26471	202.35	03/02/2021::10:28:48	192.95-201.43
377	HDLC	26471	57.77	03/02/2021::10:29:40	45.37-71.97
378	LDLC	26471	141.27	03/02/2021::10:29:40	125.50-144.62

Qc results

From: 01/01/2021 To: 12/03/2021 Concentration: Normal

No.	Test name	QC lot	Result	Run date	QC range
379	CAL	26471	9.29	03/04/2021::10:19:29	8.25-10.01
380	ALP	26471	113.31	03/04/2021::10:20:12	96.26-111.74
381	SGPT	26471	33.69	03/04/2021::10:20:25	31.10-35.86
382	ALB	26471	3.99	03/04/2021::10:20:48	3.75-4.35
383	DBIL	26471	0.47	03/04/2021::10:21:10	0.33-0.49
384	GGT	26471	62.07	03/04/2021::10:21:24	59.00-64.60
385	CREAT	26471	2.25	03/04/2021::10:22:16	1.60-2.48
386	UREA	26471	33.54	03/04/2021::10:22:16	29.38-36.30
387	SGOT	26471	43.50	03/04/2021::10:23:08	38.48-43.96
388	UA	26471	4.19	03/04/2021::10:25:27	3.82-5.06
389	TBIL	26471	1.17	03/04/2021::10:26:46	0.78-1.46
390	TP	26471	6.44	03/04/2021::10:26:59	5.93-6.89
391	GLUC	26471	84.65	03/04/2021::10:27:12	81.69-90.21
392	CHOL	26471	250.29	03/04/2021::10:28:06	247.81-266.89
393	TG	26471	195.44	03/04/2021::10:29:50	192.95-201.43
394	HDLC	26471	45.22	03/04/2021::10:30:42	45.37-71.97
395	LDLC	26471	153.64	03/04/2021::10:30:42	125.50-144.62
396	CAL	26471	8.50	03/06/2021::11:24:52	8.25-10.01
397	ALP	26471	99.37	03/06/2021::11:25:34	96.26-111.74
398	SGPT	26471	33.21	03/06/2021::11:25:48	31.10-35.86
399	ALB	26471	3.84	03/06/2021::11:26:11	3.75-4.35
400	DBIL	26471	0.48	03/06/2021::11:26:33	0.33-0.49
401	GGT	26471	62.33	03/06/2021::11:26:46	59.00-64.60
402	CREAT	26471	1.96	03/06/2021::11:27:38	1.60-2.48
403	UREA	26471	31.93	03/06/2021::11:27:51	29.38-36.30
404	SGOT	26471	43.12	03/06/2021::11:28:30	38.48-43.96
405	UA	26471	4.81	03/06/2021::11:30:48	3.82-5.06
406	TBIL	26471	1.02	03/06/2021::11:32:07	0.78-1.46
407	TP	26471	6.05	03/06/2021::11:32:20	5.93-6.89
408	GLUC	26471	82.91	03/06/2021::11:32:33	81.69-90.21
409	CHOL	26471	253.56	03/06/2021::11:33:27	247.81-266.89
410	TG	26471	196.82	03/06/2021::11:35:12	192.95-201.43
411	HDLC	26471	48.09	03/06/2021::11:36:04	45.37-71.97
412	LDLC	26471	130.08	03/06/2021::11:36:04	125.50-144.62
413	CAL	26471	8.98	03/09/2021::11:02:21	8.25-10.01
414	ALP	26471	101.17	03/09/2021::11:03:03	96.26-111.74
415	SGPT	26471	32.15	03/09/2021::11:03:16	31.10-35.86
416	ALB	26471	3.82	03/09/2021::11:03:39	3.75-4.35
417	DBIL	26471	0.45	03/09/2021::11:04:01	0.33-0.49
418	GGT	26471	63.00	03/09/2021::11:04:14	59.00-64.60
419	CREAT	26471	2.06	03/09/2021::11:05:07	1.60-2.48
420	UREA	26471	33.98	03/09/2021::11:05:07	29.38-36.30
421	SGOT	26471	39.89	03/09/2021::11:06:12	38.48-43.96
422	UA	26471	4.96	03/09/2021::11:08:17	3.82-5.06
423	TBIL	26471	1.16	03/09/2021::11:09:36	0.78-1.46
424	TP	26471	6.18	03/09/2021::11:09:49	5.93-6.89
425	GLUC	26471	83.09	03/09/2021::11:10:02	81.69-90.21
426	CHOL	26471	256.97	03/09/2021::11:10:57	247.81-266.89
427	TG	26471	198.41	03/09/2021::11:12:41	192.95-201.43
428	HDLC	26471	58.09	03/09/2021::11:13:33	45.37-71.97
429	LDLC	26471	132.48	03/09/2021::11:13:33	125.50-144.62
430	CAL	26471	8.91	03/12/2021::09:08:43	8.25-10.01
431	ALP	26471	104.37	03/12/2021::09:09:18	96.26-111.74
432	SGPT	26471	31.52	03/12/2021::09:09:37	31.10-35.86

Qc results

From: 01/01/2021

To: 12/03/2021

Concentration: Normal

No.	Test name	QC lot	Result	Run date	QC range
433	ALB	26471	3.75	03/12/2021::09:10:12	3.75-4.35
434	DBIL	26471	0.44	03/12/2021::09:10:36	0.33-0.49
435	GGT	26471	60.07	03/12/2021::09:10:36	59.00-64.60
436	CREAT	26471	2.03	03/12/2021::09:11:29	1.60-2.48
437	UREA	26471	35.70	03/12/2021::09:11:29	29.38-36.30
438	SGOT	26471	38.78	03/12/2021::09:12:30	38.48-43.96
439	UA	26471	3.95	03/12/2021::09:14:33	3.82-5.06
440	TP	26471	6.30	03/12/2021::09:16:04	5.93-6.89
441	GLUC	26471	88.97	03/12/2021::09:16:26	81.69-90.21
442	TBIL	26471	1.15	03/12/2021::09:16:27	0.78-1.46
443	CHOL	26471	258.04	03/12/2021::09:17:14	247.81-266.89
444	TG	26471	196.60	03/12/2021::09:18:59	192.95-201.43
445	HDLC	26471	66.34	03/12/2021::09:20:14	45.37-71.97
446	LDLC	26471	139.86	03/12/2021::09:20:25	125.50-144.62