



DIAGNØ PØINT DISTRIBUTØRS

(A Complete Solution of Lab Instruments & Reagents)

- ▶ Semi-Auto analyzer Cum Coagulation with inbuilt incubator.
- ▶ Fully auto biochemistry instrument, fluorescence enzyme immunoassay
- ▶ Electrolyte Analyzer. Hematology Automated cell Counter.
- ▶ Elisa Strip & Plate Reader. Blood Gas Analyzer
- ▶ Olympus & labomed Microscope. Remi Centrifuge & Blood Bank Product.
- ▶ All other Lab instruments & All Range of Bio - Chemistry, Elisa Kits, Rapid Cards

TIN : 3303 4901776

CST NO : 154982

Plot# 1, Shempagapoo 2nd Street,
Sriram Nagar, Sammtipuram,
Madurai - 625 010.

Off : 0452 2385456

Cell : 93666 23506, 94869 56087

email : diagnopointdistributors@gmail.com

CALIBRATION CERTIFICATE

Report No: 290051	Report Date : 02/02/2021
Prod Cat No: AF-0010	Pipette SrNo:424342
Micro Pipette Volume Range: 10	Bal Sensitivity :0.01
Temperature :25.00	Rf Humamity:60.00 %
Air pressure : 1013 hPA	Z.Factor : 1.0040 ul/mg

TEST DETAILS

Micro Pipette	No of Measurements	Mean Weight	Mean Volume
10.00ul	10	4.9610 mg	4.9808ul

SUMMARY STATISTICS

Micro Pipette	SD	Inaccuracy E%			Imprecision CV%		
		Actual	Target		Actual	Target	Status
10.00ul	0.011	-0.3831	1.3	Pass	0.2218	1.2	Pass
STATUS	PASSED	Performed by Mittal		Incharge QC- Calibration Method According to ENISO8655-6			
Pfact	Fine Care Corporation Pvt Ltd						

CALIBRATION CERTIFICATE

Report No: 290052	Report Date : 02/02/2021
Prod Cat No: AF-00500	Pipette SrNo:309205
Micro Pipette Volume Range: 500 ul	Bal Sensitivity :0.01
Temperature :25.00	Rf Humamity:60.00 %
Air pressure : 1013 hPA	Z.Factor : 1.0040 ul/mg

TEST DETAILS

Micro Pipette	No of Measurements	Mean Weight	Mean Volume
500.00ul	10	4.9610 mg	4.9808ul

SUMMARY STATISTICS

Micro Pipette	SD	Inaccuracy E%			Imprecision CV%		
		Actual	Target		Actual	Target	Status
500.00ul	0.011	-0.3831	1.3	Pass	0.2218	1.2	Pass
STATUS	PASSED	Performed by Mittal		Incharge QC- Calibration Method According to ENISO8655-6			
Pfact	Fine Care Corporation Pvt Ltd						