



DIAGNO POINT DISTRIBUTORS

(A Complete Solution of Lab Instruments & Reagents)

- ▶ Semi-Auto analyzer Cum Coagulation with inbuilt Incubator.
- ▶ Fully auto biochemistry instrument, fluorescence enzyme immunoassay
- ▶ Electrolyte Analyzer. Hematology Automated cell Counter.
- ▶ Elisa Strip & Plate Reader. Blood Gas Analyzer
- ▶ Olympus & labomed Microscope. Remi Centrifuge & Blood Bank Product.
- ▶ All other Lab instruments & All Range of Bio - Chemistry, Elisa Kits, Rapid Cards

TIN : 3303 4901776

CST NO : 154982

Plot# 1, Shenpagapoo 2nd Street,
Sriram Nagar, Sammtipuram,
Madurai - 625 010.
Off : 0452 2385456

Cell : 93666 23506, 94869 56087

email : diagnopointdistributors@gmail.com

CALIBRATION CERTIFICATE

Report No: 290053

Prod Cat No: AF-001000

Micro Pipette Volume Range: 1000 ul

Temperature :25.00

Air pressure : 1013 hPA

Report Date : 02/02/2021

Pipette SrNo:

Bal Sensitivity :0.01

Rf Humamity:60.00 %

Z.Factor : 1.0040 ul/mg

TEST DETAILS

| Micro Pipette | No of Measurements | Mean Weight | Mean Volume |
|---------------|--------------------|-------------|-------------|
| 1000.00ul | 10 | 4.9610 mg | 4.9808ul |

SUMMARY STATISTICS

| Micro Pipette | SD | Inaccuracy E% | | | Imprecision CV% | | |
|---------------|-------------------------------|---------------------|--------|--|-----------------|--------|--------|
| | | Actual | Target | | Actual | Target | Status |
| 1000.00ul | 0.011 | -0.3831 | 1.3 | Pass | 0.2218 | 1.2 | Pass |
| STATUS | PASSED | Performed by Mittal | | Incharge QC- Calibration Method According to ENISO8655-6 | | | |
| Pfact | Fine Care Corporation Pvt Ltd | | | | | | |


DIAGNO POINT DISTRIBUTORS

Plot #1, Shanbagapoo 2nd Street,
Sriram Nagar, Sammattipuram,
Kalavasal, Madurai - 625 016