



DIAGNO POINT DISTRIBUTORS

(A Complete Solution of Lab Instruments & Reagents)

- ▶ Semi-Auto analyzer Cum Coagulation with inbuilt Incubator.
- ▶ Fully auto biochemistry instrument, fluorescence enzyme immunoassay
- ▶ Electrolyte Analyzer. Hematology Automated cell Counter.
- ▶ Elisa Strip & Plate Reader. Blood Gas Analyzer
- ▶ Olympus & labomed Microscope. Remi Centrifuge & Blood Bank Product.
- ▶ All other Lab instruments & All Range of Bio - Chemistry, Elisa Kits, Rapid Cards

TIN : 3303 4901776

CST NO : 154982

Plot# 1, Shenpagapoo 2nd Street,
Sriram Nagar, Sammtipuram,
Madurai - 625 010.
Off : 0452 2385456

Cell : 93666 23506, 94869 56087

email : diagnopointdistributors@gmail.com

CALIBRATION CERTIFICATE

Report No: 193993

Report Date : 02/02/2021

Prod Cat No: AV-600

Pipette SrNo:NC404713

Micro Pipette Volume Range: 10-100ul

Bal Sensitivity :0.01

Temperature :25.00

Rf Humamity:61.00 %

Air pressure : 1012


Z.Factor : 1.0040 ul/mg

TEST DETAILS

Micro Pipette	No of Measurements	Mean Weight	Mean Volume
100.00ul	10	100.1680mg	100.5687 ul
500.00ul	10	499.4900mg	501.4880ul
1000.00ul	10	996.4830 mg	1000.4689 ul

SUMMARY STATISTICS

Micro Pipette	SD	Inaccuracy E%			Imprecision CV%		
		Actual	Target		Actual	Target	Status
100.00ul	10	0.5687	2	Pass	0.3007	0.65	Pass
500.00ul	10	0.2976	1	Pass	0.1529	0.42	Pass
1000.00ul	10	0.0469	0.6	pass	0.095	0.21	Pass
STATUS	PASSED	Performed by Mittal		Incharge QC- Calibration Method According to ENISO8655			
Pfact	Fine Care Corporation Pvt Ltd						


DIAGNO POINT DISTRIBUTORS
Plot #1, Shanbagapoo 2nd Street,
Sriram Nagar, Sammattipuram,
Kalavasal, Madurai - 625 016