

CALIBRATION CERTIFICATE

This is to Certify that Haematology Analyser Meril **CliniQuant 5** Instrument Sr. No. **CQ502000702** Installed at **Energion Health -Bangalore** has been calibrated by our engineer on **01-Oct-2020** with appropriate and sophisticated tools. Below parameters are been calibrated:

TEST ITEMS AND RESULT

Appearance	OK
AD Value Range (120<AD<350)	OK
Temperature Checking (at 37'C)	OK
Printer Test	OK
Aspiration Volume Range	OK
Stability of Pot. Dichromate	OK
Absorbency Linearity at 340	OK
PC Communication Checking	OK
Sample Test 1	OK
Sample Test 2	OK
Sample Test 3	OK

This Calibration Certificate is Valid Up to **01-Oct-2021**.

Conclusion: All parameter are within range. The Machine is working properly.
(As per Meril terms.)

Engineer Name: **Mr. PRABHAKAR RAO.K**

Designation: **Regional Service Manager**

Engineer Sign:

Stamp;

Stamp;

Stamp;

