



**PROFICIENCY TESTING REPORT**  
**ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME**  
 NABL accredited program as per ISO/IEC 17043:2010 standard  
 Organized By Department of Hematology, AIIMS, New Delhi-110029



*Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens*

EQAP CODE No. : 1196

Distribution No.: 152-D

Month/Year: February/2021

Instrument ID: SYSMEX XN 350 (11359)

Name & Contact No. of PT Co-ordinator: Dr. Seema Tyagi (Prof.), Hematology, AIIMS, Delhi,  
 Tel: 9013085730 , E-Mail : accuracy2000@gmail.com

Date of issue &amp; status of the report: 02-03-2021[Final].

### CBC and Retic Assessment

Test Parameters	S.No.	Among Lab (Accuracy Testing)						Within Lab (Precision Testing)			
		Your Result 1	Your Result 2	Your Results Sum of 2 Value	Consensus result sum of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score	Yours Results Diff. of 2 Values	Consensus Result Diff. of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score
WBC x10 <sup>3</sup> /µl	1	2.43	2.37	4.8	5.5	0.0200	-1.23	0.06	0.1	0.0050	-0.62
RBC x10 <sup>6</sup> /µl	1	4.66	4.66	9.32	9.32	0.0080	0.00	0	0.03	0.0020	-1.01
Hb g/dl	1	13.4	13.4	26.8	27	0.0200	-0.34	0	0.1	0.0070	-1.35
HCT%	1	40.6	40.6	81.2	83.9	0.1780	-0.46	0	0.3	0.0200	-0.81
MCV-fl	1	87.1	87.1	174.2	179.9	0.3170	-0.55	0	0.2	0.0170	-0.90
MCH-Pg	1	28.8	28.8	57.6	57.7	0.0540	-0.07	0	0.2	0.0140	-0.90
MCHC-g/dl	1	33	33	66	64.15	0.1300	0.41	0	0.3	0.0120	-1.01
Plt. x10 <sup>3</sup> /µl	1	96	94	190	196	0.75	-0.28	2	4	0.26	-0.45
Retic %	2	2.7	2.5	5.2	5	0.08	0.09	0.2	0.2	0.01	0.00

### P.S . Assesment

YOUR REPORT			CONSENSUS REPORT		
DLC%	3	Nrbcs=NORMOBLASTS 60% , Poly=68 L=18, E=1. Mono/Promono=5 , B1= P.M.=01, Mye=04, Meta=03, Other=	nRBC: 30 - 65, Poly: 60 - 75, Lympho: 15-30, Eos/Mono: 1-5, Blast/Myelo/Meta: 0-1		
RBC Morphology	3	ANISOPOIKILOCYTOSIS, SPHEROCYTES, MICROCYTOSIS, MACROCYTOSIS, TARGET & TEAR DROP CELLS, S HOWELL JOLLY BODIES SPHEROCYTES	Predominantly: Macrocytosis, Microcytosis, Spherocytosis, Polychromasia, Anisocytosis: Moderate: Normocytic/Normochromic, Hypo.		
Diagnosis	3	HEREDITARY SPHEROCYTOSIS	Hemolytic Anemia		



Test parameters	S.No.	Total participants covered in the current dist.	Total No. responded	% of Labs with Z Score 0-2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
				Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
WBC x10 <sup>3</sup> /µl	1	312	347	89.05	91.93	2.31	1.44	8.36	6.34
RBC x10 <sup>6</sup> /µl	1	312	347	89.63	90.78	6.92	3.46	3.17	5.19
Hb g/dl	1	312	347	91.07	93.08	6.92	2.59	2.02	4.32
HCT%	1	312	347	97.41	91.93	1.73	3.46	0.58	4.32
MCV-fl	1	312	347	97.12	85.59	1.44	8.93	0.86	4.9
MCH-Pg	1	312	347	91.35	90.78	6.05	3.46	2.31	5.48
MCHC-g/dl	1	312	347	98.27	91.93	0.29	3.75	1.15	3.75
Plt. x10 <sup>3</sup> /µl	1	312	347	93.08	91.64	3.46	5.48	3.17	2.59
ReticCount%	2	312	318	93.71	86.48	4.09	2.2	2.2	11.64
PS Assessment	3	312	335	Acceptable:91.4,Warning Signal:7.7,Unacceptable :0.9					

**Comments:**

1). Among Lab (EQA) : Results acceptable.

2). Within Lab (IQA) : Precision acceptable.

**Note-1:** EQA (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.

**IQA** ( Internal Quality Assurance) : Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

**Note-2:** Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values - Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (IQA)= (Your Result Difference of two values - Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR = 0.7413 x IQR

**Note-3:** Z score 0 to ±2: Acceptable, Z score ±2 to ±3 :Warning Signal, Z score > ±3 : Unacceptable [As per ISO/IEC 13528:2015 standard]

**Note-4:** Z score value between "0 to ±2" are texted in green colour. Z score value between "±2 to ±3" are texted in orange colour. Z score value > ±3 are texted in red colour.

**Note-5:** Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value (0.3\*SDPA). To pass the stability test, average difference in measurement values of first and last day sample ( $\bar{x}-\bar{y}$ ) should be smaller than the check value (0.3\*SDPA).

**Note-6:** ISHTM-AIIMS-EQAP does not subcontract any task of its scheme

**Note-7:** Participants are free to use methods/analyzer of their own choice.

**Note-8:** Proficiency testing (PT ) samples are sent quarterly to each participant.

**Note-9:** All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme's website [www.ishtmaiimseqap.com](http://www.ishtmaiimseqap.com).

Report authorized by,



Dr. Seema Tyagi (Prof.)

PT Co-ordinator: ISHTM-AIIMS-EQAP

Department of Hematology, AIIMS, New Delhi

-----End Of Report-----



**ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME**  
*An ISO/IEC 17043:2010 certified programme*  
Organized by  
**Department of Hematology, AIIMS, New Delhi-110029**

Email: [accuracy2000@gmail.com](mailto:accuracy2000@gmail.com)



## **PARTICIPATION CERTIFICATE**

[Certificate No. EQAP/1196/2021/04]

**Dated: 16.4.2021**

This is to certify that, Vinamra Swaraj Hospital, Navi Mumbai-400703 has participated in the "ISHTM-AIIMS External Quality Assurance Program" for the period "Jan 2020 to Dec 2020".

**Dr. Seema Tyagi (Prof, Hematology)**  
**Chief Coordinator**  
**ISHTM-AIIMS-EQAP**