

SRI SARAVANA ELECTRICAL SERVICE CENTRE**NO. 4/ 9, PERUMAL SALAI, NEW PERUNGALATHUR. CHENNAI-600 063.**

Specialist in: All kind of Scientific Instruments and Medical Equipments, Service Works Hospital, College, Clinical Lab, Blood Bank Equipments, Cooling Centrifuges, B.O.D, Incubator, Auto Clave, Furnace, Microscopes, Laminar Flow, Deep freezers, Incubators, Hot air ovens. And others Calorimeter, We Are Undertaking Annual Maintenance Contract and Instruments Calibration.

@

CALIBRATION CERTIFICATE**CUSTOMER NAME AND ADDRESS:**

STL DIAGNOSTIC CENTRE,
NO.236, C.T.H.ROAD,
CHENNAI-600 072.

CERTIFICATE NO : 3103/21
DATE OF CALIBRATION : 02-04-2021
NEXT CALIBRATION DATE : 02-04-2022

DESCRIPTION OF ITEM : REMI LABORATORY CENTRIFUGE.
MODEL NO : R.8C-
SL.NO : EFLC-2021
ROTOR HEAD : 16X15 ML . SOH
LOCATION : LABORATORY

DETAILS OF EQUIPMENT USED FOR CALIBRATION:

SR.	DESCRIPTION	CERTIFIEDBY	CERTIFICATE.NO.	VALIDITY
1.	DIGITAL TACHO METER. LUTRON MODEL. DT.2234.C. SL NO.L.207666	NORTH LAB CHENNAI. 4 NABL.	CERTIFICATE XEROX ENCLOSED	CERTIFICATE XEROX ENCLOSED

OBSERVATIONS:

SR.	INSTRUMENTSPEED DISPLAYED (RPM)	SPEED READING ON TACHOMETRE (RPM)	DIFF.IN.ACTUAL &TACHOMETRE (RPM)	REMARK
1.	1500	1550	50	Acceptable
2.	2000	2050	50	Acceptable
3.	2500	2550	50	Acceptable
4	3500	3560	60	Acceptable

Remark: 1. CALIBRATION TRACEBLE TO NATIONAL STANDARDS
2.OVERALL DEVIATION IS WITH IN SPECIFIED ACCURACY, HENCE THE INSTRUMENT IS OK.
3. RESULTS ARE RELATED ONLY TO THE ITEM CALIBRATED
4MAXIMUM TOLERANCE PERMISSIBLE IS ± 100 RPM
5 TRACEABILITY XEROX COPY ENCLOSED

AUTHORIZED SIGNATURE & CALIBRATED BY


R. DURAI BABU
(Service Engineer)