

SRI SARAVANA ELECTRICAL SERVICE CENTRE**NO. 4/ 9, PERUMAL SALAI, NEW PERUNGALATHUR. CHENNAI-600 063.**

Specialist in: All kind of Scientific Instruments and Medical Equipments, Service Works Hospital, College, Clinical Lab, Blood Bank Equipments, Cooling Centrifuges, B.O.D, Incubator, Auto Clave, Furnace, Microscopes, Laminar Flow, Deep freezers, Incubators, Hot air ovens. And others Calorimeter, We Are Undertaking Annual Maintenance Contract and Instruments Calibration.

@

CALIBRATION CERTIFICATE**CUSTOMER NAME AND ADDRESS:**

STL DIAGNOSTIC CENTRE,
NO.236, C.T.H.ROAD,
CHENNAI-600 072.

CERTIFICATE NO : 3104/21
DATE OF CALIBRATION : 02-04-2021
NEXT CALIBRATION DATE : 02-04-2022

NAME OF EQUIPMENT : PIPETTE
COMPANY NAME : BIOHIT
MODEL : VARIABLE VOLUME
SERIAL NO : 011400737
RANGE : 100-1000 μ l
LOCATION : LABORATORY
CONDITION OF THE EQUIPMENT : GOOD

Calibration Data

Test Reading (UUC)	Standard reading (Mean)	Deviation	Deviation Allowed (\pm)
μ l	μ l	μ l	μ l
100	100.01	0.01	0.5
200	200.01	0.01	0.5
500	500.01	0.01	0.5
900	900.01	0.01	0.5

- Remark: 1. CALIBRATION TRACEBLE TO NATIONAL STANDARDS
2.OVERALL DEVIATION IS WITH IN SPECIFIED ACCURACY, HENCE THE INSTRUMENT IS OK.
3. RESULTS ARE RELATED ONLY TO THE ITEM CALIBRATED
4MAXIMUM TOLERANCE PERMISSIBLE IS $\pm 0.5\mu$ l
5 TRACEABILITY XEROX COPY ENCLOSED

AUTHORIZED SIGNATURE & CALIBRATED BY

R. DURAI BABU
(Service Engineer)