



**PROFICIENCY TESTING REPORT**  
**ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME**  
 NABL accredited program as per ISO/IEC 17043:2010 standard  
 Organized By Department of Hematology, AIIMS, New Delhi-110029



*Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens*

EQAP CODE No. : 2126

Distribution No.: 149-F

Month/Year: December/2019

Instrument ID: MINDARY BC 5000

Name & Contact No. of PT Co-ordinator: Dr. Seema Tyagi (Prof.), Hematology, AIIMS, Delhi,  
 Tel: 9013085730 , E-Mail : accuracy2000@gmail.com

Date of issue &amp; status of the report: 09-03-2020[Final].

### CBC and Retic Assessment

Test Parameters	S.No.	Among Lab (Accuracy Testing)						Within Lab (Precision Testing)			
		Your Result 1	Your Result 2	Your Results Sum of 2 Value	Consensus result sum of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score	Yours Results Diff. of 2 Values	Consensus Result Diff. of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score
WBC x10 <sup>3</sup> /µl	1	6.92	4	10.92	11.63	0.0270	-0.86	2.92	0.1	0.0080	22.38
RBC x10 <sup>6</sup> /µl	1	3.77	3.5	7.27	8.06	0.0070	-3.81	0.27	0.04	0.0020	5.17
Hb g/dl	1	10.9	10	20.9	22.2	0.0200	-1.95	0.9	0.1	0.0070	5.40
HCT%	1	34.9	34	68.9	66.8	0.1410	0.46	0.9	0.3	0.0210	1.62
MCV-fl	1	92.6	80	172.6	166.25	0.3130	0.57	12.6	0.2	0.0200	27.88
MCH-Pg	1	28.9	27	55.9	55	0.0540	0.61	1.9	0.2	0.0100	7.64
MCHC-g/dl	1	31.2	31	62.2	66.7	0.1380	-0.88	0.2	0.3	0.0190	-0.27
Plt. x10 <sup>3</sup> /µl	1	166	100	266	340	0.92	-2.73	66	6	0.32	10.12
Retic %	2	5		5							

### P.S . Assesment

YOUR REPORT			CONSENSUS REPORT		
DLC%	3	Nrbcs=10-15/100 WBCs , Poly=05% L=10%, E=10%, Mono/Promono=05% , B1=09% P.M.=04%, Mye=08%, Meta=12%, Other=00%	Poly: 50-60, Lymph: 2-6, nRBC/Eo/Mono/Blast/Pro: 0-5, My: 10-20, Meta: 10-15		
RBC Morphology	3	00%	Predominantly: Normocytic Normochromic. Moderate: Micro. Mild: Hypo, Aniso.		
Diagnosis	3	CHRONIC MYELOID LEUKEMIA	Chronic Myeloid Leukemia (Chronic Phase) : CML-CP		

EQAP Code No.:  
2126

Distribution No.: 149-F Month/Year: December/2019

Instrument ID: MINDARY BC  
5000**COMBINED DATA VALUES OF TOTAL PARTICIPANTS**

Test parameters	S.No.	Total participants covered in the current dist.	Total No. responded	% of Labs with Z Score 0-2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
				Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
WBC x10 <sup>3</sup> /µl	1	450	366	90.98	84.97	3.28	6.83	5.19	7.1
RBC x10 <sup>6</sup> /µl	1	450	366	89.89	88.25	5.46	5.19	4.1	6.01
Hb g/dl	1	450	366	92.08	86.89	4.92	3.83	2.46	8.74
HCT%	1	450	366	95.08	90.16	3.01	5.74	1.37	3.55
MCV-fl	1	450	366	96.99	95.63	1.37	1.64	1.09	2.19
MCH-Pg	1	450	366	87.7	89.62	7.65	5.19	4.1	4.64
MCHC-g/dl	1	450	366	97.54	88.25	1.64	4.92	0.27	6.28
Plt. x10 <sup>3</sup> /µl	1	450	366	89.34	91.53	7.92	4.64	1.91	3.01
ReticCount%	2	450	301	94.02	74.75	2.99	20.6	2.99	7.64
PS Assessment	3	450	349	Acceptable:96.2%,Warning Signal:3.2%,Unacceptable :0.6%					

**\*Comments:**

1). **Among Lab (EQA) : CBC result for RBC unacceptable, may be due to random/human error, PS partially correct, PS Morph not reported**

2). **Within Lab (IQA) : Difference for most of the CBC results unacceptable, check precision.**

**Note-1: EQA** (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.

**IQA** ( Internal Quality Assurance) : Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

**Note-2:** Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values - Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (IQA)= (Your Result Difference of two values - Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR = 0.7413 x IQR

**Note-3:** Z score 0 to ±2: Acceptable, Z score ±2 to ±3 :Warning Signal, Z score > ±3 : Unacceptable [As per ISO/IEC 13528:2015 standard]

**Note-4:** Z score value between "0 to ±2" are texted in green colour. Z score value between "±2 to ±3" are texted in orange colour. Z score value > ±3 are texted in red colour.

**Note-5:** Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value (0.3\*SDPA). To pass the stability test, average difference in measurement values of first and last day sample ( $\bar{x}-\bar{y}$ ) should be smaller than the check value (0.3\*SDPA).

**Note-6:** ISHTM-AIIMS-EQAP does not subcontract any task of its scheme

**Note-7:** Participants are free to use methods/analyzer of their own choice.

**Note-8:** Proficiency testing (PT ) samples are sent quarterly to each participant.

**Note-9:** All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme's website [www.ishtmaiimseqap.com](http://www.ishtmaiimseqap.com).

Report authorized by,



Dr. R. Saxena

Prof & Head, Hematology, AIIMS, Delhi.

PT Co-ordinator: ISHTM-AIIMS-EQAP

-----End Of Report-----



**PROFICIENCY TESTING REPORT**  
**ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME**  
 NABL accredited program as per ISO/IEC 17043:2010 standard  
 Organized By Department of Hematology, AIIMS, New Delhi-110029



*Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens*

EQAP CODE No. : 2126

Distribution No.: 152-F

Month/Year: February/2021

Instrument ID: H560 (K11042015033)

Name & Contact No. of PT Co-ordinator: Dr. Seema Tyagi (Prof.), Hematology, AIIMS, Delhi,  
 Tel: 9013085730 , E-Mail : accuracy2000@gmail.com

Date of issue &amp; status of the report: 25-03-2021[Final].

### CBC and Retic Assessment

Test Parameters	S.No.	Among Lab (Accuracy Testing)						Within Lab (Precision Testing)			
		Your Result 1	Your Result 2	Your Results Sum of 2 Value	Consensus result sum of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score	Yours Results Diff. of 2 Values	Consensus Result Diff. of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score
WBC x10 <sup>3</sup> /µl	1	3.7	3.4	7.1	6.7	0.0230	0.75	0.3	0.1	0.0060	3.17
RBC x10 <sup>6</sup> /µl	1	4.16	4.15	8.31	8.3	0.0090	0.04	0.01	0.04	0.0030	-1.01
Hb g/dl	1	13.4	13.3	26.7	27.2	0.0300	-0.69	0.1	0.1	0.0080	0.00
HCT%	1	44	43.9	87.9	85.1	0.1870	0.52	0.1	0.4	0.0260	-0.95
MCV-fl	1	105.8	105.7	211.5	206	0.3690	0.64	0.1	0.3	0.0280	-0.39
MCH-Pg	1	32.1	32.1	64.2	65.5	0.0880	-0.70	0	0.25	0.0190	-0.84
MCHC-g/dl	1	30.4	30.3	60.7	64.1	0.1500	-0.92	0.1	0.3	0.0190	-0.67
Plt. x10 <sup>3</sup> /µl	1	112	111	223	204	1.39	0.61	1	4	0.34	-0.67
Retic %	2	6.5	6.5	13	8.5	0.17	1.11	0	0.3	0.03	-0.54

### P.S . Assesment

YOUR REPORT			CONSENSUS REPORT		
DLC%	3	Nrbcs=5-7/100 WBCs , Poly=40% L=08%, E=05%, Mono/Promono=09% , B1=05% P.M.=02%, Mye=10%, Meta=11%, Other=Stab Cells and Band cells 06% +03%	Poly: 35 - 65, Myelo: 10 - 30, Meta: 5 - 15, Promyelo: 1 - 10, nRBC/Baso/Blast/Lympho/Eos/Mono: 0 - 5		
RBC Morphology	3	Normocytic Normochromic cells	Predominantly: Normocytic/Normochromic; Moderate: Microcytosis, Anisocytosis; Mild: Hypochromia, Macrocytosis, Poikilocytosis		
Diagnosis	3	Chronic myeloid leukemia (CML) Chronic phase	Chronic Myeloid Leukemia (CML)		

EQAP Code No.: 2126 Distribution No.: 152-F Month/Year: February/2021 Instrument ID: H560 (K11042015033)

### COMBINED DATA VALUES OF TOTAL PARTICIPANTS

Test parameters	S.No.	Total participants covered in the current dist.	Total No. responded	% of Labs with Z Score 0-2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
				Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
WBC x10 <sup>3</sup> /µl	1	236	255	89.41	87.84	3.14	0.78	7.45	10.2
RBC x10 <sup>6</sup> /µl	1	236	255	89.8	91.76	5.49	3.92	4.71	3.92
Hb g/dl	1	236	255	93.33	92.16	3.53	3.53	3.14	4.31
HCT%	1	236	255	94.12	89.8	5.49	2.75	0	6.67
MCV-fl	1	236	255	91.76	93.33	5.88	2.75	1.57	3.53
MCH-Pg	1	236	255	85.1	88.63	9.41	5.1	5.1	5.49
MCHC-g/dl	1	236	255	96.08	89.02	2.75	3.14	0.39	7.06
Plt. x10 <sup>3</sup> /µl	1	236	255	89.8	88.63	5.88	3.53	3.92	7.45
ReticCount%	2	236	235	92.34	92.34	4.68	5.11	3.4	2.98
PS Assessment	3	236	244	Acceptable:95.3,Warning Signal:0.9,Unacceptable :3.8					

**\*Comments:**

1). **Among Lab (EQA) : Results acceptable.**

2). **Within Lab (IQA) : Difference in the CBC measurement values for WBC unacceptable, may be due to random/human error.**

**Note-1: EQA** (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.

**IQA** ( Internal Quality Assurance) : Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

**Note-2:** Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values - Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (IQA)= (Your Result Difference of two values - Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR = 0.7413 x IQR

**Note-3:** Z score 0 to ±2: Acceptable, Z score ±2 to ±3 :Warning Signal, Z score > ±3 : Unacceptable [As per ISO/IEC 13528:2015 standard]

**Note-4:** Z score value between "0 to ±2" are texted in green colour. Z score value between "±2 to ±3" are texted in orange colour. Z score value > ±3 are texted in red colour.

**Note-5:** Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value (0.3\*SDPA). To pass the stability test, average difference in measurement values of first and last day sample ( $\bar{x}-\bar{y}$ ) should be smaller than the check value (0.3\*SDPA).

**Note-6:** ISHTM-AIIMS-EQAP does not subcontract any task of its scheme

**Note-7:** Participants are free to use methods/analyzer of their own choice.

**Note-8:** Proficiency testing (PT ) samples are sent quarterly to each participant.

**Note-9:** All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme's website www.ishtmaiimseqap.com.

Report authorized by,



Dr. Seema Tyagi (Prof.)

PT Co-ordinator: ISHTM-AIIMS-EQAP

Department of Hematology, AIIMS, New Delhi

-----End Of Report-----