Digital Thermometer Calibration Verification Record Sheet Location: RH Shikrapur, Counseling Room Property No/ Model No: TA 298 Procedure using Digital Thermometer: In house verification of Digital thermometer using standard calibrated thermometer Date: 30/04/2021 Certified thermometer ID: CRV/MICRO/VIRO/03/2017-18, 201705038856 Verification Done by: Ms Bandgar Rupali, Technical Officer, SRL, AFMC, Pune Range of refrigerator: 2-8 Degree Celcius Sr.No. Time Test Temp Degree **Certified Temp Temp Difference** Celcius **Degree Celcius Degree Celcius** 1 11.00 am 2.5 3 0.5 2 11.30 am 3.1 1 2.1 3 12.00 pm 2 2.8 **8.0** 5 12.30 pm 4 3.5 0.5 Mean 2.5 3.225 0.7 SD 0.53 %CV 16.48 % Accuracy= 29 Accept/ Reject: Accepted since average temp difference is within +/- 1 degree celcius Final Interpretation: Digital thermometer is in working condition Action Taken:

Correction factor:

Remark: NA

