

Timer Calibration Verification Record Sheet

Location: ICTC Lab, RH Shikrapur

Property No: DT 1045

Procedure using Timer: Inhouse verification of timer using certified standard calibrated Timer

Date: 13/02/2021

Certified timer ID: CRV/MICRO/VIRO/03/2017-18, DT-02

Verification Done by: Ms Bandgar Rupali, Technical Officer, SRL,AFMC,Pune

Sr.No.	Timer Setting (Sec)	Certified Time (Sec)	Time Difference (Sec)	Timer Setting (Sec)	Certified Time (Sec)	Time Difference (Sec)	Timer Setting (Sec)	Certified Time (Sec)
1	300	298	-2	600	597	-3	900	904
2	300	300	0	600	598	-2	900	900
3	300	300	0	600	603	3	900	900
Mean	300	299.3	-0.7	600	599.3	-0.6	900	899.3
SD	-	1.15	-	-	3.21	-	-	2.31
%CV	-	0.3858004	-	-	0.5363842	-	-	0.2567999
%	-	-0.2333333	-	-	-0.1166667	-	-	-0.0777778

Accept/ Reject: Accepted since variation is within +/- 60 Sec

Final Interpretation: Timer is in good working condition

Action Taken: -

Remark: -

Ms. Rupali Bandgar
Technical Officer
SRL, MSACS
AFMC, Pune - 411 040