

Timer Calibration Verification Record Sheet

Location: Room No.08

Property No: Khadaki Cantonment Hosp, Hosp property

Procedure using Timer: Inhouse verification of timer using certified standard calibrated Timer

Date: 17/03/2021

Certified timer ID: CRV/MICRO/VIRO/03/2017-18, DT-02

Verification Done by: Ms Bandgar Rupali, Technical Officer, SRL,AFMC,Pune

Sr.No.	Timer Setting (Sec)	Certified Time (Sec)	Time Difference (Sec)	Timer Setting (Sec)	Certified Time (Sec)	Time Difference (Sec)	Timer Setting (Sec)	Certified Time (Sec)	Time Difference (Sec)
1	300	299	-1	600	598	-2	900	900	0
2	300	298	-2	600	598	-2	900	900	0
3	300	300	0	600	600	0	900	899	-1
Mean	300	299	-1	600	598.6	-1.3	900	899.6	-0.3
SD	-	1.00	-	-	1.15	-	-	0.58	-
%CV	-	0.3344482	-	-	0.1929002	-	-	0.0641786	-
% Accuracy	-	-0.3333333	-	-	-0.2333333	-	-	-0.0444444	-

Accept/ Reject: Accepted since there is no variation.

Final Interpretation: Timer is in good working condition

Action Taken: -NA

Remark: -NA


Ms. Rupali Bandgar
Technical Officer
SRL, MSACS
AFMC, Pune - 411 040