

arkray

arkray healthcare Pvt. Ltd.

Plot no. 336, 338, 340, Road no. 3, G.I.D.C.,

Sachin 394 230, (Surat), India. Telephone: 0261-239 7712 to 15

Facsimile: 0261-239 7719, URI: <http://www.arkray.com>

# *Certificate Of Calibration*


**This is to certify that Urine Analyser U500 with Serial number  
297B20005F9; installed at St. Therisas Multi Specialty Hospital  
Opposite Rythy Bazar, Erragadda,  
Hyderabad, Telangana 500018  
has been calibrated on 02/02/2021.**

**Controls Lot Nos : C1L9F39,C1L9F40**

**Controls Expiry : 08/2021**

**Control has passed in the machine and results are  
Came with in the range.**

**Note: Next Calibration (Due date : 02/08/2021)**

<p>For ARKRAY Healthcare Pvt. Ltd</p>  <p>Authorised Signatory <b>(Asst. Manager , Technical support cell)</b></p>	<p><b>Customer's Signature and Stamp</b></p>
---	--

QUALITY CHECK

02-02-2021 15:30

LEVEL 1

REFERENCE VALUES	RESULTS
LEU - neg -	neg
NIT - neg -	neg
URO - 0.2mg/dl -	0.2mg/dl
PRO - neg -	neg
pH - 6.0 -	6.0
BLO - neg -	neg
SG - 1.000 -	1.000
KET - neg -	neg
BIL - neg -	neg
GLU - neg -	neg

QUALITY CHECK

02-02-2021 15:32

LEVEL 2

REFERENCE VALUES	RESULTS
LEU - 3+ 500Leu/uL	- 3+ 500Leu/uL
NIT - + POS	- pos
URO - 3+ 8mg/dl	- 3+ 8mg/dl
PRO - 3+ 300mg/dl	- 3+ 300mg/dl
pH - 10	- 9.5
BLO - 3+ 200Ery/uL	- 3+ 200Ery/uL
SG - 1.020	- 1.015
KET - 3+ 80mg/dl	- 3+ 80mg/dl
BIL - 3+ 4mg/dL	- 3+ 4mg/dL
GLU - 3+ 1000mg/dL	- 3+ 1000mg/dL