



CC-2806

CALIBRATION CERTIFICATE

In accordance with ISO / IEC-17025 : 2017

F10-CC-03

Page : 1 of 1

Certificate No. : SIMCO/2011/761/MES/05	Issue Date : 19/11/2020
1. Customer Name & Address: M/S. ST THERESA'S MULTI SPECIALITY HOSPITAL, Erragadda, Hyderabad.	ULR- CC 2806 20100007873 F
	Reference Date : 15/11/2020
	Calibration Date : 17/11/2020
	Calibration Due Date : 16/11/2021

2. Details of Instrument Under Calibration:

Description : Centrifuge	Model : RM12
Make : REMI	
Range : Upto 5000 RPM	
ID.No : STH-BME-LAB-CENT-003	
Serial No : IJL2654	

3. Detail of Standard Instruments Used :

Instrument Used	Serial/ID. No.	Valid up to	Certificate No.
Digital Tachometer	175-0010V	06/03/2021	CC/ECL/2263/19-20

4.Environmental Conditions:

Standard Temperature : (25±4)°C

Relative Humidity : (50±20)% Rh

Actual Temperature : 25.3 °C

Actual Relative Humidity : 58.9 % Rh

5.Calibration Procedure: SOP-RPM-01

6.Calibration Results:

SL. No.	Set (RPM)	Standard Reading (RPM)	UUC Reading (RPM)	Error (RPM)	Expanded Uncertainty in ± (RPM)
1	500	500.0	500	0.0	6
2	1000	999.5	1000	0.5	6
3	2000	1999.3	2000	0.7	6
4	4000	3999.1	4000	0.9	6
5	5000	4997.7	5000	2.3	6

7. Remarks:

- The instrument was received in good condition and was calibrated at Site.
- This certificate pertains only to the item calibrated.
- The calibration results reported in this certificate are valid at the time of and at the stated environmental conditions.
- The calibration interval is determined based on customer's requirements.
- The calibration is traceable to National standards as per traceability details given in the certificate.
- This calibration certificate shall not be reproduced in full, except with prior written approval of Managing Director, SIMCO Calibration Laboratory.
- This calibration certificate is meant for scientific and industrial purpose only.
- The NABL Symbol is used as per NABL guidelines in NABL-133.
- The Measurement Uncertainty is reported approximately at 95% confidence level with coverage factor $k = 2$

Gintreddy
Calibrated by

N.V. Kameswara Rao
Mr.N.V.Kameswara Rao
Technical Manager
AUTHORISED SIGNATORY