

Timer Calibration Verification Record Sheet

Location: ICTC Lab, RGMC Kalwa

Property No: RGMC Kalwa

Procedure using Timer: Inhouse verification of timer using certified standard calibrated Timer

Date: 24/04/2021

Certified timer ID: CRV/MICRO/VIRO/03/2017-18, DT-02

Verification Done by: Ms Bandgar Rupali, Technical Officer, SRL,AFMC,Pune

Sr.No.	Timer Setting (Sec)	Certified Time (Sec)	Time Difference (Sec)	Timer Setting (Sec)	Certified Time (Sec)	Time Difference (Sec)	Timer Setting (Sec)	Certified Time (Sec)	Time Difference (Sec)
1	300	300	0	600	598	-2	900	899	-1
2	300	300	0	600	599	-1	900	900	0
3	300	300	0	600	599	-1	900	900	0
Mean	300	300	0	600	598.66667	-1.3333333	900	899.66667	-0.3333333
SD	-	0.00	-	-	0.58	-	-	0.58	-
%CV	-	0	-	-	0.0964394	-	-	0.0641738	-
%	-	0	-	-	-0.222222	-	-	-0.037037	-

Accept/ Reject: Accepted since average difference is within +/- 60 Sec

Final Interpretation: Timer is in good working condition

Action Taken: - NA

Remark: - NA

Rupali
Ms. Rupali Bandgar
 Technical Officer
 SRL, MSACS
 EMG, Pune - 411 040