

## Calibration Certificate

This is to Certify that Biochemistry Analyzer Meril **AutoQuant-200i** Instrument Sr. No. **AQ091044** Installed at **Energion Health-Bangalore** has been calibrated by our engineer on **01-Oct-2020** with appropriate and sophisticated tools. Below parameter has been calibrated:

### TEST ITEMS AND RESULT

Appearance	No Imbalance Probe Movement
Leakage in Water tubing or CRU	No water leakage
Ambient temp ( 15- 32 ) Degree	24.0 degree C
Lamp Check	Ok
Sample Dispersing Volume	In range ( 0 to 30 ul)
Reagent cooling Temperature	2.8 ° C ( 0 to 15 Degree)
Optical Path length	0.65
Waster & Water Sensor	Working Properly
Threshold value	±34000 (≥25,000)
Valve & Motor	Working Properly
Reagent Dispensing Volume	In range ( 0 to 150 ul)
Power Supply and Earthing	230 Volts and Grounding less then 5 V
CV-Kinetic/End point/two Point	CV- 1.77% /2.42% /2.80%
SD-Kinetic/End point/two Point	SD- 1.15/0.39/0.69
Mean-Kinetic/End point/two Point	Mean- 97.55/14.34/24.79
Hardware Check-Functional Test	OK5415152524141112
Light Quality Check	OK
Cuvette Absorbance	OK
Cuvette Blank Value	OK >33000
Incubation temp.	37.0 ° C
Room temp.	25.0 degree C

This Calibration Certificate is Valid Up to **01-Oct-2021**.

Conclusion: All parameter are within range. The Machine is working properly.  
(As per Meril terms.)

Engineer Name: **Mr. PRABHAKAR RAO.K**

Designation: **Regional Service Manager**

Engineer Sign:

Stamp;



### Qc results

From: 01/01/2021

To: 12/03/2021

Concentration: Normal

No.	Test name	QC lot	Result	Run date	QC range
1	CAL	26471	9.27	01/05/2021::17:19:59	8.25-10.01
2	ALP	26471	102.31	01/05/2021::17:20:10	96.26-111.74
3	SGPT	26471	31.97	01/05/2021::17:20:31	31.10-35.86
4	ALB	26471	4.12	01/05/2021::17:21:34	3.75-4.35
5	GGT	26471	62.27	01/05/2021::17:22:10	59.00-64.60
6	CREAT	26471	1.98	01/05/2021::17:22:24	1.60-2.48
7	DBIL	26471	0.46	01/05/2021::17:22:24	0.33-0.49
8	UREA	26471	36.35	01/05/2021::17:23:42	29.38-36.30
9	SGOT	26471	41.91	01/05/2021::17:24:23	38.48-43.96
10	CAL	26471	9.28	01/05/2021::17:48:30	8.25-10.01
11	ALP	26471	100.68	01/05/2021::17:48:40	96.26-111.74
12	SGPT	26471	33.64	01/05/2021::17:49:02	31.10-35.86
13	ALB	26471	4.44	01/05/2021::17:50:16	3.75-4.35
14	CREAT	26471	1.96	01/05/2021::17:51:06	1.60-2.48
15	GGT	26471	61.34	01/05/2021::17:51:06	59.00-64.60
16	DBIL	26471	0.43	01/05/2021::17:51:19	0.33-0.49
17	UREA	26471	34.05	01/05/2021::17:52:24	29.38-36.30
18	SGOT	26471	41.85	01/05/2021::17:53:05	38.48-43.96
19	UA	26471	4.37	01/05/2021::17:54:21	3.82-5.06
20	TBIL	26471	1.16	01/05/2021::17:55:12	0.78-1.46
21	TP	26471	6.46	01/05/2021::17:55:28	5.93-6.89
22	GLUC	26471	89.12	01/05/2021::17:55:41	81.69-90.21
23	CHOL	26471	260.82	01/05/2021::17:58:17	247.81-266.89
24	HDLC	26471	66.12	01/05/2021::18:00:28	45.37-71.97
25	LDLC	26471	134.87	01/05/2021::18:00:28	125.50-144.62
26	TG	26471	199.53	01/05/2021::18:20:52	192.95-201.43
27	CAL	26471	9.41	01/07/2021::11:18:38	8.25-10.01
28	ALP	26471	103.01	01/07/2021::11:19:22	96.26-111.74
29	SGPT	26471	35.32	01/07/2021::11:19:38	31.10-35.86
30	ALB	26471	4.25	01/07/2021::11:19:53	3.75-4.35
31	DBIL	26471	0.43	01/07/2021::11:20:26	0.33-0.49
32	GGT	26471	60.17	01/07/2021::11:20:47	59.00-64.60
33	CREAT	26471	1.77	01/07/2021::11:21:27	1.60-2.48
34	UREA	26471	33.36	01/07/2021::11:21:27	29.38-36.30
35	SGOT	26471	41.91	01/07/2021::11:22:19	38.48-43.96
36	UA	26471	4.37	01/07/2021::11:24:25	3.82-5.06
37	TBIL	26471	1.16	01/07/2021::11:26:08	0.78-1.46
38	TP	26471	6.55	01/07/2021::11:26:08	5.93-6.89
39	GLUC	26471	87.92	01/07/2021::11:26:21	81.69-90.21
40	CHOL	26471	260.82	01/07/2021::11:27:15	247.81-266.89
41	TG	26471	199.53	01/07/2021::11:28:59	192.95-201.43
42	HDLC	26471	66.12	01/07/2021::11:30:04	45.37-71.97
43	LDLC	26471	147.03	01/07/2021::12:53:05	125.50-144.62
44	DBIL	26471	0.43	01/07/2021::15:20:50	0.33-0.49
45	CAL	26471	9.45	01/09/2021::10:56:44	8.25-10.01
46	ALP	26471	105.14	01/09/2021::10:57:25	96.26-111.74
47	SGPT	26471	34.73	01/09/2021::10:57:38	31.10-35.86
48	ALB	26471	4.29	01/09/2021::10:58:01	3.75-4.35
49	DBIL	26471	0.41	01/09/2021::10:58:23	0.33-0.49
50	GGT	26471	60.79	01/09/2021::10:58:36	59.00-64.60
51	UREA	26471	33.43	01/09/2021::10:59:42	29.38-36.30
52	SGOT	26471	41.11	01/09/2021::11:00:21	38.48-43.96
53	UA	26471	4.44	01/09/2021::11:02:38	3.82-5.06
54	TBIL	26471	1.18	01/09/2021::11:03:57	0.78-1.46

### Qc results

From: 01/01/2021 To: 12/03/2021 Concentration: Normal

No.	Test name	QC lot	Result	Run date	QC range
55	TP	26471	6.64	01/09/2021::11:04:10	5.93-6.89
56	GLUC	26471	88.36	01/09/2021::11:04:23	81.69-90.21
57	CHOL	26471	259.31	01/09/2021::11:05:17	247.81-266.89
58	TG	26471	198.09	01/09/2021::11:07:02	192.95-201.43
59	LDLC	26471	141.04	01/09/2021::11:07:54	125.50-144.62
60	HDLC	26471	63.96	01/09/2021::11:08:07	45.37-71.97
61	ALP	26471	99.81	01/12/2021::10:53:11	96.26-111.74
62	SGPT	26471	32.37	01/12/2021::10:53:24	31.10-35.86
63	ALB	26471	4.12	01/12/2021::10:53:47	3.75-4.35
64	DBIL	26471	0.43	01/12/2021::10:54:10	0.33-0.49
65	GGT	26471	62.03	01/12/2021::10:54:23	59.00-64.60
66	CREAT	26471	2.47	01/12/2021::10:55:15	1.60-2.48
67	UREA	26471	30.67	01/12/2021::10:55:15	29.38-36.30
68	SGOT	26471	39.52	01/12/2021::10:56:20	38.48-43.96
69	UA	26471	4.28	01/12/2021::10:58:25	3.82-5.06
70	TBIL	26471	1.17	01/12/2021::10:59:44	0.78-1.46
71	TP	26471	6.29	01/12/2021::10:59:57	5.93-6.89
72	GLUC	26471	84.73	01/12/2021::11:00:10	81.69-90.21
73	CHOL	26471	265.50	01/12/2021::11:01:04	247.81-266.89
74	TG	26471	195.07	01/12/2021::11:02:48	192.95-201.43
75	HDLC	26471	60.04	01/12/2021::11:03:40	45.37-71.97
76	LDLC	26471	133.53	01/12/2021::11:03:41	125.50-144.62
77	CAL	26471	8.53	01/14/2021::11:27:21	8.25-10.01
78	ALP	26471	100.87	01/14/2021::11:28:04	96.26-111.74
79	SGPT	26471	32.08	01/14/2021::11:28:17	31.10-35.86
80	ALB	26471	3.91	01/14/2021::11:28:40	3.75-4.35
81	DBIL	26471	0.44	01/14/2021::11:29:02	0.33-0.49
82	GGT	26471	62.34	01/14/2021::11:29:15	59.00-64.60
83	CREAT	26471	2.07	01/14/2021::11:30:08	1.60-2.48
84	UREA	26471	34.48	01/14/2021::11:30:08	29.38-36.30
85	SGOT	26471	41.39	01/14/2021::11:31:13	38.48-43.96
86	UA	26471	4.36	01/14/2021::11:33:18	3.82-5.06
87	TBIL	26471	1.16	01/14/2021::11:34:37	0.78-1.46
88	TP	26471	6.29	01/14/2021::11:34:50	5.93-6.89
89	GLUC	26471	85.53	01/14/2021::11:35:03	81.69-90.21
90	CHOL	26471	258.31	01/14/2021::11:35:57	247.81-266.89
91	TG	26471	196.64	01/14/2021::11:37:42	192.95-201.43
92	HDLC	26471	68.23	01/14/2021::11:38:33	45.37-71.97
93	LDLC	26471	137.40	01/14/2021::11:38:34	125.50-144.62
94	ALP	26471	105.14	01/16/2021::11:16:59	96.26-111.74
95	ALB	26471	3.92	01/16/2021::11:17:25	3.75-4.35
96	SGPT	26471	33.55	01/16/2021::11:17:25	31.10-35.86
97	GGT	26471	62.66	01/16/2021::11:18:20	59.00-64.60
98	DBIL	26471	0.38	01/16/2021::11:18:57	0.33-0.49
99	CREAT	26471	1.94	01/16/2021::11:19:10	1.60-2.48
100	UREA	26471	34.05	01/16/2021::11:20:12	29.38-36.30
101	SGOT	26471	42.78	01/16/2021::11:21:04	38.48-43.96
102	CAL	26471	8.62	01/16/2021::11:21:56	8.25-10.01
103	UA	26471	4.32	01/16/2021::11:23:10	3.82-5.06
104	TBIL	26471	1.19	01/16/2021::11:23:36	0.78-1.46
105	TP	26471	6.42	01/16/2021::11:23:49	5.93-6.89
106	GLUC	26471	84.58	01/16/2021::11:24:02	81.69-90.21
107	CHOL	26471	248.76	01/16/2021::11:25:10	247.81-266.89
108	TG	26471	197.60	01/16/2021::11:27:33	192.95-201.43

### Qc results

From: 01/01/2021

To: 12/03/2021

Concentration: Normal

No.	Test name	QC lot	Result	Run date	QC range
109	LDLC	26471	136.12	01/16/2021::11:28:12	125.50-144.62
110	HDLC	26471	54.00	01/16/2021::11:28:25	45.37-71.97
111	CAL	26471	9.42	01/23/2021::11:07:27	8.25-10.01
112	ALP	26471	104.43	01/23/2021::11:08:09	96.26-111.74
113	SGPT	26471	34.43	01/23/2021::11:08:22	31.10-35.86
114	ALB	26471	3.97	01/23/2021::11:08:45	3.75-4.35
115	DBIL	26471	0.40	01/23/2021::11:09:07	0.33-0.49
116	GGT	26471	63.59	01/23/2021::11:09:21	59.00-64.60
117	CREAT	26471	2.30	01/23/2021::11:10:13	1.60-2.48
118	UREA	26471	30.52	01/23/2021::11:10:13	29.38-36.30
119	SGOT	26471	40.00	01/23/2021::11:11:05	38.48-43.96
120	UA	26471	4.41	01/23/2021::11:13:23	3.82-5.06
121	TBIL	26471	1.17	01/23/2021::11:14:42	0.78-1.46
122	GLUC	26471	86.87	01/23/2021::11:15:08	81.69-90.21
123	CHOL	26471	259.16	01/23/2021::11:16:02	247.81-266.89
124	TG	26471	194.48	01/23/2021::11:17:46	192.95-201.43
125	HDLC	26471	52.04	01/23/2021::11:18:38	45.37-71.97
126	LDLC	26471	133.38	01/23/2021::11:18:39	125.50-144.62
127	CAL	26471	9.58	01/27/2021::10:51:05	8.25-10.01
128	ALP	26471	104.43	01/27/2021::10:51:48	96.26-111.74
129	SGPT	26471	32.08	01/27/2021::10:52:01	31.10-35.86
130	ALB	26471	3.97	01/27/2021::10:52:24	3.75-4.35
131	DBIL	26471	0.37	01/27/2021::10:52:46	0.33-0.49
132	GGT	26471	59.55	01/27/2021::10:52:59	59.00-64.60
133	CREAT	26471	1.97	01/27/2021::10:53:52	1.60-2.48
134	UREA	26471	32.05	01/27/2021::10:53:52	29.38-36.30
135	SGOT	26471	41.11	01/27/2021::10:54:44	38.48-43.96
136	UA	26471	4.27	01/27/2021::10:57:03	3.82-5.06
137	TBIL	26471	1.13	01/27/2021::10:58:21	0.78-1.46
138	GLUC	26471	86.48	01/27/2021::10:58:47	81.69-90.21
139	CHOL	26471	256.10	01/27/2021::10:59:41	247.81-266.89
140	TG	26471	198.57	01/27/2021::11:01:26	192.95-201.43
141	HDLC	26471	54.12	01/27/2021::11:02:18	45.37-71.97
142	LDLC	26471	131.36	01/27/2021::11:02:18	125.50-144.62
143	CAL	26471	9.23	01/29/2021::10:48:33	8.25-10.01
144	ALP	26471	113.66	01/29/2021::10:49:16	96.26-111.74
145	SGPT	26471	34.43	01/29/2021::10:49:29	31.10-35.86
146	ALB	26471	3.90	01/29/2021::10:49:51	3.75-4.35
147	DBIL	26471	0.49	01/29/2021::10:50:14	0.33-0.49
148	GGT	26471	63.90	01/29/2021::10:50:27	59.00-64.60
149	CREAT	26471	2.24	01/29/2021::10:51:19	1.60-2.48
150	UREA	26471	31.62	01/29/2021::10:51:20	29.38-36.30
151	SGOT	26471	43.79	01/29/2021::10:52:25	38.48-43.96
152	UA	26471	4.34	01/29/2021::10:54:30	3.82-5.06
153	TBIL	26471	1.19	01/29/2021::10:55:48	0.78-1.46
154	TP	26471	6.22	01/29/2021::10:56:01	5.93-6.89
155	GLUC	26471	85.92	01/29/2021::10:56:15	81.69-90.21
156	CHOL	26471	257.45	01/29/2021::10:57:09	247.81-266.89
157	TG	26471	197.30	01/29/2021::10:58:53	192.95-201.43
158	HDLC	26471	56.65	01/29/2021::10:59:45	45.37-71.97
159	LDLC	26471	133.79	01/29/2021::10:59:45	125.50-144.62
160	CAL	26471	9.36	02/01/2021::11:21:30	8.25-10.01
161	ALP	26471	107.98	02/01/2021::11:22:12	96.26-111.74
162	SGPT	26471	34.14	02/01/2021::11:22:25	31.10-35.86

### Qc results

From: 01/01/2021 To: 12/03/2021 Concentration: Normal

No.	Test name	QC lot	Result	Run date	QC range
163	ALB	26471	3.97	02/01/2021::11:22:48	3.75-4.35
164	DBIL	26471	0.41	02/01/2021::11:23:11	0.33-0.49
165	GGT	26471	64.21	02/01/2021::11:23:24	59.00-64.60
166	CREAT	26471	1.98	02/01/2021::11:24:17	1.60-2.48
167	UREA	26471	34.40	02/01/2021::11:24:17	29.38-36.30
168	SGOT	26471	44.35	02/01/2021::11:25:09	38.48-43.96
169	UA	26471	4.42	02/01/2021::11:27:27	3.82-5.06
170	TBIL	26471	1.19	02/01/2021::11:28:46	0.78-1.46
171	TP	26471	6.26	02/01/2021::11:28:59	5.93-6.89
172	GLUC	26471	86.84	02/01/2021::11:29:12	81.69-90.21
173	CHOL	26471	257.60	02/01/2021::11:30:06	247.81-266.89
174	TG	26471	192.66	02/01/2021::11:31:50	192.95-201.43
175	HDLC	26471	58.00	02/01/2021::11:32:42	45.37-71.97
176	LDLC	26471	129.31	02/01/2021::11:32:42	125.50-144.62
177	CAL	26471	9.54	02/03/2021::10:50:42	8.25-10.01
178	ALP	26471	107.27	02/03/2021::10:51:24	96.26-111.74
179	SGPT	26471	33.55	02/03/2021::10:51:37	31.10-35.86
180	ALB	26471	3.96	02/03/2021::10:52:00	3.75-4.35
181	DBIL	26471	0.44	02/03/2021::10:52:22	0.33-0.49
182	GGT	26471	62.97	02/03/2021::10:52:35	59.00-64.60
183	CREAT	26471	1.85	02/03/2021::10:53:28	1.60-2.48
184	UREA	26471	33.12	02/03/2021::10:53:41	29.38-36.30
185	SGOT	26471	44.07	02/03/2021::10:54:20	38.48-43.96
186	UA	26471	4.35	02/03/2021::10:56:38	3.82-5.06
187	TBIL	26471	1.19	02/03/2021::10:57:57	0.78-1.46
188	TP	26471	6.36	02/03/2021::10:58:10	5.93-6.89
189	GLUC	26471	86.04	02/03/2021::10:58:23	81.69-90.21
190	CHOL	26471	255.54	02/03/2021::10:59:17	247.81-266.89
191	TG	26471	199.11	02/03/2021::11:01:01	192.95-201.43
192	HDLC	26471	63.39	02/03/2021::11:01:54	45.37-71.97
193	LDLC	26471	132.88	02/03/2021::11:01:54	125.50-144.62
194	CAL	26471	8.88	02/05/2021::11:18:58	8.25-10.01
195	ALP	26471	99.81	02/05/2021::11:19:40	96.26-111.74
196	SGPT	26471	34.14	02/05/2021::11:19:53	31.10-35.86
197	ALB	26471	3.93	02/05/2021::11:20:16	3.75-4.35
198	DBIL	26471	0.46	02/05/2021::11:20:39	0.33-0.49
199	GGT	26471	63.59	02/05/2021::11:20:52	59.00-64.60
200	CREAT	26471	1.93	02/05/2021::11:21:44	1.60-2.48
201	UREA	26471	33.05	02/05/2021::11:21:57	29.38-36.30
202	SGOT	26471	42.66	02/05/2021::11:22:36	38.48-43.96
203	UA	26471	4.45	02/05/2021::11:24:55	3.82-5.06
204	TBIL	26471	1.20	02/05/2021::11:26:01	0.78-1.46
205	TP	26471	6.54	02/05/2021::11:26:27	5.93-6.89
206	GLUC	26471	85.30	02/05/2021::11:26:40	81.69-90.21
207	CHOL	26471	257.23	02/05/2021::11:27:34	247.81-266.89
208	TG	26471	195.68	02/05/2021::11:29:19	192.95-201.43
209	HDLC	26471	56.73	02/05/2021::11:30:11	45.37-71.97
210	LDLC	26471	132.00	02/05/2021::11:30:11	125.50-144.62
211	CAL	26471	9.09	02/08/2021::11:10:04	8.25-10.01
212	ALP	26471	96.97	02/08/2021::11:10:42	96.26-111.74
213	SGPT	26471	32.08	02/08/2021::11:11:01	31.10-35.86
214	GGT	26471	62.34	02/08/2021::11:12:05	59.00-64.60
215	DBIL	26471	0.42	02/08/2021::11:13:19	0.33-0.49
216	UREA	26471	31.75	02/08/2021::11:14:12	29.38-36.30

### Qc results

From: 01/01/2021

To: 12/03/2021

Concentration: Normal

No.	Test name	QC lot	Result	Run date	QC range
217	SGOT	26471	41.53	02/08/2021::11:15:04	38.48-43.96
218	UA	26471	4.57	02/08/2021::11:15:58	3.82-5.06
219	TP	26471	6.23	02/08/2021::11:17:38	5.93-6.89
220	GLUC	26471	83.65	02/08/2021::11:17:51	81.69-90.21
221	TBIL	26471	1.16	02/08/2021::11:18:04	0.78-1.46
222	CHOL	26471	258.47	02/08/2021::11:18:43	247.81-266.89
223	UREA	26471	30.75	02/08/2021::11:20:14	29.38-36.30
224	TG	26471	193.08	02/08/2021::11:20:27	192.95-201.43
225	LDLC	26471	142.43	02/08/2021::11:22:24	125.50-144.62
226	HDLC	26471	57.08	02/08/2021::11:22:37	45.37-71.97
227	ALB	26471	3.77	02/08/2021::11:36:43	3.75-4.35
228	UREA	26471	30.56	02/08/2021::11:37:09	29.38-36.30
229	CREAT	26471	2.78	02/08/2021::11:39:19	1.60-2.48
230	CAL	26471	9.05	02/10/2021::11:16:00	8.25-10.01
231	ALP	26471	98.39	02/10/2021::11:16:43	96.26-111.74
232	SGPT	26471	32.96	02/10/2021::11:16:56	31.10-35.86
233	ALB	26471	3.90	02/10/2021::11:17:19	3.75-4.35
234	DBIL	26471	0.48	02/10/2021::11:17:41	0.33-0.49
235	GGT	26471	62.34	02/10/2021::11:17:54	59.00-64.60
236	CREAT	26471	2.06	02/10/2021::11:18:47	1.60-2.48
237	UREA	26471	31.31	02/10/2021::11:18:47	29.38-36.30
238	SGOT	26471	40.96	02/10/2021::11:19:39	38.48-43.96
239	UA	26471	4.62	02/10/2021::11:21:57	3.82-5.06
240	TBIL	26471	1.18	02/10/2021::11:23:16	0.78-1.46
241	TP	26471	6.51	02/10/2021::11:23:29	5.93-6.89
242	GLUC	26471	82.57	02/10/2021::11:23:42	81.69-90.21
243	CHOL	26471	259.2	02/10/2021::11:24:36	247.81-266.89
244	TG	26471	197.79	02/10/2021::11:26:20	192.95-201.43
245	HDLC	26471	57.23	02/10/2021::11:27:13	45.37-71.97
246	LDLC	26471	140.04	02/10/2021::11:27:13	125.50-144.62
247	CAL	26471	8.94	02/12/2021::11:05:54	8.25-10.01
248	ALP	26471	103.01	02/12/2021::11:06:37	96.26-111.74
249	SGPT	26471	34.43	02/12/2021::11:06:50	31.10-35.86
250	ALB	26471	3.88	02/12/2021::11:07:13	3.75-4.35
251	DBIL	26471	0.48	02/12/2021::11:07:35	0.33-0.49
252	GGT	26471	65.14	02/12/2021::11:07:48	59.00-64.60
253	CREAT	26471	2.12	02/12/2021::11:08:41	1.60-2.48
254	UREA	26471	32.15	02/12/2021::11:08:54	29.38-36.30
255	SGOT	26471	40.96	02/12/2021::11:09:33	38.48-43.96
256	UA	26471	4.35	02/12/2021::11:11:51	3.82-5.06
257	TBIL	26471	1.22	02/12/2021::11:13:10	0.78-1.46
258	TP	26471	6.69	02/12/2021::11:13:23	5.93-6.89
259	GLUC	26471	86.12	02/12/2021::11:13:36	81.69-90.21
260	CHOL	26471	253.03	02/12/2021::11:14:30	247.81-266.89
261	TG	26471	199.96	02/12/2021::11:16:15	192.95-201.43
262	HDLC	26471	57.54	02/12/2021::11:17:07	45.37-71.97
263	CAL	26471	8.67	02/15/2021::10:53:54	8.25-10.01
264	ALP	26471	109.04	02/15/2021::10:54:36	96.26-111.74
265	SGPT	26471	34.43	02/15/2021::10:54:49	31.10-35.86
266	ALB	26471	3.94	02/15/2021::10:55:12	3.75-4.35
267	DBIL	26471	0.39	02/15/2021::10:55:34	0.33-0.49
268	GGT	26471	63.28	02/15/2021::10:55:47	59.00-64.60
269	CREAT	26471	2.21	02/15/2021::10:56:40	1.60-2.48
270	UREA	26471	31.95	02/15/2021::10:56:53	29.38-36.30

### Qc results

From: 01/01/2021

To: 12/03/2021

Concentration: Normal

No.	Test name	QC lot	Result	Run date	QC range
271	SGOT	26471	42.94	02/15/2021::10:57:32	38.48-43.96
272	UA	26471	4.29	02/15/2021::10:59:50	3.82-5.06
273	TBIL	26471	1.20	02/15/2021::11:01:08	0.78-1.46
274	TP	26471	6.79	02/15/2021::11:01:22	5.93-6.89
275	GLUC	26471	84.99	02/15/2021::11:01:35	81.69-90.21
276	CHOL	26471	250.82	02/15/2021::11:02:29	247.81-266.89
277	TG	26471	194.52	02/15/2021::11:04:13	192.95-201.43
278	HDLC	26471	56.85	02/15/2021::11:05:05	45.37-71.97
279	CAL	26471	8.94	02/17/2021::13:00:50	8.25-10.01
280	ALP	26471	112.60	02/17/2021::13:01:32	96.26-111.74
281	SGPT	26471	35.53	02/17/2021::13:01:45	31.10-35.86
282	ALB	26471	3.86	02/17/2021::13:02:08	3.75-4.35
283	DBIL	26471	0.45	02/17/2021::13:02:31	0.33-0.49
284	GGT	26471	63.59	02/17/2021::13:02:44	59.00-64.60
285	CREAT	26471	2.15	02/17/2021::13:03:36	1.60-2.48
286	UREA	26471	32.74	02/17/2021::13:03:49	29.38-36.30
287	SGOT	26471	46.33	02/17/2021::13:04:28	38.48-43.96
288	UA	26471	4.50	02/17/2021::13:06:46	3.82-5.06
289	TBIL	26471	1.19	02/17/2021::13:08:05	0.78-1.46
290	TP	26471	6.69	02/17/2021::13:08:19	5.93-6.89
291	GLUC	26471	85.32	02/17/2021::13:08:31	81.69-90.21
292	CHOL	26471	251.68	02/17/2021::13:09:25	247.81-266.89
293	TG	26471	197.23	02/17/2021::13:11:24	192.95-201.43
294	HDLC	26471	49.77	02/17/2021::13:12:03	45.37-71.97
295	CAL	26471	8.91	02/19/2021::11:05:34	8.25-10.01
296	ALP	26471	106.91	02/19/2021::11:06:16	96.26-111.74
297	SGPT	26471	33.38	02/19/2021::11:06:29	31.10-35.86
298	ALB	26471	3.87	02/19/2021::11:06:52	3.75-4.35
299	DBIL	26471	0.43	02/19/2021::11:07:15	0.33-0.49
300	GGT	26471	60.79	02/19/2021::11:07:28	59.00-64.60
301	CREAT	26471	1.77	02/19/2021::11:08:20	1.60-2.48
302	UREA	26471	30.21	02/19/2021::11:08:33	29.38-36.30
303	SGOT	26471	44.35	02/19/2021::11:09:12	38.48-43.96
304	UA	26471	4.48	02/19/2021::11:11:30	3.82-5.06
305	TBIL	26471	1.15	02/19/2021::11:12:49	0.78-1.46
306	TP	26471	6.45	02/19/2021::11:13:02	5.93-6.89
307	GLUC	26471	85.90	02/19/2021::11:13:15	81.69-90.21
308	CHOL	26471	257.21	02/19/2021::11:14:09	247.81-266.89
309	TG	26471	198.62	02/19/2021::11:15:53	192.95-201.43
310	HDLC	26471	53.31	02/19/2021::11:16:46	45.37-71.97
311	CAL	26471	8.87	02/22/2021::10:35:05	8.25-10.01
312	ALP	26471	101.23	02/22/2021::10:35:47	96.26-111.74
313	SGPT	26471	34.61	02/22/2021::10:36:00	31.10-35.86
314	ALB	26471	3.89	02/22/2021::10:36:23	3.75-4.35
315	GGT	26471	62.66	02/22/2021::10:36:59	59.00-64.60
316	DBIL	26471	0.41	02/22/2021::10:37:12	0.33-0.49
317	CREAT	26471	2.16	02/22/2021::10:37:51	1.60-2.48
318	UREA	26471	32.89	02/22/2021::10:38:04	29.38-36.30
319	SGOT	26471	42.66	02/22/2021::10:38:44	38.48-43.96
320	UA	26471	4.29	02/22/2021::10:41:01	3.82-5.06
321	TBIL	26471	1.15	02/22/2021::10:42:20	0.78-1.46
322	TP	26471	6.29	02/22/2021::10:42:33	5.93-6.89
323	GLUC	26471	83.78	02/22/2021::10:42:47	81.69-90.21
324	CHOL	26471	258.41	02/22/2021::10:43:40	247.81-266.89

## Qc results

From: 01/01/2021 To: 12/03/2021 Concentration: Normal

No.	Test name	QC lot	Result	Run date	QC range
325	TG	26471	199.82	02/22/2021::10:45:12	192.95-201.43
326	HDLC	26471	48.62	02/22/2021::10:46:17	45.37-71.97
327	LDLC	26471	139.32	02/22/2021::10:46:17	125.50-144.62
328	CAL	26471	8.94	02/24/2021::11:16:41	8.25-10.01
329	ALP	26471	97.68	02/24/2021::11:17:24	96.26-111.74
330	SGPT	26471	34.91	02/24/2021::11:17:37	31.10-35.86
331	ALB	26471	3.75	02/24/2021::11:18:00	3.75-4.35
332	DBIL	26471	0.43	02/24/2021::11:18:22	0.33-0.49
333	GGT	26471	63.59	02/24/2021::11:18:35	59.00-64.60
334	CREAT	26471	2.11	02/24/2021::11:19:28	1.60-2.48
335	UREA	26471	32.41	02/24/2021::11:19:41	29.38-36.30
336	SGOT	26471	39.83	02/24/2021::11:20:20	38.48-43.96
337	UA	26471	4.39	02/24/2021::11:22:37	3.82-5.06
338	TBIL	26471	1.19	02/24/2021::11:23:56	0.78-1.46
339	TP	26471	6.58	02/24/2021::11:24:09	5.93-6.89
340	GLUC	26471	87.20	02/24/2021::11:24:22	81.69-90.21
341	CHOL	26471	259.69	02/24/2021::11:25:16	247.81-266.89
342	TG	26471	198.54	02/24/2021::11:27:01	192.95-201.43
343	HDLC	26471	63.08	02/24/2021::11:27:53	45.37-71.97
344	LDLC	26471	140.85	02/24/2021::11:27:53	125.50-144.62
345	CAL	26471	9.02	02/26/2021::10:53:45	8.25-10.01
346	ALP	26471	94.13	02/26/2021::10:54:28	96.26-111.74
347	SGPT	26471	32.46	02/26/2021::10:54:40	31.10-35.86
348	ALB	26471	3.69	02/26/2021::10:55:04	3.75-4.35
349	DBIL	26471	0.44	02/26/2021::10:55:26	0.33-0.49
350	GGT	26471	61.69	02/26/2021::10:55:39	59.00-64.60
351	CREAT	26471	1.99	02/26/2021::10:56:31	1.60-2.48
352	UREA	26471	30.42	02/26/2021::10:56:32	29.38-36.30
353	SGOT	26471	40.96	02/26/2021::10:57:37	38.48-43.96
354	UA	26471	4.16	02/26/2021::10:59:42	3.82-5.06
355	TBIL	26471	1.16	02/26/2021::11:01:00	0.78-1.46
356	TP	26471	6.39	02/26/2021::11:01:14	5.93-6.89
357	GLUC	26471	86.23	02/26/2021::11:01:27	81.69-90.21
358	CHOL	26471	265.23	02/26/2021::11:02:21	247.81-266.89
359	TG	26471	198.23	02/26/2021::11:04:05	192.95-201.43
360	HDLC	26471	54.47	02/26/2021::11:04:57	45.37-71.97
361	LDLC	26471	142.70	02/26/2021::11:04:57	125.50-144.62
362	CAL	26471	9.43	03/02/2021::10:18:28	8.25-10.01
363	ALP	26471	108.69	03/02/2021::10:19:10	96.26-111.74
364	SGPT	26471	32.50	03/02/2021::10:19:23	31.10-35.86
365	ALB	26471	4.44	03/02/2021::10:19:46	3.75-4.35
366	DBIL	26471	0.41	03/02/2021::10:20:08	0.33-0.49
367	GGT	26471	64.52	03/02/2021::10:20:21	59.00-64.60
368	CREAT	26471	2.18	03/02/2021::10:21:14	1.60-2.48
369	UREA	26471	34.54	03/02/2021::10:21:27	29.38-36.30
370	SGOT	26471	45.65	03/02/2021::10:22:06	38.48-43.96
371	UA	26471	4.91	03/02/2021::10:24:24	3.82-5.06
372	TBIL	26471	1.15	03/02/2021::10:25:43	0.78-1.46
373	TP	26471	6.90	03/02/2021::10:25:56	5.93-6.89
374	GLUC	26471	90.45	03/02/2021::10:26:09	81.69-90.21
375	CHOL	26471	268.13	03/02/2021::10:27:03	247.81-266.89
376	TG	26471	202.35	03/02/2021::10:28:48	192.95-201.43
377	HDLC	26471	57.77	03/02/2021::10:29:40	45.37-71.97
378	LDLC	26471	141.27	03/02/2021::10:29:40	125.50-144.62



### Qc results

From: 01/01/2021 To: 12/03/2021 Concentration: Normal

No.	Test name	QC lot	Result	Run date	QC range
379	CAL	26471	9.29	03/04/2021::10:19:29	8.25-10.01
380	ALP	26471	113.31	03/04/2021::10:20:12	96.26-111.74
381	SGPT	26471	33.69	03/04/2021::10:20:25	31.10-35.86
382	ALB	26471	3.99	03/04/2021::10:20:48	3.75-4.35
383	DBIL	26471	0.47	03/04/2021::10:21:10	0.33-0.49
384	GGT	26471	62.07	03/04/2021::10:21:24	59.00-64.60
385	CREAT	26471	2.25	03/04/2021::10:22:16	1.60-2.48
386	UREA	26471	33.54	03/04/2021::10:22:16	29.38-36.30
387	SGOT	26471	43.50	03/04/2021::10:23:08	38.48-43.96
388	UA	26471	4.19	03/04/2021::10:25:27	3.82-5.06
389	TBIL	26471	1.17	03/04/2021::10:26:46	0.78-1.46
390	TP	26471	6.44	03/04/2021::10:26:59	5.93-6.89
391	GLUC	26471	84.65	03/04/2021::10:27:12	81.69-90.21
392	CHOL	26471	250.29	03/04/2021::10:28:06	247.81-266.89
393	TG	26471	195.44	03/04/2021::10:29:50	192.95-201.43
394	HDLC	26471	45.22	03/04/2021::10:30:42	45.37-71.97
395	LDLC	26471	153.64	03/04/2021::10:30:42	125.50-144.62
396	CAL	26471	8.50	03/06/2021::11:24:52	8.25-10.01
397	ALP	26471	99.37	03/06/2021::11:25:34	96.26-111.74
398	SGPT	26471	33.21	03/06/2021::11:25:48	31.10-35.86
399	ALB	26471	3.84	03/06/2021::11:26:11	3.75-4.35
400	DBIL	26471	0.48	03/06/2021::11:26:33	0.33-0.49
401	GGT	26471	62.33	03/06/2021::11:26:46	59.00-64.60
402	CREAT	26471	1.96	03/06/2021::11:27:38	1.60-2.48
403	UREA	26471	31.93	03/06/2021::11:27:51	29.38-36.30
404	SGOT	26471	43.12	03/06/2021::11:28:30	38.48-43.96
405	UA	26471	4.81	03/06/2021::11:30:48	3.82-5.06
406	TBIL	26471	1.02	03/06/2021::11:32:07	0.78-1.46
407	TP	26471	6.05	03/06/2021::11:32:20	5.93-6.89
408	GLUC	26471	82.91	03/06/2021::11:32:33	81.69-90.21
409	CHOL	26471	253.56	03/06/2021::11:33:27	247.81-266.89
410	TG	26471	196.82	03/06/2021::11:35:12	192.95-201.43
411	HDLC	26471	48.09	03/06/2021::11:36:04	45.37-71.97
412	LDLC	26471	130.08	03/06/2021::11:36:04	125.50-144.62
413	CAL	26471	8.98	03/09/2021::11:02:21	8.25-10.01
414	ALP	26471	101.17	03/09/2021::11:03:03	96.26-111.74
415	SGPT	26471	32.15	03/09/2021::11:03:16	31.10-35.86
416	ALB	26471	3.82	03/09/2021::11:03:39	3.75-4.35
417	DBIL	26471	0.45	03/09/2021::11:04:01	0.33-0.49
418	GGT	26471	63.00	03/09/2021::11:04:14	59.00-64.60
419	CREAT	26471	2.06	03/09/2021::11:05:07	1.60-2.48
420	UREA	26471	33.98	03/09/2021::11:05:07	29.38-36.30
421	SGOT	26471	39.89	03/09/2021::11:06:12	38.48-43.96
422	UA	26471	4.96	03/09/2021::11:08:17	3.82-5.06
423	TBIL	26471	1.16	03/09/2021::11:09:36	0.78-1.46
424	TP	26471	6.18	03/09/2021::11:09:49	5.93-6.89
425	GLUC	26471	83.09	03/09/2021::11:10:02	81.69-90.21
426	CHOL	26471	256.97	03/09/2021::11:10:57	247.81-266.89
427	TG	26471	198.41	03/09/2021::11:12:41	192.95-201.43
428	HDLC	26471	58.09	03/09/2021::11:13:33	45.37-71.97
429	LDLC	26471	132.48	03/09/2021::11:13:33	125.50-144.62
430	CAL	26471	8.91	03/12/2021::09:08:43	8.25-10.01
431	ALP	26471	104.37	03/12/2021::09:09:18	96.26-111.74
432	SGPT	26471	31.52	03/12/2021::09:09:37	31.10-35.86

### Qc results

From: 01/01/2021

To: 12/03/2021

Concentration: Normal

No.	Test name	QC lot	Result	Run date	QC range
433	ALB	26471	3.75	03/12/2021::09:10:12	3.75-4.35
434	DBIL	26471	0.44	03/12/2021::09:10:36	0.33-0.49
435	GGT	26471	60.07	03/12/2021::09:10:36	59.00-64.60
436	CREAT	26471	2.03	03/12/2021::09:11:29	1.60-2.48
437	UREA	26471	35.70	03/12/2021::09:11:29	29.38-36.30
438	SGOT	26471	38.78	03/12/2021::09:12:30	38.48-43.96
439	UA	26471	3.95	03/12/2021::09:14:33	3.82-5.06
440	TP	26471	6.30	03/12/2021::09:16:04	5.93-6.89
441	GLUC	26471	88.97	03/12/2021::09:16:26	81.69-90.21
442	TBIL	26471	1.15	03/12/2021::09:16:27	0.78-1.46
443	CHOL	26471	258.04	03/12/2021::09:17:14	247.81-266.89
444	TG	26471	196.60	03/12/2021::09:18:59	192.95-201.43
445	HDLC	26471	66.34	03/12/2021::09:20:14	45.37-71.97
446	LDLC	26471	139.86	03/12/2021::09:20:25	125.50-144.62

### Qc results

From: 01/01/2021

To: 13/03/2021

Concentration: High

No.	Test name	QC lot	Result	Run date	QC range
	ALB	26472	2.65	01/05/2021::17:13:49	2.55-2.99
	GGT	26472	144.98	01/05/2021::17:14:54	136.01-149.73
	TBIL	26472	4.32	01/05/2021::17:20:09	3.98-4.70
	TP	26472	4.22	01/05/2021::17:20:09	3.51-4.59
	GLUC	26472	282.80	01/05/2021::17:20:30	272.97-287.53
	CHOL	26472	100.79	01/05/2021::17:21:21	97.37-102.77
	CAL	26472	11.09	01/05/2021::17:42:22	9.65-12.89
	ALP	26472	398.80	01/05/2021::17:42:57	381.62-419.94
	SGPT	26472	110.09	01/05/2021::17:43:16	100.63-113.23
0	DBIL	26472	1.53	01/05/2021::17:44:00	1.45-1.77
1	GGT	26472	144.36	01/05/2021::17:44:28	136.01-149.73
2	CREAT	26472	4.94	01/05/2021::17:45:09	4.30-5.66
3	UREA	26472	100.03	01/05/2021::17:45:09	82.60-104.52
4	SGOT	26472	193.23	01/05/2021::17:46:01	177.94-201.54
5	UA	26472	9.50	01/05/2021::17:48:10	7.49-10.45
6	TP	26472	4.28	01/05/2021::17:49:50	3.51-4.59
7	TBIL	26472	4.79	01/05/2021::17:49:52	3.98-4.70
8	GLUC	26472	279.81	01/05/2021::17:50:03	272.97-287.53
9	CHOL	26472	100.64	01/05/2021::17:50:52	97.37-102.77
10	LDLC	26472	64.16	01/05/2021::17:53:44	61.20-69.96
11	HDLC	26472	27.47	01/05/2021::17:53:57	17.38-31.46
12	CAL	26472	10.70	01/06/2021::10:46:09	9.65-12.89
13	SGPT	26472	105.65	01/06/2021::10:47:04	100.63-113.23
14	DBIL	26472	1.64	01/06/2021::10:47:50	1.45-1.77
15	GGT	26472	139.41	01/06/2021::10:48:03	136.01-149.73
16	CREAT	26472	5.12	01/06/2021::10:48:55	4.30-5.66
17	UREA	26472	93.65	01/06/2021::10:49:08	82.60-104.52
18	SGOT	26472	182.71	01/06/2021::10:49:47	177.94-201.54
19	UA	26472	9.35	01/06/2021::10:52:05	7.49-10.45
20	TBIL	26472	4.18	01/06/2021::10:53:24	3.98-4.70
21	TP	26472	4.17	01/06/2021::10:53:37	3.51-4.59
22	GLUC	26472	283.78	01/06/2021::10:53:50	272.97-287.53
23	CHOL	26472	101.34	01/06/2021::10:54:44	97.37-102.77
24	HDLC	26472	19.78	01/06/2021::10:57:20	17.38-31.46
25	LDLC	26472	62.89	01/06/2021::10:57:21	61.20-69.96
26	CAL	26472	11.79	01/08/2021::10:59:18	9.65-12.89
27	ALP	26472	405.98	01/08/2021::11:00:00	381.62-419.94
28	SGPT	26472	111.24	01/08/2021::11:00:13	100.63-113.23
29	ALB	26472	2.93	01/08/2021::11:00:36	2.55-2.99
30	DBIL	26472	1.57	01/08/2021::11:00:58	1.45-1.77
31	GGT	26472	143.30	01/08/2021::11:01:11	136.01-149.73
32	UREA	26472	91.47	01/08/2021::11:02:17	82.60-104.52
33	SGOT	26472	187.80	01/08/2021::11:02:56	177.94-201.54
34	UA	26472	10.20	01/08/2021::11:05:14	7.49-10.45
35	TBIL	26472	4.33	01/08/2021::11:06:32	3.98-4.70
36	TP	26472	4.23	01/08/2021::11:06:46	3.51-4.59
37	GLUC	26472	282.57	01/08/2021::11:06:59	272.97-287.53
38	CHOL	26472	100.70	01/08/2021::11:07:53	97.37-102.77
39	TG	26472	89.16	01/08/2021::11:09:37	83.94-96.86
40	HDLC	26472	30.69	01/08/2021::11:10:29	17.38-31.46
41	LDLC	26472	67.22	01/08/2021::11:10:29	61.20-69.96
42	CAL	26472	12.44	01/11/2021::11:27:07	9.65-12.89
43	ALP	26472	404.92	01/11/2021::11:27:49	381.62-419.94
44	SGPT	26472	108.89	01/11/2021::11:28:02	100.63-113.23

### Qc results

From: 01/01/2021 To: 13/03/2021 Concentration: High

No.	Test name	QC lot	Result	Run date	QC range
5	ALB	26472	2.90	01/11/2021::11:28:25	2.55-2.99
6	DBIL	26472	1.61	01/11/2021::11:28:47	1.45-1.77
7	GGT	26472	144.85	01/11/2021::11:29:00	136.01-149.73
8	UREA	26472	92.11	01/11/2021::11:29:52	82.60-104.52
9	SGOT	26472	189.39	01/11/2021::11:30:57	177.94-201.54
10	UA	26472	9.86	01/11/2021::11:33:02	7.49-10.45
11	TBIL	26472	4.34	01/11/2021::11:34:21	3.98-4.70
12	TP	26472	4.22	01/11/2021::11:34:34	3.51-4.59
13	GLUC	26472	279.92	01/11/2021::11:34:47	272.97-287.53
14	CHOL	26472	99.24	01/11/2021::11:35:41	97.37-102.77
15	TG	26472	88.49	01/11/2021::11:37:25	83.94-96.86
16	HDLC	26472	28.65	01/11/2021::11:38:17	17.38-31.46
17	LDLC	26472	70.59	01/11/2021::11:38:18	61.20-69.96
18	CAL	26472	12.94	01/13/2021::11:13:19	9.65-12.89
19	ALP	26472	389.29	01/13/2021::11:14:01	381.62-419.94
20	SGPT	26472	108.01	01/13/2021::11:14:14	100.63-113.23
21	ALB	26472	2.89	01/13/2021::11:14:38	2.55-2.99
22	DBIL	26472	1.63	01/13/2021::11:15:00	1.45-1.77
23	GGT	26472	140.51	01/13/2021::11:15:13	136.01-149.73
24	UREA	26472	90.50	01/13/2021::11:16:19	82.60-104.52
25	UA	26472	8.45	01/13/2021::11:19:16	7.49-10.45
26	TBIL	26472	4.26	01/13/2021::11:20:34	3.98-4.70
27	TP	26472	4.26	01/13/2021::11:20:48	3.51-4.59
28	GLUC	26472	278.58	01/13/2021::11:21:01	272.97-287.53
29	CHOL	26472	98.54	01/13/2021::11:21:55	97.37-102.77
30	TG	26472	88.98	01/13/2021::11:23:40	83.94-96.86
31	HDLC	26472	25.12	01/13/2021::11:24:32	17.38-31.46
32	LDLC	26472	66.36	01/13/2021::11:24:33	61.20-69.96
33	SGOT	26472	188.61	01/13/2021::12:21:46	177.94-201.54
34	CREAT	26472	4.90	01/13/2021::12:22:51	4.30-5.66
35	CAL	26472	11.32	01/15/2021::10:45:43	9.65-12.89
36	ALP	26472	384.67	01/15/2021::10:46:26	381.62-419.94
37	SGPT	26472	106.53	01/15/2021::10:46:39	100.63-113.23
38	ALB	26472	2.76	01/15/2021::10:47:02	2.55-2.99
39	DBIL	26472	1.72	01/15/2021::10:47:24	1.45-1.77
40	GGT	26472	141.44	01/15/2021::10:47:37	136.01-149.73
41	CREAT	26472	4.94	01/15/2021::10:48:30	4.30-5.66
42	UREA	26472	100.40	01/15/2021::10:48:30	82.60-104.52
43	SGOT	26472	187.22	01/15/2021::10:49:22	177.94-201.54
44	UA	26472	8.55	01/15/2021::10:51:40	7.49-10.45
45	TBIL	26472	4.22	01/15/2021::10:52:59	3.98-4.70
46	TP	26472	4.26	01/15/2021::10:53:12	3.51-4.59
47	GLUC	26472	275.72	01/15/2021::10:53:25	272.97-287.53
48	CHOL	26472	97.14	01/15/2021::10:54:19	97.37-102.77
49	TG	26472	91.21	01/15/2021::10:56:03	83.94-96.86
50	HDLC	26472	22.81	01/15/2021::10:56:55	17.38-31.46
51	LDLC	26472	66.12	01/15/2021::10:56:56	61.20-69.96
52	CAL	26472	10.68	01/18/2021::12:37:49	9.65-12.89
53	ALP	26472	406.34	01/18/2021::12:38:31	381.62-419.94
54	SGPT	26472	106.53	01/18/2021::12:38:44	100.63-113.23
55	ALB	26472	2.80	01/18/2021::12:39:07	2.55-2.99
56	DBIL	26472	1.60	01/18/2021::12:39:29	1.45-1.77
57	GGT	26472	146.09	01/18/2021::12:39:42	136.01-149.73
58	CREAT	26472	4.63	01/18/2021::12:40:35	4.30-5.66

### Qc results

From: 01/01/2021

To: 13/03/2021

Concentration: High

No.	Test name	QC lot	Result	Run date	QC range
109	UREA	26472	98.97	01/18/2021::12:40:48	82.60-104.52
110	SGOT	26472	198.06	01/18/2021::12:41:27	177.94-201.54
111	UA	26472	8.35	01/18/2021::12:43:45	7.49-10.45
112	TBIL	26472	4.23	01/18/2021::12:45:03	3.98-4.70
113	TP	26472	4.10	01/18/2021::12:45:16	3.51-4.59
114	GLUC	26472	275.77	01/18/2021::12:45:29	272.97-287.53
115	CHOL	26472	100.75	01/18/2021::12:46:23	97.37-102.77
116	TG	26472	89.64	01/18/2021::12:48:08	83.94-96.86
117	LDLC	26472	66.22	01/18/2021::12:49:00	61.20-69.96
118	HDLC	26472	23.62	01/18/2021::12:49:13	17.38-31.46
119	CAL	26472	11.51	01/22/2021::11:10:51	9.65-12.89
120	ALP	26472	386.45	01/22/2021::11:11:33	381.62-419.94
121	SGPT	26472	98.88	01/22/2021::11:11:46	100.63-113.23
122	ALB	26472	2.71	01/22/2021::11:12:09	2.55-2.99
123	DBIL	26472	1.72	01/22/2021::11:12:31	1.45-1.77
124	GGT	26472	143.55	01/22/2021::11:12:45	136.01-149.73
125	CREAT	26472	4.84	01/22/2021::11:13:37	4.30-5.66
126	UREA	26472	82.24	01/22/2021::11:13:38	82.60-104.52
127	SGOT	26472	187.30	01/22/2021::11:14:29	177.94-201.54
128	UA	26472	8.40	01/22/2021::11:16:48	7.49-10.45
129	TBIL	26472	4.13	01/22/2021::11:18:06	3.98-4.70
130	TP	26472	3.67	01/22/2021::11:18:08	3.51-4.59
131	GLUC	26472	275.72	01/22/2021::11:18:33	272.97-287.53
132	CHOL	26472	100.05	01/22/2021::11:19:25	97.37-102.77
133	TG	26472	88.07	01/22/2021::11:21:09	83.94-96.86
134	LDLC	26472	62.19	01/22/2021::11:21:59	61.20-69.96
135	HDLC	26472	18.54	01/22/2021::11:22:12	17.38-31.46
136	SGOT	26472	178.33	01/22/2021::11:22:38	177.94-201.54
137	CAL	26472	11.27	01/25/2021::11:24:47	9.65-12.89
138	ALP	26472	412.38	01/25/2021::11:25:30	381.62-419.94
139	SGPT	26472	109.18	01/25/2021::11:25:43	100.63-113.23
140	ALB	26472	2.83	01/25/2021::11:26:06	2.55-2.99
141	DBIL	26472	1.49	01/25/2021::11:26:28	1.45-1.77
142	GGT	26472	142.99	01/25/2021::11:26:41	136.01-149.73
143	CREAT	26472	4.86	01/25/2021::11:27:34	4.30-5.66
144	UREA	26472	93.41	01/25/2021::11:27:34	82.60-104.52
145	SGOT	26472	190.83	01/25/2021::11:28:26	177.94-201.54
146	UA	26472	8.27	01/25/2021::11:30:45	7.49-10.45
147	TBIL	26472	4.18	01/25/2021::11:32:03	3.98-4.70
148	TP	26472	3.66	01/25/2021::11:32:16	3.51-4.59
149	GLUC	26472	286.75	01/25/2021::11:32:30	272.97-287.53
150	CHOL	26472	100.45	01/25/2021::11:33:23	97.37-102.77
151	TG	26472	88.80	01/25/2021::11:35:08	83.94-96.86
152	HDLC	26472	26.31	01/25/2021::11:36:00	17.38-31.46
153	LDLC	26472	65.99	01/25/2021::11:36:00	61.20-69.96
154	TP	26472	3.46	01/25/2021::11:48:45	3.51-4.59
155	CAL	26472	10.25	01/28/2021::10:46:14	9.65-12.89
156	ALP	26472	401.72	01/28/2021::10:46:56	381.62-419.94
157	SGPT	26472	105.65	01/28/2021::10:47:09	100.63-113.23
158	DBIL	26472	1.57	01/28/2021::10:47:55	1.45-1.77
159	GGT	26472	147.33	01/28/2021::10:48:08	136.01-149.73
160	CREAT	26472	4.80	01/28/2021::10:49:01	4.30-5.66
161	UREA	26472	86.66	01/28/2021::10:49:01	82.60-104.52
162	SGOT	26472	195.20	01/28/2021::10:49:53	177.94-201.54

### Qc results

From: 01/01/2021 To: 13/03/2021 Concentration: High

No.	Test name	QC lot	Result	Run date	QC range
63	UA	26472	9.98	01/28/2021::10:52:11	7.49-10.45
64	TBIL	26472	4.38	01/28/2021::10:53:30	3.98-4.70
65	GLUC	26472	281.05	01/28/2021::10:53:56	272.97-287.53
66	CHOL	26472	99.29	01/28/2021::10:54:50	97.37-102.77
67	TG	26472	92.59	01/28/2021::10:56:34	83.94-96.86
68	HDLC	26472	24.77	01/28/2021::10:57:26	17.38-31.46
69	LDLC	26472	65.38	01/28/2021::10:57:26	61.20-69.96
70	ALB	26472	2.75	01/28/2021::11:40:09	2.55-2.99
71	SGOT	26472	195.20	01/28/2021::11:41:53	177.94-201.54
72	UA	26472	8.10	01/28/2021::11:44:17	7.49-10.45
73	TP	26472	4.28	01/28/2021::11:48:11	3.51-4.59
74	CAL	26472	10.91	01/30/2021::11:22:49	9.65-12.89
75	ALB	26472	2.78	01/30/2021::11:23:05	2.55-2.99
76	ALP	26472	404.21	01/30/2021::11:23:26	381.62-419.94
77	SGPT	26472	105.65	01/30/2021::11:23:39	100.63-113.23
78	DBIL	26472	1.57	01/30/2021::11:24:53	1.45-1.77
79	CREAT	26472	4.87	01/30/2021::11:25:32	4.30-5.66
80	UREA	26472	97.79	01/30/2021::11:25:45	82.60-104.52
81	SGOT	26472	198.31	01/30/2021::11:26:37	177.94-201.54
82	UA	26472	8.41	01/30/2021::11:28:43	7.49-10.45
83	TBIL	26472	4.26	01/30/2021::11:29:22	3.98-4.70
84	TP	26472	3.80	01/30/2021::11:30:02	3.51-4.59
85	GLUC	26472	283.45	01/30/2021::11:30:30	272.97-287.53
86	CHOL	26472	102.31	01/30/2021::11:31:22	97.37-102.77
87	TG	26472	90.42	01/30/2021::11:33:06	83.94-96.86
88	HDLC	26472	22.31	01/30/2021::11:33:58	17.38-31.46
89	LDLC	26472	65.97	01/30/2021::11:33:58	61.20-69.96
90	CAL	26472	11.55	02/02/2021::10:58:54	9.65-12.89
91	ALP	26472	410.25	02/02/2021::10:59:36	381.62-419.94
92	SGPT	26472	107.42	02/02/2021::10:59:49	100.63-113.23
93	ALB	26472	2.78	02/02/2021::11:00:12	2.55-2.99
94	DBIL	26472	1.51	02/02/2021::11:00:49	1.45-1.77
95	GGT	26472	147.95	02/02/2021::11:00:49	136.01-149.73
96	CREAT	26472	4.94	02/02/2021::11:01:38	4.30-5.66
97	UREA	26472	99.09	02/02/2021::11:01:51	82.60-104.52
98	SGOT	26472	196.05	02/02/2021::11:02:30	177.94-201.54
99	UA	26472	8.61	02/02/2021::11:04:39	7.49-10.45
200	TBIL	26472	4.31	02/02/2021::11:06:07	3.98-4.70
201	TP	26472	4.26	02/02/2021::11:06:20	3.51-4.59
202	GLUC	26472	282.57	02/02/2021::11:06:34	272.97-287.53
203	CHOL	26472	101.50	02/02/2021::11:07:28	97.37-102.77
204	TG	26472	89.40	02/02/2021::11:09:12	83.94-96.86
205	HDLC	26472	28.96	02/02/2021::11:10:04	17.38-31.46
206	LDLC	26472	67.44	02/02/2021::11:10:30	61.20-69.96
207	CAL	26472	11.31	02/04/2021::10:32:19	9.65-12.89
208	ALP	26472	405.27	02/04/2021::10:33:01	381.62-419.94
209	SGPT	26472	105.65	02/04/2021::10:33:14	100.63-113.23
210	ALB	26472	2.78	02/04/2021::10:33:37	2.55-2.99
211	DBIL	26472	1.56	02/04/2021::10:33:59	1.45-1.77
212	GGT	26472	148.88	02/04/2021::10:34:12	136.01-149.73
213	CREAT	26472	5.12	02/04/2021::10:35:05	4.30-5.66
214	UREA	26472	92.01	02/04/2021::10:35:18	82.60-104.52
215	SGOT	26472	191.81	02/04/2021::10:35:57	177.94-201.54
216	UA	26472	8.33	02/04/2021::10:38:15	7.49-10.45

### Qc results

From: 01/01/2021

To: 13/03/2021

Concentration: High

No.	Test name	QC lot	Result	Run date	QC range
217	TBIL	26472	4.31	02/04/2021::10:39:34	
218	TP	26472	4.12	02/04/2021::10:39:47	3.98-4.70
219	GLUC	26472	277.24	02/04/2021::10:40:00	3.51-4.59
220	CHOL	26472	100.25	02/04/2021::10:40:54	272.97-287.53
221	TG	26472	85.90	02/04/2021::10:42:38	97.37-102.77
222	HDLC	26472	27.08	02/04/2021::10:43:30	83.94-96.86
223	LDLC	26472	63.67	02/04/2021::10:43:31	17.38-31.46
224	CAL	26472	11.26	02/06/2021::11:09:19	61.20-69.96
225	ALP	26472	396.39	02/06/2021::11:10:02	9.65-12.89
226	SGPT	26472	104.18	02/06/2021::11:10:15	381.62-419.94
227	ALB	26472	2.80	02/06/2021::11:10:38	100.63-113.23
228	DBIL	26472	1.57	02/06/2021::11:11:00	2.55-2.99
229	GGT	26472	145.16	02/06/2021::11:11:13	1.45-1.77
230	CREAT	26472	4.98	02/06/2021::11:12:06	136.01-149.73
231	UREA	26472	91.88	02/06/2021::11:12:19	4.30-5.66
232	SGOT	26472	195.20	02/06/2021::11:12:58	82.60-104.52
233	UA	26472	8.94	02/06/2021::11:15:16	177.94-201.54
234	TBIL	26472	4.31	02/06/2021::11:16:35	7.49-10.45
235	TP	26472	4.51	02/06/2021::11:16:48	3.98-4.70
236	GLUC	26472	276.19	02/06/2021::11:17:01	3.51-4.59
237	CHOL	26472	100.50	02/06/2021::11:17:55	272.97-287.53
238	TG	26472	89.40	02/06/2021::11:19:39	97.37-102.77
239	HDLC	26472	27.27	02/06/2021::11:20:31	83.94-96.86
240	LDLC	26472	64.77	02/06/2021::11:20:31	17.38-31.46
241	CAL	26472	11.13	02/09/2021::10:35:08	61.20-69.96
242	ALP	26472	397.46	02/09/2021::10:35:50	9.65-12.89
243	SGPT	26472	105.95	02/09/2021::10:36:03	381.62-419.94
244	ALB	26472	2.78	02/09/2021::10:36:26	100.63-113.23
245	GGT	26472	147.64	02/09/2021::10:37:01	2.55-2.99
246	DBIL	26472	1.46	02/09/2021::10:37:15	136.01-149.73
247	CREAT	26472	5.16	02/09/2021::10:37:54	1.45-1.77
248	UREA	26472	89.42	02/09/2021::10:38:07	4.30-5.66
249	SGOT	26472	194.07	02/09/2021::10:38:46	82.60-104.52
250	UA	26472	8.99	02/09/2021::10:41:05	177.94-201.54
251	TBIL	26472	4.27	02/09/2021::10:42:23	7.49-10.45
252	TP	26472	4.31	02/09/2021::10:42:37	3.98-4.70
253	GLUC	26472	278.35	02/09/2021::10:42:50	3.51-4.59
254	CHOL	26472	98.14	02/09/2021::10:43:44	272.97-287.53
255	TG	26472	92.05	02/09/2021::10:45:15	97.37-102.77
256	HDLC	26472	24.77	02/09/2021::10:46:20	83.94-96.86
257	LDLC	26472	68.53	02/09/2021::10:46:20	17.38-31.46
258	CREAT	26472	5.16	02/09/2021::11:28:25	61.20-69.96
259	TP	26472	4.31	02/09/2021::11:33:51	4.30-5.66
260	CAL	26472	10.67	02/11/2021::11:21:40	3.51-4.59
261	ALP	26472	402.08	02/11/2021::11:22:22	9.65-12.89
262	SGPT	26472	103.00	02/11/2021::11:22:35	381.62-419.94
263	ALB	26472	2.74	02/11/2021::11:22:58	100.63-113.23
264	DBIL	26472	1.61	02/11/2021::11:23:21	2.55-2.99
265	GGT	26472	141.75	02/11/2021::11:23:34	1.45-1.77
266	CREAT	26472	5.06	02/11/2021::11:24:26	136.01-149.73
267	UREA	26472	94.36	02/11/2021::11:24:39	4.30-5.66
268	SGOT	26472	187.29	02/11/2021::11:25:18	82.60-104.52
269	UA	26472	8.67	02/11/2021::11:27:37	177.94-201.54
270	TBIL	26472	4.30	02/11/2021::11:28:56	7.49-10.45
					3.98-4.70

### Qc results

From: 01/01/2021 To: 13/03/2021 Concentration: High

No.	Test name	QC lot	Result	Run date	QC range
271	TP	26472	4.30	02/11/2021::11:29:09	3.51-4.59
272	GLUC	26472	277.71	02/11/2021::11:29:22	272.97-287.53
273	CHOL	26472	97.34	02/11/2021::11:30:16	97.37-102.77
274	TG	26472	88.61	02/11/2021::11:32:01	83.94-96.86
275	HDLC	26472	24.81	02/11/2021::11:32:53	17.38-31.46
276	LDLC	26472	66.85	02/11/2021::11:32:53	61.20-69.96
277	CAL	26472	10.87	02/13/2021::11:04:25	9.65-12.89
278	ALP	26472	416.64	02/13/2021::11:05:07	381.62-419.94
279	SGPT	26472	108.30	02/13/2021::11:05:21	100.63-113.23
280	ALB	26472	2.77	02/13/2021::11:05:44	2.55-2.99
281	DBIL	26472	1.49	02/13/2021::11:06:06	1.45-1.77
282	GGT	26472	148.88	02/13/2021::11:06:19	136.01-149.73
283	CREAT	26472	5.11	02/13/2021::11:07:12	4.30-5.66
284	UREA	26472	90.35	02/13/2021::11:07:25	82.60-104.52
285	SGOT	26472	201.70	02/13/2021::11:08:04	177.94-201.54
286	UA	26472	8.90	02/13/2021::11:10:22	7.49-10.45
287	TBIL	26472	4.28	02/13/2021::11:11:41	3.98-4.70
288	TP	26472	4.35	02/13/2021::11:11:54	3.51-4.59
289	GLUC	26472	278.09	02/13/2021::11:12:07	272.97-287.53
290	CHOL	26472	98.49	02/13/2021::11:13:01	97.37-102.77
291	TG	26472	87.77	02/13/2021::11:14:45	83.94-96.86
292	HDLC	26472	24.81	02/13/2021::11:15:37	17.38-31.46
293	LDLC	26472	67.18	02/13/2021::11:15:38	61.20-69.96
294	CAL	26472	11.14	02/16/2021::11:25:36	9.65-12.89
295	ALP	26472	419.02	02/16/2021::11:26:21	381.62-419.94
296	ALB	26472	2.66	02/16/2021::11:26:56	2.55-2.99
297	GGT	26472	144.54	02/16/2021::11:27:45	136.01-149.73
298	DBIL	26472	1.55	02/16/2021::11:27:59	1.45-1.77
299	CREAT	26472	5.08	02/16/2021::11:28:25	4.30-5.66
300	UREA	26472	92.65	02/16/2021::11:29:05	82.60-104.52
301	UA	26472	8.67	02/16/2021::11:31:27	7.49-10.45
302	TBIL	26472	4.33	02/16/2021::11:32:18	3.98-4.70
303	TP	26472	4.38	02/16/2021::11:32:57	3.51-4.59
304	GLUC	26472	276.96	02/16/2021::11:33:25	272.97-287.53
305	CHOL	26472	98.29	02/16/2021::11:34:04	97.37-102.77
306	TG	26472	85.72	02/16/2021::11:35:49	83.94-96.86
307	HDLC	26472	23.42	02/16/2021::11:37:07	17.38-31.46
308	LDLC	26472	67.57	02/16/2021::11:37:20	61.20-69.96
309	DBIL	26472	1.53	02/16/2021::12:01:20	1.45-1.77
310	SGPT	26472	112.70	02/16/2021::12:03:17	100.63-113.23
311	SGOT	26472	196.40	02/16/2021::12:03:30	177.94-201.54
312	PHOS	26472	0.06	02/16/2021::12:06:07	6.49-8.01
313	PHOS	26472	6.44	02/16/2021::12:18:39	6.49-8.01
314	CAL	26472	11.53	02/18/2021::10:44:44	9.65-12.89
315	ALP	26472	418.61	02/18/2021::10:45:26	381.62-419.94
316	SGPT	26472	113.72	02/18/2021::10:45:39	100.63-113.23
317	ALB	26472	2.85	02/18/2021::10:46:02	2.55-2.99
318	DBIL	26472	1.48	02/18/2021::10:46:24	1.45-1.77
319	GGT	26472	146.92	02/18/2021::10:46:37	136.01-149.73
320	CREAT	26472	4.79	02/18/2021::10:47:30	4.30-5.66
321	UREA	26472	95.61	02/18/2021::10:47:43	82.60-104.52
322	SGOT	26472	200.17	02/18/2021::10:48:22	177.94-201.54
323	UA	26472	8.71	02/18/2021::10:50:40	7.49-10.45
324	TBIL	26472	4.30	02/18/2021::10:51:58	3.98-4.70



### Qc results

From: 01/01/2021 To: 13/03/2021 Concentration: High

No.	Test name	QC lot	Result	Run date	QC range
325	TP	26472	4.70	02/18/2021::10:52:12	3.51-4.59
326	GLUC	26472	277.37	02/18/2021::10:52:25	272.97-287.53
327	CHOL	26472	100.45	02/18/2021::10:53:19	97.37-102.77
328	TG	26472	92.17	02/18/2021::10:55:03	83.94-96.86
329	HDLC	26472	24.88	02/18/2021::10:55:55	17.38-31.46
330	LDLC	26472	68.40	02/18/2021::10:55:55	61.20-69.96
331	CAL	26472	11.28	02/20/2021::11:32:15	9.65-12.89
332	ALP	26472	406.69	02/20/2021::11:32:57	381.62-419.94
333	SGPT	26472	106.68	02/20/2021::11:33:10	100.63-113.23
334	ALB	26472	2.97	02/20/2021::11:33:33	2.55-2.99
335	DBIL	26472	1.67	02/20/2021::11:34:05	1.45-1.77
336	GGT	26472	148.88	02/20/2021::11:34:05	136.01-149.73
337	CREAT	26472	5.12	02/20/2021::11:34:59	4.30-5.66
338	UREA	26472	87.99	02/20/2021::11:35:15	82.60-104.52
339	SGOT	26472	195.63	02/20/2021::11:35:55	177.94-201.54
340	UA	26472	8.47	02/20/2021::11:38:05	7.49-10.45
341	TBIL	26472	4.25	02/20/2021::11:39:39	3.98-4.70
342	TP	26472	4.46	02/20/2021::11:39:39	3.51-4.59
343	GLUC	26472	268.20	02/20/2021::11:39:57	272.97-287.53
344	CHOL	26472	102.65	02/20/2021::11:40:40	97.37-102.77
345	TG	26472	96.93	02/20/2021::11:42:38	83.94-96.86
346	HDLC	26472	28.96	02/20/2021::11:43:52	17.38-31.46
347	LDLC	26472	69.91	02/20/2021::11:43:52	61.20-69.96
348	CAL	26472	11.21	02/23/2021::10:55:56	9.65-12.89
349	ALP	26472	397.46	02/23/2021::10:56:39	381.62-419.94
350	SGPT	26472	110.86	02/23/2021::10:56:52	100.63-113.23
351	ALB	26472	2.78	02/23/2021::10:57:15	2.55-2.99
352	DBIL	26472	1.58	02/23/2021::10:57:37	1.45-1.77
353	GGT	26472	147.02	02/23/2021::10:57:50	136.01-149.73
354	CREAT	26472	5.20	02/23/2021::10:58:42	4.30-5.66
355	UREA	26472	96.46	02/23/2021::10:58:55	82.60-104.52
356	SGOT	26472	193.34	02/23/2021::10:59:34	177.94-201.54
357	UA	26472	8.77	02/23/2021::11:01:52	7.49-10.45
358	TBIL	26472	4.40	02/23/2021::11:03:11	3.98-4.70
359	TP	26472	4.16	02/23/2021::11:03:24	3.51-4.59
360	GLUC	26472	280.26	02/23/2021::11:03:37	272.97-287.53
361	CHOL	26472	103.92	02/23/2021::11:04:31	97.37-102.77
362	TG	26472	94.15	02/23/2021::11:06:16	83.94-96.86
363	HDLC	26472	18.27	02/23/2021::11:07:11	17.38-31.46
364	LDLC	26472	75.07	02/23/2021::11:07:11	61.20-69.96
365	TG	26472	94.16	02/23/2021::11:19:32	83.94-96.86
366	CHOL	26472	103.92	02/23/2021::11:21:16	97.37-102.77
367	TG	26472	94.15	02/23/2021::11:36:12	83.94-96.86
368	CAL	26472	11.36	02/25/2021::10:17:07	9.65-12.89
369	ALP	26472	408.82	02/25/2021::10:17:49	381.62-419.94
370	SGPT	26472	114.84	02/25/2021::10:18:02	100.63-113.23
371	ALB	26472	2.77	02/25/2021::10:18:25	2.55-2.99
372	DBIL	26472	1.55	02/25/2021::10:18:48	1.45-1.77
373	GGT	26472	147.95	02/25/2021::10:19:01	136.01-149.73
374	CREAT	26472	4.98	02/25/2021::10:19:53	4.30-5.66
375	UREA	26472	98.22	02/25/2021::10:19:54	82.60-104.52
376	SGOT	26472	191.32	02/25/2021::10:20:46	177.94-201.54
377	UA	26472	7.78	02/25/2021::10:23:04	7.49-10.45
378	TBIL	26472	4.24	02/25/2021::10:24:22	3.98-4.70

### Qc results

From: 01/01/2021

To: 13/03/2021

Concentration: High

No.	Test name	QC lot	Result	Run date	QC range
379	TP	26472	4.09	02/25/2021::10:24:36	3.51-4.59
380	GLUC	26472	279.22	02/25/2021::10:24:49	272.97-287.53
381	CHOL	26472	99.59	02/25/2021::10:25:43	97.37-102.77
382	TG	26472	91.80	02/25/2021::10:27:27	83.94-96.86
383	HDLC	26472	23.45	02/25/2021::10:28:19	17.38-31.46
384	LDLC	26472	65.23	02/25/2021::10:28:19	61.20-69.96
385	CAL	26472	10.87	02/27/2021::10:43:02	9.65-12.89
386	ALP	26472	424.10	02/27/2021::10:43:44	381.62-419.94
387	SGPT	26472	111.78	02/27/2021::10:43:57	100.63-113.23
388	ALB	26472	2.88	02/27/2021::10:44:20	2.55-2.99
389	DBIL	26472	1.48	02/27/2021::10:44:42	1.45-1.77
390	GGT	26472	143.61	02/27/2021::10:44:55	136.01-149.73
391	CREAT	26472	5.13	02/27/2021::10:45:48	4.30-5.66
392	UREA	26472	92.85	02/27/2021::10:45:48	82.60-104.52
393	SGOT	26472	206.50	02/27/2021::10:46:40	177.94-201.54
394	UA	26472	8.49	02/27/2021::10:48:58	7.49-10.45
395	TBIL	26472	4.33	02/27/2021::10:50:17	3.98-4.70
396	TP	26472	4.16	02/27/2021::10:50:30	3.51-4.59
397	GLUC	26472	282.73	02/27/2021::10:50:43	272.97-287.53
398	CHOL	26472	100.70	02/27/2021::10:51:38	97.37-102.77
399	TG	26472	86.28	02/27/2021::10:53:22	83.94-96.86
400	HDLC	26472	25.21	02/27/2021::10:54:14	17.38-31.46
401	LDLC	26472	77.03	02/27/2021::10:54:14	61.20-69.96
402	CAL	26472	10.64	03/03/2021::10:56:45	9.65-12.89
403	ALP	26472	415.46	03/03/2021::10:57:27	381.62-419.94
404	SGPT	26472	110.86	03/03/2021::10:57:40	100.63-113.23
405	ALB	26472	2.70	03/03/2021::10:58:03	2.55-2.99
406	DBIL	26472	1.55	03/03/2021::10:58:26	1.45-1.77
407	GGT	26472	135.86	03/03/2021::10:58:38	136.01-149.73
408	CREAT	26472	5.07	03/03/2021::10:59:31	4.30-5.66
409	UREA	26472	91.02	03/03/2021::10:59:44	82.60-104.52
410	SGOT	26472	197.46	03/03/2021::11:00:23	177.94-201.54
411	UA	26472	8.20	03/03/2021::11:02:41	7.49-10.45
412	TBIL	26472	4.32	03/03/2021::11:04:00	3.98-4.70
413	TP	26472	3.66	03/03/2021::11:04:13	3.51-4.59
414	GLUC	26472	276.52	03/03/2021::11:04:26	272.97-287.53
415	CHOL	26472	99.05	03/03/2021::11:05:20	97.37-102.77
416	TG	26472	84.86	03/03/2021::11:07:04	83.94-96.86
417	HDLC	26472	20.13	03/03/2021::11:07:56	17.38-31.46
418	LDLC	26472	75.86	03/03/2021::11:07:57	61.20-69.96
419	CAL	26472	10.84	03/05/2021::11:00:31	9.65-12.89
420	ALP	26472	393.71	03/05/2021::11:01:13	381.62-419.94
421	SGPT	26472	110.04	03/05/2021::11:01:27	100.63-113.23
422	ALB	26472	2.70	03/05/2021::11:01:50	2.55-2.99
423	DBIL	26472	1.63	03/05/2021::11:02:12	1.45-1.77
424	GGT	26472	137.93	03/05/2021::11:02:25	136.01-149.73
425	CREAT	26472	4.95	03/05/2021::11:03:18	4.30-5.66
426	UREA	26472	107.56	03/05/2021::11:03:18	82.60-104.52
427	SGOT	26472	189.99	03/05/2021::11:04:10	177.94-201.54
428	UA	26472	7.83	03/05/2021::11:06:29	7.49-10.45
429	TBIL	26472	4.30	03/05/2021::11:07:47	3.98-4.70
430	TP	26472	3.80	03/05/2021::11:08:00	3.51-4.59
431	GLUC	26472	275.74	03/05/2021::11:08:13	272.97-287.53
432	CHOL	26472	97.82	03/05/2021::11:09:08	97.37-102.77

### Qc results

From: 01/01/2021

To: 13/03/2021

Concentration: High

No.	Test name	QC lot	Result	Run date	QC range
433	TG	26472	87.67	03/05/2021::11:10:52	83.94-96.86
434	HDLC	26472	21.81	03/05/2021::11:11:44	17.38-31.46
435	LDLC	26472	64.99	03/05/2021::11:11:44	61.20-69.96
436	CAL	26472	10.77	03/08/2021::11:53:17	9.65-12.89
437	ALP	26472	388.18	03/08/2021::11:53:59	381.62-419.94
438	SGPT	26472	109.24	03/08/2021::11:54:12	100.63-113.23
439	ALB	26472	2.74	03/08/2021::11:54:36	2.55-2.99
440	DBIL	26472	1.59	03/08/2021::11:54:58	1.45-1.77
441	GGT	26472	147.88	03/08/2021::11:55:11	136.01-149.73
442	CREAT	26472	4.65	03/08/2021::11:56:04	4.30-5.66
443	UREA	26472	104.31	03/08/2021::11:56:04	82.60-104.52
444	SGOT	26472	190.30	03/08/2021::11:56:56	177.94-201.54
445	UA	26472	7.91	03/08/2021::11:59:14	7.49-10.45
446	TBIL	26472	4.42	03/08/2021::12:00:32	3.98-4.70
447	TP	26472	4.23	03/08/2021::12:00:46	3.51-4.59
448	GLUC	26472	279.30	03/08/2021::12:00:59	272.97-287.53
449	CHOL	26472	101.06	03/08/2021::12:01:53	97.37-102.77
450	TG	26472	90.82	03/08/2021::12:03:37	83.94-96.86
451	HDLC	26472	30.37	03/08/2021::12:04:29	17.38-31.46
452	LDLC	26472	69.41	03/08/2021::12:04:29	61.20-69.96
453	CAL	26472	11.10	03/10/2021::08:51:15	9.65-12.89
454	ALP	26472	384.18	03/10/2021::08:51:57	381.62-419.94
455	SGPT	26472	111.11	03/10/2021::08:52:10	100.63-113.23
456	ALB	26472	2.74	03/10/2021::08:52:33	2.55-2.99
457	DBIL	26472	1.51	03/10/2021::08:52:56	1.45-1.77
458	GGT	26472	143.90	03/10/2021::08:53:09	136.01-149.73
459	CREAT	26472	4.57	03/10/2021::08:54:01	4.30-5.66
460	UREA	26472	103.49	03/10/2021::08:54:14	82.60-104.52
461	SGOT	26472	183.66	03/10/2021::08:54:53	177.94-201.54
462	UA	26472	8.15	03/10/2021::08:57:12	7.49-10.45
463	TBIL	26472	4.24	03/10/2021::08:58:31	3.98-4.70
464	TP	26472	3.81	03/10/2021::08:58:43	3.51-4.59
465	GLUC	26472	279.01	03/10/2021::08:58:56	272.97-287.53
466	CHOL	26472	98.95	03/10/2021::08:59:50	97.37-102.77
467	TG	26472	93.23	03/10/2021::09:01:35	83.94-96.86
468	HDLC	26472	30.62	03/10/2021::09:02:27	17.38-31.46
469	LDLC	26472	66.47	03/10/2021::09:02:27	61.20-69.96
470	CAL	26472	10.65	03/13/2021::10:55:25	9.65-12.89
471	ALP	26472	393.15	03/13/2021::10:56:08	381.62-419.94
472	SGPT	26472	103.81	03/13/2021::10:56:21	100.63-113.23
473	ALB	26472	2.83	03/13/2021::10:56:44	2.55-2.99
474	DBIL	26472	1.62	03/13/2021::10:57:06	1.45-1.77
475	GGT	26472	137.27	03/13/2021::10:57:19	136.01-149.73
476	CREAT	26472	4.53	03/13/2021::10:58:12	4.30-5.66
477	UREA	26472	102.99	03/13/2021::10:58:25	82.60-104.52
478	SGOT	26472	175.35	03/13/2021::10:59:04	177.94-201.54
479	UA	26472	8.77	03/13/2021::11:01:22	7.49-10.45
480	TBIL	26472	3.96	03/13/2021::11:02:40	3.98-4.70
481	TP	26472	2.51	03/13/2021::11:02:54	3.51-4.59
482	GLUC	26472	279.80	03/13/2021::11:03:07	272.97-287.53
483	CHOL	26472	99.79	03/13/2021::11:04:01	97.37-102.77
484	TG	26472	91.53	03/13/2021::11:05:45	83.94-96.86
485	HDLC	26472	21.19	03/13/2021::11:06:37	17.38-31.46
486	LDLC	26472	64.05	03/13/2021::11:06:37	61.20-69.96