

15-04-2021 13:42:30
13.0176318
77.6433729





VAIDYANATHESHWARA INSTRUMENTS

CERTIFICATE OF CALIBRATION



No. 301/A, 9th Main Road, 3rd Cross, Rajiv Gandhi Nagar, J.B. Kaval, Nandhini Layout Post, Bangalore - 560 096.
Contact : 080-23377266, Mob : 9986586789 / 9448080177 / 9964308118 | Email : info@viplgroup.com Web : www.viplgroup.com

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2017 With vide Certificate No: CC-2473

1 Name and Address of the Customer

M/s.: Energon Health.,
No.703, 3rd Cross Road, HRBR Layout,
1st Block, Kalyan Nagar, Bangalore - 560 043.

2 Customer Reference

- 2.1 ULR No. : CC2473215000023414F
- 2.2 SRF No : 0037
- 2.3 Certificate No. : VI/21-22/0037-05
- 2.4 Format No. : VI-FRM-ME-046 & VI-FRM-ET-011
- 2.5 Date of Issue : 09-04-2021

3 Details Of Device Under Calibration(DUC).

- 3.1 Nomenclature : Centrifuge
- 3.2 Make / Model : Remi / R-8C
- 3.3 Id.No. : EH/CF/01
- 3.4 DUC Condition : Satisfactory
- 3.5 Calibration Procedure No. : SOP-16- 42 & SOP-37-11
- 3.6 No.of Pages : 2
- 3.7 Calibration Date : 08-04-2021
- 3.8 Calibration Due : 07-04-2022
- 3.9 Calibration done at : Onsite
- 3.10 Instrument Location : Lab
- 3.11 Discipline : Mechanical

4 Environmental Condition

Temperature 25.3 °C Humidity: 52 %Rh

5 Standards Used for calibration


Sl. No.	Nomenclature	Make & Model	Sl. No. / Id.No.	Certificate No.	Validity
1	Digital Tachometer	SENTRY / ST723	SN1323004225	ULR- CC287920000004732F	10-09-2021
2	Time Totalizer	Beltronics / 501	251	VI/20-21/INT-ETH-426	23-02-2022

6 Notes:

- 6.1. The Calibration Certificate relates only to the above DUC
- 6.2. Publication or reproduction of this Certificate in any form other than by complete set of the whole report & in the language, written, is not permitted without the written consent of VI Lab.
- 6.3. Corrections/erasing, invalidate the Calibration Certificate.
- 6.4. Calibration of the DUC are traceable to National standards/International Standards
- 6.5. Any error in this Certificate should be brought to our knowledge within 30 days from the date of this Cert.
- 6.6. Results Reported are valid at the time of and under the stated conditions of measurements
- 6.7. The usage of NABL Symbol is as per NABL guidelines given in NABL-133

Calibrated By
F. Tejo
Madhusudhan
(Calibration Engineer)

Checked By
P. Raju
Umesh D
(Onsite In-Charge)





VAIDYANATHESHWARA INSTRUMENTS

CERTIFICATE OF CALIBRATION



COMMITTED TO THE
CUSTOMER SINCE - 1996

No. 301/A, 9th Main Road, 3rd Cross, Rajiv Gandhi Nagar, J.B. Kaval, Nandhini Layout Post, Bangalore - 560 096.

Contact : 080-23377266, Mob : 9986586789 / 9448080177 / 9964308118 | Email : info@viplgroup.com Web : www.viplgroup.com

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2017 With vide Certificate No: CC-2473

Certificate No.

Page No: 2 of 2

RPM Results:

Range: 0 to 5250 RPM

Resolution: 1 RPM

Sl. No.	Std Reading (RPM)	DUC Reading (RPM)	Error Observed in (RPM)	Measurement Uncertainty in \pm (%)	k Factor
1	199.9	200	0.1	1.2	2.0
2	599.7	600	0.3		2.0
3	999.5	1000	0.5		2.0
4	1999.2	2000	0.8		2.0
5	2998.9	3000	1.1		2.0
6	3998.6	4000	1.4		2.0
7	5248.3	5250	1.7		2.0

Timer Results:

Range: 0 to 99 min

Resolution: 1min


Sl. No.	Std Reading (hr:min:sec:msec)	DUC Reading (min)	Error Observed in (hr:min:sec:msec)	Measurement Uncertainty in \pm (%)	k Factor
1	00:00:00:000	0	00:00:00:000	----	----
2	00:15:00:012	15	-00:00:00:012	0.05	2.0
3	00:30:00:026	30	-00:00:00:026	0.05	2.0
4	00:45:00:042	45	-00:00:00:042	0.05	2.0
5	01:00:00:059	60	-00:00:00:059	0.05	2.0
6	01:39:00:067	99	-00:00:00:067	0.05	2.0

Conclusion :-

1. Measurement uncertainty reported is at 95.45 % confidence level.

Calibrated By
P. Tejo
Madhusudhan
(Calibration Engineer)

Checked By
P. Tejo
Umesh D
(Onsite In-Charge)





Vaidyanatheshwara Instruments

CERTIFICATE OF CALIBRATION



No. 301/A, 9th Main Road, 3rd Cross, Rajiv Gandhi Nagar, J.B. Kaval, Nandhini Layout Post, Bangalore - 560 096.
Contact : 080-23377266, Mob : 9986586789 / 9448080177 / 9964308118 | Email : info@viplgroup.com Web : www.viplgroup.com

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2017 With vide Certificate No: CC-2473

1 Name and Address of the Customer
M/s.: Energon Health ,
No.703, 3rd Cross Road, HRBR Layout ,
1 st Block, Kalyan Nagar, Bangalore - 560 043.

2 Customer Reference
2.1 ULR No. : CC247321500022900F
2.2 Format No. : VI-FRM-TH-006
2.3 SRF No. : 0032
2.4 Certificate No. : VI/21-22/0032-01
2.5 Date of Issue : 07-04-2021

3 Details Of Device Under Calibration(DUC).
3.1 Nomenclature : Refrigerator
3.2 Make : LG
3.3 Id.No. : EH/REF/01
3.4 Range / Resolution : 2 to 8 °C / 0.1 °C
3.5 DUC Condition : Satisfactory
3.6 Calibration Procedure No. : SOP-38-06
3.7 No.of Pages : 2
3.8 Calibration Date : 06-04-2021
3.9 Calibration Due : 05-04-2022
3.10 Calibration done at : Onsite
3.11 Instrument Location : Sample Storage
3.12 Discipline : Thermal

4 Environmental Condition
Temperature 26.4 °C Humidity 54 %RH

5 Standards Used for Temperature Calibration


Sl. No.	Nomenclature	Make / Model	Sl. No.	Certificate. No.	Validity
1	Precision Thermometer with RTD Sensor	Lutron / TM-917	Q601311	VI/20-21/INT-TH-55	28-08-2021

6 Note:

- 6.1.The Calibration Certificate relates only to the above DUC
- 6.2.Publication or reproduction of this Certificate in any form other than by complete set of the whole report & in the language, written, is not permitted without the written consent of VI.
- 6.3.Corrections/erasing, invalidate the Calibration Certificate.
- 6.4.Calibration of the DUC are traceable to National standards/International Standards
- 6.5.Any error in this Certificate should be brought to our knowledge within 30 days from the date of this Cert.
- 6.6.Results Reported are valid at the time of and under the stated conditions of measurements.
- 6.7.The usage of NABL Symbol is as per NABL guidelines given in NABL-133

Calibrated By
F Prathi
Madhusudhan
(Calibration Engineer)

Checked By
P. Umesh
Umesh D
(Onsite In-Charge)





VAIDYANATHESHWARA INSTRUMENTS

CERTIFICATE OF CALIBRATION



COMMITTED TO THE
CUSTOMER SINCE - 1996

No. 301/A, 9th Main Road, 3rd Cross, Rajiv Gandhi Nagar, J.B. Kaval, Nandhini Layout Post, Bangalore - 560 096.
Contact : 080-23377266, Mob : 9986586789 / 9448080177 / 9964308118 | Email : info@viplgroup.com Web : www.viplgroup.com

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2017 With vide Certificate No: CC-2473

Certificate No. VI/21-22/0032-01


Results:

Range : 2 to 8 °C

Resolution: 0.1 °C

SI No	STD Reading (°C)	DUC Reading (°C)	Error Observed (°C)	Measurement Uncertainty ± (°C)	k Factor
1	2.05	2.0	-0.05	1.6	2.0
2	3.09	3.0	-0.09	1.6	2.0
3	4.13	4.0	-0.13	1.6	2.0
4	6.18	6.0	-0.18	1.6	2.0
5	8.25	8.0	-0.25	1.6	2.0

Pictorial Representation of Sensors Placement



Note :

- S1 Indicates Placement of Standard Sensor

Conclusion / Remarks:

1 Measurement Uncertainty reported is at 95.45% confidence level with k = 2.

Calibrated By

J. Prathe
Madhusudan
(Calibration Engineer)

Checked By

Umesh D
Umesh D
(Onsite In-Charge)

Authorised By





VI Vaidyanatheshwara Ins 23377266
Cert. No.: VI/21-22/0125-01
SIID No.: 317582
Cal On: 09-04-2021
Cal Due: 08-04-2022



VI Vaidyanatheshwara Ins 23377266
Cert. No.: VI/21-22/0125-02
SIID No.: 317974
Cal On: 09-04-2021
Cal Due: 08-04-2022

Meril Life More Meril Life More Meril Life More
BEST SEAL IS BROKEN IF SE REPORT
REPORT IF SE REPORT

Vaidyanatheshwara
INSTRUMENTS

9800
x/B
086

	V Vaidyanatheshwara INS  23377266
Cert. No.:	VI/21-22/0086-02
SI/ID No.:	YE209A0003272
Cal On:	07/04/2021
Cal Due:	06/04/2022

	V Vaidyanatheshwara INS  23377266
Cert. No.:	VI/21-22/0086-01
SI/ID No.:	YE209A50035842
Cal On:	07/04/2021
Cal Due:	06/04/2022

9800
x/B
086

Vaidyanatheshwara
INSTRUMENTS



CERTIFICATE OF CALIBRATION

No. 301/A, 9th Main Road, 3rd Cross, Rajiv Gandhi Nagar, J.B. Kaval, Nandhini Layout Post, Bangalore - 560 096.

Contact : 080-23377266, Mob : 9986586789 / 9448080177 / 9964308118 | Email : info@viplgroup.com Web : www.viplgroup.com

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2017 With vide Certificate No: CC-2473

1 Name and Address of the Customer : M/s. DIALOGIX HEALTH SERVICES PVT. LTD.,
Kalyan Nagar , Bangalore.

2 Customer Reference

- 2.1 ULR No : CC247321100009950F
- 2.2 SRF No : 0086
- 2.3 Certificate No. : VI/21-22/0086-01
- 2.4 Format No. : VI-FRM-ME-105
- 2.5 Receipt Date : 07 - 04 - 2021
- 2.6 Date Of Issue : 07 - 04 - 2021

3 Details Of Device Under Calibration(DUC).

- 3.1 Nomenclature : Pipette
- 3.2 Make : DRAGON LAB
- 3.3 Range : 5 - 50 µl
- 3.4 SL No. : YE209AS0035842
- 3.5 DUC Condition : Satisfactory
- 3.6 Calibration Procedure No. : SOP-16-71 Based On ISO 8655-6:2002 E
- 3.7 No.of Pages : 2
- 3.8 Calibration Date : 07-04-2021
- 3.9 Calibration Due : 06-04-2022
- 3.10 Calibration done at : VI Volumetric Lab
- 3.11 Discipline : Mechanical (Mass & volume)

4 Environmental Condition

Temperature 20.4 °C Humidity 52 %Rh

5 Standards Used for calibration

Sl. No.	Nomenclature	Make & Model	Sl. No	Range	Traceable to /Cert. No.	Validity
1	Digital Weighing Balance	RADWAG/MYA5.4Y	544953	0.1mg to 5g	VI/20-21/INT-ME-121	18-09-2021
2	E1 Class Weights	LCGC	VI-E1-ME-001	1mg to 200g	TVCSPL 19/04/339-01	02 - 04 - 2022

6 Note:

- 6.1. The Calibration Certificate relates only to the above DUC
- 6.2. Publication or reproduction of this Certificate in any form other than by complete set of the whole report & in the language, written, is not permitted without the written consent of VI Lab..
- 6.3. Corrections/erasing, invalidate the Calibration Certificate.
- 6.4. Calibration of the DUC are traceable to National standards/International Standards
- 6.5. Any error in this Certificate should be brought to our knowledge within 30 days from the date of this Cert.
- 6.6. Results Reported are valid at the time of and under the stated conditions of measurements
- 6.7. The usage of NABL symbol is as per NABL guidelines given in NABL-133.

Calibrated By
Veda Murthy K.S
Veda Murthy K.S
(Calibration Engineer)

Checked By
P. Santhosh Kumar
P. Santhosh Kumar
(Lab- In-Charge)





Vaidyanatheshwara Instruments



CERTIFICATE OF CALIBRATION

No. 301/A, 9th Main Road, 3rd Cross, Rajiv Gandhi Nagar, J.B. Kaval, Nandhini Layout Post, Bangalore - 560 096.

Contact : 080-23377266, Mob : 9986586789 / 9448080177 / 9964308118 | Email : info@viplgroup.com Web : www.viplgroup.com

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2017 With vide Certificate No: CC-2473

Certificate No: VI/21-22/0086-01

Page 2 of 2

Range: 5 - 50 µl

Results:

Range	Cal Point	Volume observed	Expanded Uncertainty ±
	µl	µl	µl
5 - 50 µl	5.0	5.014	0.02
	20.0	20.025	0.04
	50.0	50.046	0.20

Note:

1. Visual Inspection : Found Well.
2. Triple Distilled water is used to Calibrate the MicroPipette.


Conclusion /Remarks:

1. Ref. standard used are traceable to National/International Standard
2. The Expanded Uncertainty of associated with measurement at approximate 95.45% confidence level with coverage factor k=2

Calibrated By

Veda Murthy K.S
(Calibration Engineer)

Checked By

P. Santhosh Kumar
(Lab- In-Charge)

Gangadhar C.K



Vaidyanatheshwara Instruments



COMMITTED TO THE CUSTOMER SINCE - 1996

CERTIFICATE OF CALIBRATION

No. 301/A, 9th Main Road, 3rd Cross, Rajiv Gandhi Nagar, J.B. Kaval, Nandhini Layout Post, Bangalore - 560 096.

Contact : 080-23377266, Mob : 9986586789 / 9448080177 / 9964308118 | Email : info@viplgroup.com Web : www.viplgroup.com

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2017 With vide Certificate No: CC-2473

Page 1 of 2

1 Name and Address of the Customer : M/s. DIALOGIX HEALTH SERVICES PVT. LTD.,
Kalyan Nagar , Bangalore.

2 Customer Reference

- 2.1 ULR No : CC247321100009951F
- 2.2 SRF No : 0086
- 2.3 Certificate No. : VI/21-22/0086-02
- 2.4 Format No. : VI-FRM-ME-105
- 2.5 Receipt Date : 07 - 04 - 2021
- 2.6 Date Of Issue : 07 - 04 - 2021

3 Details Of Device Under Calibration(DUC).

- 3.1 Nomenclature : Pipette
- 3.2 Make : DRAGON LAB
- 3.3 Range : 100 - 1000 µl
- 3.4 SL No. : YEAA9AD0013272
- 3.5 DUC Condition : Satisfactory
- 3.6 Calibration Procedure No. : SM/WI/M-10 Based On ISO 8655-6:2002 E
- 3.7 No.of Pages : 2
- 3.8 Calibration Date : 07-04-2021
- 3.9 Calibration Due : 06-04-2022
- 3.10 Calibration done at : VI Volumetric Lab
- 3.11 Discipline : Mechanical (Mass & Volume)

4 Environmental Condition

Temperature 20.4 °C Humidity 51 %Rh

5 Standards Used for calibration


Sl. No.	Nomenclature	Make & Model	Sl. No	Range	Traceable to /Cert. No.	Validity
1	Digital Weighing Balance	RADWAG/MYA5.4Y	544953	0.1mg to 5g	VI/20-21/INT-ME-121	18-09-2021
2	E1 Class Weights	LCGC	VI-E1-ME-001	1mg to 200g	TVCSP1 19/04/339-01	02 - 04 - 2022

6 Note:

- 6.1. The Calibration Certificate relates only to the above DUC
- 6.2. Publication or reproduction of this Certificate in any form other than by complete set of the whole report & in the language, written, is not permitted without the written consent of VI Lab..
- 6.3. Corrections/erasing, invalidate the Calibration Certificate.
- 6.4. Calibration of the DUC are traceable to National standards/International Standards
- 6.5. Any error in this Certificate should be brought to our knowledge within 30 days from the date of this Cert.
- 6.6. Results Reported are valid at the time of and under the stated conditions of measurements
- 6.7. The usage of NABL symbol is as per NABL guidelines given in NABL-133.

Calibrated By
Veda Murthy K.S
Veda Murthy K.S
(Calibration Engineer)

Checked By
P Santhosh Kumar
P Santhosh Kumar
(Lab In-Charge)





COMMITTED TO THE
CUSTOMER SINCE - 1996

Vaidyanatheshwara INSTRUMENTS

CERTIFICATE OF CALIBRATION



No. 301/A, 9th Main Road, 3rd Cross, Rajiv Gandhi Nagar, J.B. Kaval, Nandhini Layout Post, Bangalore - 560 096.
Contact : 080-23377266, Mob : 9986586789 / 9448080177 / 9964308118 | Email : info@viplgroup.com Web : www.viplgroup.com

NABL Accredited Calibration Lab as per ISO/IEC 17025 : 2017 With vide Certificate No: CC-2473

Certificate No: VI/21-22/0086-02

Page 2 of 2

Range: 100 - 1000 μ l

Results:

Range	Cal Point	Volume observed	Expanded Uncertainty \pm
	μ l	μ l	μ l
100 - 1000 μ l	100.0	99.459	0.62
	500.0	498.838	3.10
	1000.0	998.159	6.20

Note:

1. Visual Inspection : Found Well.
2. Tripple Distilled water is used to Calibrate the MicroPipette.

Conclusion /Remarks:

1. Ref. standard used are traceable to National/International Standard
2. The Expanded Uncertainty of associated with measurement at approximate 95.45% confidence level with coverage factor k=2

Calibrated By

Veda Murthy K.S
Veda Murthy K.S
(Calibration Engineer)

Checked By

P Santhosh Kumar
P Santhosh Kumar
(Lab In-Charge)

