

**MEASURE TECHNO LAB**  
**2, B.T. ROAD (JAYANTI CINEMA COMPLEX)**  
**BARRACKPORE, KOLKATA - 700120, W.B.**

Form No. - MTL/22/2006

**CALIBRATION CERTIFICATE OF  
 MICRO PIPETTE**



Phone : 033 - 2215 - 0032, 2215 - 9687, 8100875519,  
 Mobile: 9831190974, LAB:- 8100143376,  
 E-mail: measuretechno@yahoo.co.in

CALIBRATION CERTIFICATE NO.: MTL / SDC / R03 / 05 - 21

ULR - CC25452100009550F

Page: 1 of 1

1.0 Service Request No.: MTL / 11 / 05 / 21- 22

1.1 Issued to: M/s. Sun Diagnostic Centre,  
 IMA Building, Ranihat, Near SCB Medical College,  
 Cuttack, Odisha - 753007.

1.2 Description & Identification of item to be Calibrated:

a) Name:	Micro Pipette	b) Code No.:	SDC / LAB / PIPETTE / 01
c) Sl. No.:	YE168AA0027340	d) Make:	Dragon Lab
e) Model / Type:	Micropette	f) Range:	5 µl to 50 µl
g) Sensor:	N.A.	h) Resolution:	0.5 µl
i) End User:	Pathology	j) Accuracy:	N.S.

k) Calibration done at: On Site / √ In House

1.3 Date of receipt of item : 10-05-21

1.4 Physical Condition of DUC : OK

1.5 Date of calibration : 11-05-21

1.6 Recommended date of next calibration : 11-05-22

1.7 Date of Issue : 12-05-21

1.8 Environmental Conditions During Calibration:

Temperature:	23 °C ± 2 °C
Humidity:	50 % RH ± 10 % RH
Pressure:	1012.5 mbar

1.9 Method of Calibration: SOP / MASS / 02 ( As Per ISO : 8655 - 6 : 2002)

2.0 Traceability :

- a) Standards used for calibration are traceable to National standards through NABL Accredited Laboratory.
- b) The following standards / Equipment have been used.

- i) 1 mg to 200 g Weight Box (E1) Cal. Certificate No. WI / May / 19 / 010 (WEIGH INDIA, New Delhi) (Cal. Date: 17/05/19, Due Date: 17/05/22)
- ii) Data Acquisition Switch Unit Cal. Certificate No. JRPM - CCTR - ET - 2020 - 0795 (JRPM, Chennai) (Cal. Date: 01/09/20, Due Date: 31/08/21)
- iii) RTD (PT - 100) Cal. Certificate No. TL / 020 / 891.2.1 (TEMPSSENS, Udaipur) (Cal. Date: 27/10/20, Due Date: 26/10/21)

2.1 Result :

Mechanical Calibration

Sl. No.	Parameter/ Range	Nominal Value µl	Mass of Water mg	Volume of Water at 20 °C µl	Error µl	Measurement Expanded Uncertainty ± µl
1.	Volume	5	4.9740	4.9939	-0.0061	0.039
2.	5 µl to 50 µl	25	24.8873	24.9868	-0.0132	0.39
3.		50	49.7846	49.9837	-0.0163	0.39

Remarks: i) Cubical Expansion co - efficient of pipette material taken as  $10^{-5} \mu\text{l} / ^\circ\text{k}$ .

- ii) This result has an expanded uncertainty with a coverage factor  $k=2$  at approximately 95% confidence level.
- iii) The calibration certificate issued for this instrument is to be used for scientific or industrial purposes only.

DUC - Device Under Calibration

N.S. - Not Specified


N.A. - Not Applicable

Opinions and Interpretations

Calibrated	√	Accepted / Valid for use
Limited Use		Rejected / Out of use

Calibrated by:  
*Testing Engineer*  
**Measure Techno Lab**  
 Testing Engineer  
*SM 12/05/21*

Checked / Approved by:  
  
 Quality & Technical Manager  
**S. Pandey**

<b>MEASURE TECHNO LAB</b> <b>2, B.T. ROAD (JAYANTI CINEMA COMPLEX)</b> <b>BARRACKPORE, KOLKATA - 700120, W.B.</b> Phone : 033 - 2215 - 0032, 2215 - 9687, 8100875519, Mobile: 9831190974, LAB:- 8100143376, E-mail: measuretechno@yahoo.co.in	Form No. - MTL/22/2006	 NABL ACCREDITED LABORATORY Certificate No. CC - 2545
	<b>CALIBRATION CERTIFICATE OF MICRO PIPETTE</b>	
CALIBRATION CERTIFICATE NO.: MTL / SDC / R04 / 05 - 21	ULR - CC254521000009551F	Page: 1 of 1

1.0 Service Request No.: MTL / 11 / 05 / 21- 22

1.1 Issued to: M/s. Sun Diagnostic Centre,  
 IMA Building, Ranihat, Near SCB Medical College,  
 Cuttack, Odisha - 753007.

1.2 Description & Identification of item to be Calibrated:	a) Name:	Micro Pipette	b) Code No.:	SDC / LAB / PIPETTE / 02
	c) Sl. No.:	YE6F725706	d) Make:	Dragon Lab
	e) Model / Type:	Micropette	f) Range:	100 µl to 1000 µl
	g) Sensor:	N.A.	h) Resolution:	5 µl
	i) End User:	Pathology	j) Accuracy:	N.S.
	k) Calibration done at:	On Site / √ In House		

1.3 Date of receipt of item : 10-05-21

1.4 Physical Condition of DUC : OK

1.5 Date of calibration : 11-05-21

1.6 Recommended date of next calibration : 11-05-22

1.7 Date of Issue : 12-05-21

1.8 Environmental Conditions During Calibration:

Temperature:	23 °C ± 2 °C
Humidity:	50 % RH ± 10 % RH
Pressure:	1012.5 mbar

1.9 Method of Calibration: SOP / MASS / 02 ( As Per ISO : 8655 - 6 : 2002)

2.0 Traceability :

a) Standards used for calibration are traceable to National standards through NABL Accredited Laboratory.

b) The following standards / Equipment have been used.

- i) 1 mg to 200 g Weight Box (E1) Cal. Certificate No. WI / May / 19 / 010 (WEIGH INDIA, New Delhi) (Cal. Date: 17/05/19, Due Date: 17/05/22)
- ii) Data Acquisition Switch Unit Cal. Certificate No. JRPM - CCTR - ET - 2020 - 0795 (JRPM, Chennai) (Cal. Date: 01/09/20, Due Date: 31/08/21)
- iii) RTD (PT - 100) Cal. Certificate No. TL / 020 / 891.2.1 (TEMSENS, Udaipur) (Cal. Date: 27/10/20, Due Date: 26/10/21)

2.1 Result :

Mechanical Calibration

Sl. No.	Parameter/ Range	Nominal Value µl	Mass of Water mg	Volume of Water at 20 °C µl	Error µl	Measurement Expanded Uncertainty ± µl
1.	Volume	100	99.5830	99.9813	-0.0187	
2.	100 µl to 1000 µl	500	497.9670	499.9589	-0.0411	0.39
3.		1000	995.9380	999.9218	-0.0782	

- Remarks: i) Cubical Expansion co - efficient of pipette material taken as  $10^{-5} \mu\text{l} / ^\circ\text{k}$ .  
 ii) This result has an expanded uncertainty with a coverage factor  $k=2$  at approximately 95% confidence level.  
 iii) The calibration certificate issued for this instrument is to be used for scientific or industrial purposes only.

DUC - Device Under Calibration


N.S. - Not Specified

N.A. - Not Applicable

Opinions and Interpretations

Calibrated	√	Accepted / Valid for use
Limited Use		Rejected / Out of use

Calibrated by:  
**Testing Engineer**  
 S. Mahapatra  
 Testing Engineer  
 Kolkata

Checked / Approved by:  
  
 Quality & Technical Manager  
 S. Pandey

**MEASURE TECHNO LAB**  
**2, B.T. ROAD (JAYANTI CINEMA COMPLEX)**  
**BARRACKPORE, KOLKATA - 700120, W.B.**

Phone : 033 - 2215 - 0032, 2215 - 9687, 8100875519,  
 Mobile: 9831190974, LAB:- 8100143376,  
 E-mail: measuretechno@yahoo.co.in

Form No. - MTL/22/2006

**CALIBRATION CERTIFICATE OF  
 MICRO PIPETTE**



**CALIBRATION CERTIFICATE NO.:** MTL / SDC / R05 / 05 - 21

**ULR -** CC254521000009552F

Page: 1 of 1

**1.0 Service Request No.:** MTL / 11 / 05 / 21- 22

**1.1 Issued to:** M/s. Sun Diagnostic Centre,  
 IMA Building, Ranihat, Near SCB Medical College,  
 Cuttack, Odisha - 753007.

<b>1.2 Description &amp; Identification of item to be Calibrated:</b>	<b>a) Name:</b>	Micro Pipette	<b>b) Code No.:</b>	SDC / LAB / PIPETTE / 03
	<b>c) Sl. No.:</b>	YE176AF0016085	<b>d) Make:</b>	Dragon Lab
	<b>e) Model / Type:</b>	Micropette	<b>f) Range:</b>	1000 µl (Fixed)
	<b>g) Sensor:</b>	N.A.	<b>h) Resolution:</b>	N.A.
	<b>i) End User:</b>	Pathology	<b>j) Accuracy:</b>	N.S.
	<b>k) Calibration done at:</b>	On Site / √ In House		

**1.3 Date of receipt of item :** 10-05-21

**1.4 Physical Condition of DUC :** OK

**1.5 Date of calibration :** 11-05-21

**1.6 Recommended date of next calibration :** 11-05-22

**1.7 Date of Issue :** 12-05-21

**1.8 Environmental Conditions During Calibration:**

<b>Temperature:</b>	23 °C ± 2 °C
<b>Humidity:</b>	50 % RH ± 10 % RH
<b>Pressure:</b>	1012.5 mbar

**1.9 Method of Calibration:** SOP / MASS / 02 ( As Per ISO : 8655 - 6 : 2002)

**2.0 Traceability :**

- a) Standards used for calibration are traceable to National standards through NABL Accredited Laboratory.  
 b) The following standards / Equipment have been used.
- i) 1 mg to 200 g Weight Box (E1) Cal. Certificate No. WI / May / 19 / 010 (WEIGH INDIA, New Delhi) (Cal. Date: 17/05/19, Due Date: 17/05/22)
  - ii) Data Acquisition Switch Unit Cal. Certificate No. JRPM - CCTR - ET - 2020 - 0795 (JRPM, Chennai) (Cal. Date: 01/09/20, Due Date: 31/08/21)
  - iii) RTD (PT - 100) Cal. Certificate No. TL / 020 / 891.2.1 (TEMPSENS, Udaipur) (Cal. Date: 27/10/20, Due Date: 26/10/21)

**2.1 Result :**

**Mechanical Calibration**

Sl. No.	Parameter/ Range	Nominal Value µl	Mass of Water mg	Volume of Water at 20 °C µl	Error µl	Measurement Expanded Uncertainty ± µl
1.	Volume 1000 µl (Fixed)	1000	995.9371	999.9208	-0.0792	0.39


- Remarks: i) Cubical Expansion co - efficient of pipette material taken as  $10^{-5} \mu\text{l} / ^\circ\text{k}$ .  
 ii) This result has an expanded uncertainty with a coverage factor  $k=2$  at approximately 95% confidence level.  
 iii) The calibration certificate issued for this instrument is to be used for scientific or industrial purposes only.

DUC - Device Under Calibration                      N.S. - Not Specified                      N.A. - Not Applicable

Opinions and Interpretations	
Calibrated	√ Accepted / Valid for use
Limited Use	Rejected / Out of use

Calibrated by:  
**Testing Engineer**  
 S. Mahapatra  
 Testing Engineer  
 Kolkata

Checked / Approved by:  
  
 Quality & Technical Manager  
 S. Pandey

<b>MEASURE TECHNO LAB</b> <b>2, B.T. ROAD (JAYANTI CINEMA COMPLEX)</b> <b>BARRACKPORE, KOLKATA - 700120, W.B.</b>	Form No. - MTL/22/2006	 NABL ACCREDITED LABORATORY Certificate No. CC - 2545
	<b>CALIBRATION CERTIFICATE OF MICRO PIPETTE</b>	
Phone : 033 - 2215 - 0032, 2215 - 9687, 8100875519, Mobile: 9831190974, LAB:- 8100143376, E-mail: measuretechno@yahoo.co.in	<b>CALIBRATION CERTIFICATE NO.:</b> MTL / SDC / R06 / 05 - 21	
<b>ULR - CC254521000009553F</b>		<b>Page: 1 of 1</b>

1.0 Service Request No.: MTL / 11 / 05 / 21- 22

1.1 Issued to: M/s. Sun Diagnostic Centre,  
 IMA Building, Ranihat, Near SCB Medical College,  
 Cuttack, Odisha - 753007.

1.2 Description & Identification of item to be Calibrated:

a) Name:	Micro Pipette	b) Code No.:	SDC / LAB / PIPETTE / 04
c) Sl. No.:	V58804	d) Make:	Accupipet
e) Model / Type:	Micropette	f) Range:	1 µl to 100 µl
g) Sensor:	N.A.	h) Resolution:	1 µl
i) End User:	Pathology	j) Accuracy:	N.S.
k) Calibration done at:	On Site / √ In House		

1.3 Date of receipt of item : 10-05-21

1.4 Physical Condition of DUC : OK

1.5 Date of calibration : 11-05-21

1.6 Recommended date of next calibration : 11-05-22

1.7 Date of Issue : 12-05-21

1.8 Environmental Conditions During Calibration:

Temperature:	23 °C ± 2 °C
Humidity:	50 % RH ± 10 % RH
Pressure:	1012.5 mbar

1.9 Method of Calibration: SOP / MASS / 02 ( As Per ISO : 8655 – 6 : 2002)

2.0 Traceability :

a) Standards used for calibration are traceable to National standards through NABL Accredited Laboratory.

b) The following standards / Equipment have been used.

- i) 1 mg to 200 g Weight Box (E1) Cal. Certificate No. WI / May / 19 / 010 (WEIGH INDIA, New Delhi) (Cal. Date: 17/05/19, Due Date: 17/05/22)
- ii) Data Acquisition Switch Unit Cal. Certificate No. JRPM - CCTR - ET - 2020 - 0795 (JRPM, Chennai) (Cal. Date: 01/09/20, Due Date: 31/08/21)
- iii) RTD (PT - 100) Cal. Certificate No. TL / 020 / 891.2.1 (TEMPSENS, Udaipur) (Cal. Date: 27/10/20, Due Date: 26/10/21)

2.1 Result :

Mechanical Calibration

Sl. No.	Parameter/ Range	Nominal Value µl	Mass of Water mg	Volume of Water at 20 °C µl	Error µl	Measurement Expanded Uncertainty ± µl
1.	Volume	1	0.9921	0.9961	-0.0039	0.039
2.	1 µl to 100 µl	50	49.7847	49.9838	-0.0162	0.39
3.		100	99.5826	99.9809	-0.0191	0.39

Remarks: i) Cubical Expansion co - efficient of pipette material taken as  $10^{-5} \mu\text{l} / ^\circ\text{k}$ .

ii) This result has an expanded uncertainty with a coverage factor  $k=2$  at approximately 95% confidence level.

iii) The calibration certificate issued for this instrument is to be used for scientific or industrial purposes only.

DUC - Device Under Calibration

N.S. - Not Specified

N.A. - Not Applicable

Opinions and Interpretations

Calibrated	√	Accepted / Valid for use
Limited Use		Rejected / Out of use

Calibrated by: **Testing Engineer**  
**Measure Techno Lab**  
 S. Manapatra  
 Testing Engineer  
 Kolkata  
 12/05/21

Checked & Approved By:  
  
**Quality & Technical Manager**  
**S. Pandey**