

Timer Calibration Verification Record Sheet

ICTC Lab, Bhosari Hospital

No: ICTC Bhosari

Using Timer: Inhouse verification of timer using certified standard calibrated Timer

03/2021

Timer ID: CRV/MICRO/VIRO/03/2017-18, DT-02


Done by: Ms Bandgar Rupali, Technical Officer, SRL, AFMC, Pune


Timer Setting (Sec)	Certified Time (Sec)	Time Difference (Sec)	Timer Setting (Sec)	Certified Time (Sec)	Time Difference (Sec)	Timer Setting (Sec)	Certified Time (Sec)	Time Difference (Sec)
300	299	-1	600	598	-2	900	900	0
300	298	-2	600	600	0	900	901	1
300	300	0	600	603	3	900	900	0
300	299	-1	600	600.3	0.33	900	900.3	0.3
-	1.00	-	-	2.52	-	-	0.58	-
-	0.3344482	-	-	0.4192256	-	-	0.0641287	-
-	-0.3333333	-	-	0.05	-	-	0.0333333	-

Result: Accepted since variation is within +/- 60 Sec

Interpretation: Timer is in good working condition

Remarks: -


 Ms. Rupali Bandgar
 Technical Officer
 SRL, MSACS
 AFMC, Pune - 411 040


 Incharge Medical Officer
 ICTC, Bhosari Hospital PCMC