



EXTERNAL QUALITY ASSESSMENT SCHEME FORM

Name of Proficiency Testing Provider: NARI, Pune

Year: 2021 Round: 1<sup>st</sup>

Date Panel Dispatched: 16/1/21

Closing Date: 16/1/21

Date proficiency panel received: 14/1/21

Date of testing proficiency panel: 15/1/21

Date report sent: 16/1/21

**HUMAN IMMUNODEFICIENCY VIRUS TYPE-1 (HIV-1) ANTIBODY TESTING**

**NOTE :**

The HIV-1 performance evaluation samples are undiluted, unaltered individual donor material. It is the intention to provide laboratories with performance evaluation samples that closely resemble the types of specimens that laboratories encounter in their routine daily testing.

QAS Laboratory Identification No.:

--	--	--	--	--	--	--	--

(Number can be found on your panel box)

Laboratory Name: Rural Hospital Chakan, ICTC

Type of Laboratory:  NRL/SRL/ICTC/PPTCT/FI-ICTC/PPP-ICTC/Blood Bank/Private lab

Address of Laboratory (where testing is undertaken): Chakan, Tal-Chhed

Street: Pune, opposite S.T. Stand

State: Maharashtra Postal Code: 410501

Telephone No. \_\_\_\_\_

E-mail: puneictrhchakan@gmail.com

Fax No.: \_\_\_\_\_

Name of nodal officer:  
& Signature

Dr. Dhawle N. G.

वैद्यकीय अधीक्षक वर्ग-१  
ग्रामीण रुग्णालय, चाकण,  
ता. खेड, जि. पुणे



EQAS Lab ID. No.

--	--	--	--	--	--	--	--

KIT DETAILS

	Kit Details		
	Test I	Test II	Test III
Type of Test (Rapid/ELISA/Western Blot)	Rapid	Rapid	Rapid
Name of the test	Combaids	Meriscxen	Fredbo
Manufacturer	AREAY	Meril Diagnosis	AREAY
Lot #	4000023271	ME012021	ME012022
Date of Expiry	4/3/2022	06/2021	12/2021
Principle of the test	chromatography	Immuno dot assay	

FINAL RESULTS

S.No.	Panel Sample ID	Test Results			Final Interpretation
		Test I	Test II	Test III	
1	A	R	R	R	Reactive
2	B	NR	—	—	Negative
3	C	NR	—	—	Negative
4	D	NR	—	—	Negative

R : Reactive ; NR : Nonreactive; P: Positive; I : Indeterminate ; N : Negative

Remarks (if any) : —

Signature of Technician

Date: 16/1/2021

Signature of Nodal Officer

Date:

वैद्यकीय अधीक्षक वर्ग-१  
ग्रामीण रुग्णालय, चाक्रण,  
ता. खेड, जि. पुणे

External Quality Assessment Scheme for HIV Serology

EQAS REPORT (SRL to ICTCs)

Round: II(2020)

Name of Proficiency Testing Provider: NARI, Pune

Date Panel Distributed: 11/01/2021

Name of the participating ICTC: RH Chakan, Pune

Date of result Received: 13/01/2021

Sample Code	Standard Result	Results obtained by participating laboratory	% Matching
A	R	R	100%
B	NR	NR	100%
C	NR	NR	100%
D	NR	NR	100%

Keyword: R: Reactive, NR: Non reactive

Remarks: Performance is Satisfactory

*Dr. Sae Pol*  
L.T.

*[Signature]*  
Incharge,  
ICTC & SRL  
Dept. of Microbiology  
BJMC Pune

**Dr. Sae Pol**

Name of Laboratory Incharge:

Signature of Laboratory Incharge

**\*End of Report \*\***