


<b>MEASURE TECHNO LAB</b> <b>2, B.T. ROAD (JAYANTI CINEMA COMPLEX)</b> <b>BARRACKPORE, KOLKATA - 700120, W.B.</b> Phone : 033 - 2215 - 0032, 2215 - 9687, 8100875519, Mobile: 9831190974, LAB:- 8100143376, E-mail: measuretechno@yahoo.co.in	Form No. - MTL/22/2006	 NABL ACCREDITED LABORATORY Certificate No. CC - 2545
	<b>CALIBRATION CERTIFICATE OF MICRO PIPETTE</b>	
CALIBRATION CERTIFICATE NO.: MTL / SDC / R04 / 05 - 21	ULR - CC254521000009551F	Page: 1 of 1

1.0 Service Request No.: MTL / 11 / 05 / 21- 22

1.1 Issued to: M/s. Sun Diagnostic Centre,  
 IMA Building, Ranihat, Near SCB Medical College,  
 Cuttack, Odisha - 753007.

1.2 Description & Identification of item to be Calibrated:

a) Name:	Micro Pipette	b) Code No.:	SDC / LAB / PIPETTE / 02
c) Sl. No.:	YE6F725706	d) Make:	Dragon Lab
e) Model / Type:	Micropette	f) Range:	100 µl to 1000 µl
g) Sensor:	N.A.	h) Resolution:	5 µl
i) End User:	Pathology	j) Accuracy:	N.S.
k) Calibration done at:	On Site / √ In House		

1.3 Date of receipt of item : 10-05-21

1.4 Physical Condition of DUC : OK

1.5 Date of calibration : 11-05-21

1.6 Recommended date of next calibration : 11-05-22

1.7 Date of Issue : 12-05-21

1.8 Environmental Conditions During Calibration:

Temperature:	23 °C ± 2 °C
Humidity:	50 % RH ± 10 % RH
Pressure:	1012.5 mbar

1.9 Method of Calibration: SOP / MASS / 02 ( As Per ISO : 8655 - 6 : 2002)

2.0 Traceability :

- a) Standards used for calibration are traceable to National standards through NABL Accredited Laboratory.  
 b) The following standards / Equipment have been used.
- i) 1 mg to 200 g Weight Box (E1) Cal. Certificate No. WI / May / 19 / 010 (WEIGH INDIA, New Delhi) (Cal. Date: 17/05/19, Due Date: 17/05/22)
  - ii) Data Acquisition Switch Unit Cal. Certificate No. JRPM - CCTR - ET - 2020 - 0795 (JRPM, Chennai) (Cal. Date: 01/09/20, Due Date: 31/08/21)
  - iii) RTD (PT - 100) Cal. Certificate No. TL / 020 / 891.2.1 (TEMPSSENS, Udaipur) (Cal. Date: 27/10/20, Due Date: 26/10/21)

2.1 Result :

Mechanical Calibration

Sl. No.	Parameter/ Range	Nominal Value µl	Mass of Water mg	Volume of Water at 20 °C µl	Error µl	Measurement Expanded Uncertainty ± µl
1.	Volume	100	99.5830	99.9813	-0.0187	
2.	100 µl to 1000 µl	500	497.9670	499.9589	-0.0411	0.39
3.		1000	995.9380	999.9218	-0.0782	

- Remarks: i) Cubical Expansion co - efficient of pipette material taken as  $10^{-5} \mu\text{l} / ^\circ\text{k}$ .  
 ii) This result has an expanded uncertainty with a coverage factor  $k=2$  at approximately 95% confidence level.  
 iii) The calibration certificate issued for this instrument is to be used for scientific or industrial purposes only.

DUC - Device Under Calibration


N.S. - Not Specified

N.A. - Not Applicable

Opinions and Interpretations

Calibrated	√	Accepted / Valid for use
Limited Use		Rejected / Out of use

Calibrated by:  
**Testing Engineer**  
 S. Mahapatra  
 Testing Engineer  
 Kolkata

Checked / Approved by:  
  
 Quality & Technical Manager  
 S. Pandey



YE6F725706

MEASURE TECHNO LAB  
2, B. T. Road, Barrackpore, Kol-120  
Nomenclature: *MICRO PIPETTE*  
Code No: *102-101122*