



CALIBRATION REPORT

To,

Sadbhavana Pathology Lab,
Pune.

Dear Sir,

As desired by you please find enclosed herewith the Calibration report

1. Instrument: Bio-chemistry Analyzer EC-5plusV2 S.N- **S200619**
2. Range of the instrument: Optical density Linear form 0-2.5
3. Date of Calibration - 19/04/2021
4. Next Calibration Due- 18/04/2022

Test Voltages / Test Temperature -

Test Voltage	Actual Voltage	Ok / Not Ok
Logic 5 V	4.9V	OK
Lamp Voltage (11.5V – 12V)	11.90V	OK
SMPS 16.5 -18.5	17.8	OK
Flowcell Temperature	25°C,30°C,37°C ,OFF	ALL OK

Sr.No.	Filters	Pre Channel	Log Channel
1	340	40295	0.6025
2	405	70536	0.4622
3	450	65846	0.5071
4	505	64125	0.5125
5	546	71332	0.4165
6	578	82713	0.3286
7	630	71111	0.4155
8	670	78011	0.3555

Thanking you and assuring of our prompt attention at all times.

Yours faithfully,

For TRANSASIA BIO-MEDICALS LTD.

Rajesh Wankhede
Dy. Regional Service Manager

