

GSTIN:- 08CIUPP8490B1ZL

SP DIAGNOSTIC

Sales & Service Medical & Laboratory Instruments.

CALIBRATION CERTIFICATE

Equipment Type : Hematology

Model : Nihon Kohden MEK 6400

Serial# : 10700

Calibration Date : 26-03-2021

Customer Location : Vishwakarma Pathology Lab

The final results are as follows:

Parameters	Target	Obtained
Adjusting the Amplifier Offset	$\pm 5mV$	3
WBC	900 - 1400	1174
RBC	900 - 1400	1189
HB	900 - 1400	958
HCT	900 - 1400	1285
PLATLET	900 - 1400	1100
LY%	900 - 1400	943
MO%	900 - 1400	1361
RDW	900 - 1400	1214
MPV	900 - 1400	1250

Note:

1. This certificate refers only to the particular item submitted for calibration.
2. The calibration result reported in the certificate is valid at the time of and under the stated condition of the measurement.

The instrument is working satisfactorily, subsequent to calibrations of the above parameters, and the next calibration is due on 25 / 03 / 2022

Service & Application Department

G-159 SECTOR-G SHASTRI NAGAR, NEAR ST. PAULS SCHOOL, JODHPUR (RAJ.) 342001

EMAIL: SPDIAGNOSTIC2704@YAHOO.COM

Mob: +91-8233194862, +91-7791958626

Press [Count] switch to measure hematology control for calibration.

	Cal
WBC	1125
RBC	1211
HGB	970
HCT	1146
MCV	—
MCH	—
MCHC	—
PLT	1100

Navigation buttons: Left arrow, Up arrow, Right arrow, Down arrow.

Numeric keypad with buttons: 7, 8, 9, 4, 5, 6, 1, 2, 3, 0, ., C, and an Enter button.



PREV and NEXT buttons.

OK, [unlabeled], and HISTORY buttons.