



UNMATCHED SERVICE
SINCE 1979...

Date: 21-DEC-2020

To,
Dhiraj General hospital,
Vadodra.

Calibration Certificate

We hereby certify that the Erba Hb-Vario HPLC Hemoglobin system Sr. No. HBV-D0137
installed in your lab is checked for calibration/validation and found instrument validation pass.

Also verified performance of following module parameters, found satisfactory.

Sr.No	Parameters	Measured Readings	Expected Readings	Tolerance	Remarks
1	Column Temperature	27.2deg C	27.3 deg C	+/-5%	Pass
3	Detector Reference	25670	25500	+/- 500	Pass
3	Detector Dark	995	1000	+/- 20	Pass

For,

Transasia Bio-Medicals Ltd.

Jalpen
Jalpen Kasarwala
Service Engineer
Vadodra