



## CALIBRATION REPORT

### Details of the Unit Under Calibration:

Calibration Certificate No: SBS/CL/21-22/02452

Page: 1 of 2

Customer Name:	SURIYAA DIAGNOSTIC CENTRE	Customer Address:	921, VELACHERY MAIN ROAD, CAMP ROAD JUNCTION, SELAIYUR, CHENNAI-600 073
Location:	LAB	Serial No:	YE3K088150
Description:	MICRO PIPETTE	Asset No:	NA
Make :	NA	Date of Calibration:	28.05.2020
Model:	NA	Date of Due:	27.05.2021
Range:	5-50 $\mu$ L	VARIABLE	


### Traceability Details:

S.No	Standard Used	Make	Model	Certificate No	Validity
1	Electronic Weighing Balance	A&D	GH-252	2020/0914-01	14-Dec-2021

### Environmental Conditions and Standard Operating Procedure Details:

Temperature	24°C	Calibration Done At	LAB
Humidity (40-70% RH)	55%	Calibration Procedure No	-

Calibrated by:

  
P.MYILSAMY  
(Calibration Engineer)

Authorized Signatory:



**SUNSHINE BIOMEDICAL SOLUTIONS**

No:68, First floor , Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087.



Result of Calibration

REPORT				
S.No	Units	Nominal Values	Measured Values	Tolerance
1	µL	10	9.9	0.5%
2	µL	20	19.9	0.5%
3	µL	30	30.1	0.5%
4	µL	40	40.1	0.5%
5	µL	50	50.2	0.5%

CALIBRATION RESULT:

- Corrective maintenance required
- Removed from use
- Acceptable for use

Calibration Engineer Sign: *GP*

Date: 28.05.2020



## CALIBRATION REPORT

### Details of the Unit Under Calibration:

Calibration Certificate No: SBS/CL/21-22/02453

Page: 1 of 2

Customer Name:	SURIYAA DIAGNOSTIC CENTRE	Customer Address:	921, VELACHERY MAIN ROAD, CAMP ROAD JUNCTION, SELAIYUR, CHENNAI-600 073
Location:	LAB	Serial No:	MG421738
Description:	MICRO PIPETTE	Asset No:	NA
Make :	NA	Date of Calibration:	28.05.2020
Model:	NA	Date of Due:	27.05.2021
Range:	100-1000 $\mu$ L	VARIABLE	


### Traceability Details:

S.No	Standard Used	Make	Model	Certificate No	Validity
1	Electronic Weighing Balance	A&D	GH-252	2020/0914-01	14-Dec-2021

### Environmental Conditions and Standard Operating Procedure Details:

Temperature	24°C	Calibration Done At	LAB
Humidity (40-70% RH)	55%	Calibration Procedure No	-

Calibrated by:

  
P.MYILSAMY  
(Calibration Engineer)

Authorized Signatory:




**SUNSHINE BIOMEDICAL SOLUTIONS**

No:68, First floor , Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087.

REPORT				
S.No	Units	Nominal Values	Measured Values	Tolerance
1	μL	100	99.9	0.5%
2	μL	200	199.9	0.5%
3	μL	500	500.1	0.5%
4	μL	700	700.2	0.5%
5	μL	1000	1000.3	0.5%

**CALIBRATION RESULT:**

- Corrective maintenance required
- Removed from use
- Acceptable for use

Calibration Engineer Sign: 

Date: 28.05.2020



## CALIBRATION REPORT

### Details of the Unit Under Calibration:

Calibration Certificate No: SBS/CL/21-22/02454

Page: 1 of 2

Customer Name:	SURIYAA DIAGNOSTIC CENTRE	Customer Address:	921, VELACHERY MAIN ROAD, CAMP ROAD JUNCTION, SELAIYUR, CHENNAI-600 073
Location:	LAB	Serial No:	NA
Description:	MICRO PIPETTE	Asset No:	NA
Make :	NA	Date of Calibration:	28.05.2020
Model:	NA	Date of Due:	27.05.2021
Range:	50 µL	FIXED	


### Traceability Details:

S.No	Standard Used	Make	Model	Certificate No	Validity
1	Electronic Weighing Balance	A&D	GH-252	2020/0914-01	14-Dec-2021

### Environmental Conditions and Standard Operating Procedure Details:

Temperature	24°C	Calibration Done At	LAB
Humidity (40-70% RH)	55%	Calibration Procedure No	-

Calibrated by:

  
**P.MYILSAMY**  
(Calibration Engineer)

Authorized Signatory:

  
**C. SHANMUGARAJ**  
(Technical Manager)

**SUNSHINE BIOMEDICAL SOLUTIONS**


No:68, First floor , Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087.

REPORT				
S.No	Units	Nominal Values	Measured Values	Tolerance
1	$\mu$ L	50	50.1	0.5%



CALIBRATION RESULT:

- Corrective maintenance required
- Removed from use
- Acceptable for use

Calibration Engineer Sign: 

Date: 28.05.2020