

# SIMCO CALIBRATION LABORATORY

(A Division of : Sharp Industrial Machinery Maintenance Co. Pvt. Ltd.)

10-3-74/27, Plot No.151, Street No:3, Teacher's Colony, East Marredpally, Secunderabad - 26. Telangana  
Tel 040 2773 2341,2773 2342, 2773 1510, Telefax : 040-2773 2330, Mob: 77299 91231, 99488 96802, 98480 46524  
E-mail : simco.hyd@gmail.com, www.simcocalibrationlaboratory.com



CC-2806

**SIMCO**<sup>TM</sup>  
AN ISO 9001 : 2015  
CERTIFIED COMPANY

## CALIBRATION CERTIFICATE

In accordance with ISO / IEC-17025 : 2017

F10-CC-03

Page : 1 of 1

<b>Certificate No. : SIMCO/2011/381/MVL/06</b>	<b>Issue Date : 18/11/2020</b>
<b>1. Customer Name &amp; Address:</b> M/s. St Theresa's Multi Speciality Hospital, Erragadda, Hyderabad.	<b>ULR- CC 2 8 0 6 2 0 2 0 0 0 2 3 1 7 F</b>
	Reference Date : 16/11/2020 Calibration Date : 17/11/2020 Calibration Due Date : 16/11/2021

### 2. Details of Instrument Under Calibration:

Description : Micro Pipette	ID.No : STH-BME-LAB-PIPET-006
Make : Micpette	Location : Microbiology
Range : 10-100 µl	

### 3. Details of Standard Instruments Used :

Instrument Used	Serial/ID. No.	Valid up to	Certificate No.
Semi Micro Balance	SL/PMM/SMB/01	31/05/2021	SIMCO/2006/192/MVL/02

### 4. Environmental Conditions:

Standard Temperature : (23±2)°C

Relative Humidity : (50±10)% Rh

Air Pressure : (900-1100)hpa

Actual Temperature : 24.3 °C

Actual Relative Humidity : 50.6 % Rh

Actual Air Pressure : (953)hpa

### 5. Calibration Procedure : SOP-MP-01

### 6. Mechanical Calibration (Mass & Volume)

### 7. Calibration Results:

Sl. No.	Instrument Reading (µl)	Measured Value (µl)	Error (µl)	Expanded Uncertainty in (Ue) ± (µl)
1	10	9.44	-0.56	0.07
2	50	49.37	-0.63	0.07
3	100	99.28	-0.72	0.07

### 8. Remarks:

- The instrument was received in good condition and was calibrated at lab.
- This certificate pertains only to the item calibrated.
- The calibration results reported in this certificate are valid at the time of and at the stated environmental conditions.
- The calibration interval is determined based on customer's requirements.
- The calibration is traceable to National standards as per traceability details given in the certificate.
- This calibration certificate shall not be reproduced in full, except with prior written approval of Managing Director, SIMCO Calibration Laboratory.
- This calibration certificate is meant for scientific and industrial purpose only.
- The NABL Symbol is used as per NABL guidelines in NABL-133.
- The Measurement Uncertainty is reported approximately at 95% confidence level with coverage factor  $k = 2$ .
- To use this instrument at other temperatures use the formula given below

$$V_{27} = V_T (1 - \gamma (t - 27))$$

where,  $V_T$  = Volume measured at temperature  $t^\circ\text{C}$  (ml),  $V_{27}$  = Volume measured at  $27^\circ\text{C}$  (ml)

$\gamma$  = coefficient of cubical expansion of Pipette tips (0.00024 /°C)

A. Pavan  
Calibrated by

N.V. Kameswara Rao  
Technical Manager  
AUTHORISED SIGNATORY