



CALIBRATION CERTIFICATE

In accordance with ISO / IEC-17025 : 2017

F10-CC-03

Page : 1 of 1

Certificate No. : SIMCO/2011/869/THS/08	Issue Date : 19/11/2020
1. Customer Name & Address: M/S. ST THERESA'S MULTI SPECIALITY HOSPITAL, Erragadda, Hyderabad.	ULR- CC280620000007992F
	Reference Date : 15/11/2020
	Calibration Date : 17/11/2020 Calibration Due Date : 16/11/2021

2. Details of Instrument Under Calibration:

Description : Refrigerator	Model : SRC 600-GL
Make : ----	
Range : 2 to 8 °C	
Serial No. : O3DM/FF2D27845	
ID.No : STH-BME-LAB-REF-003	

3. Detail of Standard Instruments Used :

Instrument Used	Serial/ID. No.	Valid up to	Certificate No.
Universal Data Logger With Sensors (8-Channel)	SL/SMT/DL/02	03/08/2021	SIMCO/2008/240/THL/01

4.Environmental Conditions:

Standard Temperature : (25±4)°C

Relative Humidity : (50±20)% Rh

Actual Temperature : 24.5 °C

Actual Relative Humidity : 55.6 % Rh

5.Calibration Procedure: SOP-TH-01

6.Thermal Calibration :

7.Calibration Results:

SL. No.	Set Temperature in (°C)	Standard Reading (°C)	UUC Reading (°C)	Deviation (°C)	Expanded Uncertainty in ± (°C)
1	2 to 8	4.9	5.3	0.4	0.65

7. Remarks:

- The instrument was received in good condition and was calibrated at Site.
- This certificate pertains only to the item calibrated.
- The calibration results reported in this certificate are valid at the time of and at the stated environmental conditions.
- The calibration interval is determined based on customer's requirements.
- The calibration is traceable to National standards as per traceability details given in the certificate.
- This calibration certificate shall not be reproduced in full, except with prior written approval of Managing Director, SIMCO Calibration Laboratory.
- This calibration certificate is meant for scientific and industrial purpose only.
- The NABL Symbol is used as per NABL guidelines in NABL-133.
- The Measurement Uncertainty is reported approximately at 95% confidence level with coverage factor $k = 2$
- Temperature Scale ITS-90

G.N.Reddy
Calibrated by

N.V.Kameswara Rao
Mr.N.V.Kameswara Rao
Technical Manager
AUTHORISED SIGNATORY