



CALIBRATION CERTIFICATE

In accordance with ISO / IEC-17025 : 2017

F10-CC-03

Page : 1 of 1

Certificate No. : SIMCO/2011/869/THS/09	Issue Date : 19/11/2020
1. Customer Name & Address: M/S. ST THERESA'S MULTI SPECIALITY HOSPITAL, Erragadda, Hyderabad.	ULR- CC280620000007993F
	Reference Date : 15/11/2020
	Calibration Date : 17/11/2020 Calibration Due Date : 16/11/2021

2. Details of Instrument Under Calibration:

Description	: Refrigerator
Make	: SAMSUNG
Range	: 2 to 8 °C
Serial No.	: FF2A2784S
ID.No	: STH-BME-LAB-REF-004

3. Detail of Standard Instruments Used :

Instrument Used	Serial/ID. No.	Valid up to	Certificate No.
Universal Data Logger With Sensors (8-Channel)	SL/SMT/DL/02	03/08/2021	SIMCO/2008/240/THL/01

4.Environmental Conditions:

Standard Temperature : (25±4)°C

Relative Humidity : (50±20)% Rh

Actual Temperature : 24.5 °C

Actual Relative Humidity : 55.6 % Rh

5.Calibration Procedure: SOP-TH-01

6.Thermal Calibration :

7.Calibration Results:

SL. No.	Set Temperature in (°C)	Standard Reading (°C)	UUC Reading (°C)	Deviation (°C)	Expanded Uncertainty in ± (°C)
1	2 to 8	4.6	5.1	0.5	0.65

7. Remarks:

- The instrument was received in good condition and was calibrated at Site.
- This certificate pertains only to the item calibrated.
- The calibration results reported in this certificate are valid at the time of and at the stated environmental conditions.
- The calibration interval is determined based on customer's requirements.
- The calibration is traceable to National standards as per traceability details given in the certificate.
- This calibration certificate shall not be reproduced in full, except with prior written approval of Managing Director, SIMCO Calibration Laboratory.
- This calibration certificate is meant for scientific and industrial purpose only.
- The NABL Symbol is used as per NABL guidelines in NABL-133.
- The Measurement Uncertainty is reported approximately at 95% confidence level with coverage factor $k = 2$
- Temperature Scale ITS-90

G.N. Reddy
Calibrated by

N.V. Kameswara Rao
Mr.N.V.Kameswara Rao
Technical Manager
AUTHORISED SIGNATORY