

SIMCO CALIBRATION LABORATORY

(A Division of : Sharp Industrial Machinery Maintenance Co. Pvt. Ltd.)

10-3-74/27, Plot No.151, Street No.3, Teacher's Colony, East Marredpally, Secunderabad - 26. Telangana
Tel 040 2773 2341,2773 2342, 2773 1510, Telefax : 040-2773 2330, Mob: 77299 91231, 99488 96802, 98480 46524
E-mail : simco.hyd@gmail.com, www.simcocalibrationlaboratory.com



CC-2806

SIMCOTM
AN ISO 9001 : 2015
CERTIFIED COMPANY

CALIBRATION CERTIFICATE

In accordance with ISO / IEC-17025 : 2017

F10-CC-03

Page : 1 of 1

Certificate No. : SIMCO/2011/381/MVL/05	Issue Date : 18/11/2020
1. Customer Name & Address: M/s. St Theresa's Multi Speciality Hospital, Erragadda, Hyderabad.	ULR- CC 2 8 0 6 2 0 2 0 0 0 2 3 1 6 F
	Reference Date : 16/11/2020 Calibration Date : 17/11/2020 Calibration Due Date : 16/11/2021

2. Details of Instrument Under Calibration:

Description : Micro Pipette	ID.No : STH-BME-LAB-PIPET-005
Make : Thermo Scientific	Location : Microbiology
Range : 100-1000 µl	
S.No : MM17650	

3. Details of Standard Instruments Used :

Instrument Used	Serial/ID. No.	Valid up to	Certificate No.
Semi Micro Balance	SL/PMM/SMB/01	31/05/2021	SIMCO/2006/192/MVL/02

4. Environmental Conditions:

Standard Temperature : (23±2)°C
Actual Temperature : 24.3 °C

Relative Humidity : (50±10)% Rh
Actual Relative Humidity : 50.6 % Rh

Air Pressure : (900-1100)hpa
Actual Air Pressure : (953)hpa

5. Calibration Procedure : SOP-MP-01

6. Mechanical Calibration (Mass & Volume)

7. Calibration Results:

Sl. No.	Instrument Reading (µl)	Measured Value (µl)	Error (µl)	Expanded Uncertainty in (Ue) ± (µl)
1	100	97.84	-2.16	0.13
2	500	497.12	-2.88	0.13
3	1000	998.85	-1.15	0.13

8. Remarks:

- The instrument was received in good condition and was calibrated at lab.
- This certificate pertains only to the item calibrated.
- The calibration results reported in this certificate are valid at the time of and at the stated environmental conditions.
- The calibration interval is determined based on customer's requirements.
- The calibration is traceable to National standards as per traceability details given in the certificate.
- This calibration certificate shall not be reproduced in full, except with prior written approval of Managing Director, SIMCO Calibration Laboratory.
- This calibration certificate is meant for scientific and industrial purpose only.
- The NABL Symbol is used as per NABL guidelines in NABL-133.
- The Measurement Uncertainty is reported approximately at 95% confidence level with coverage factor $k = 2$.
- To use this instrument at other temperatures use the formula given below
 $V_{27} = VT (1 - \gamma (t-27))$.
where, VT = Volume measured at temperature $t^{\circ}C$ (ml), V_{27} = Volume measured at $27^{\circ}C$ (ml)
 γ = coefficient of cubical expansion of Pipette tips (0.00024 /°C)

A. Pavani
Calibrated by

N.V. Kameswara Rao
N.V. Kameswara Rao
Technical Manager
AUTHORISED SIGNATORY