



LABCARE

Care for quality • Care for customer

## Labcare Bio-Medicals

### CALIBRATION CERTIFICATE

Customer Name	FATIMA HOSPITAL
Address	Mother Teresa Road Padri Bazar Gorakhpur

### CALIBRATION REPORT FOR

Equipment	Heamatology Analyzer
Model	MEK-6420P
Serial number	01252735

### CALIBRATION VALIDITY

Calibration performed date	27-12-2020
Calibration validity	One Year
Next Calibration Due on	December 2021

### CHECKS BEFORE CALIBRATION

#### PHYSICAL CHECKS:

Outer Casing/Chassis O.K.(No Damage)	Yes <input checked="" type="checkbox"/> / No <input type="checkbox"/>
Calibration validity	Yes <input checked="" type="checkbox"/> / No <input type="checkbox"/>
Next Calibration Due on	Yes <input checked="" type="checkbox"/> / No <input type="checkbox"/>



Care for quality • Care for customer

## Labcare Bio-Medicals

### PRFORMANCE CHECKS AFTER CALIBRATION


HGB Voltage	1.920
WBC Voltage	18.35
RBC Voltage	18.10
Diluent Voltage	1.51 V
Lysing Reagent	0.99 V
HGB Unit Temp. (Degree Cent.)	26.8
MC Unit (Degree Cent.)	24.9
Power Board (Degree Cent.)	26.6

#### Calibrator Details:-

- SC – CAL PLUS
- LOT NO-PLUS 0120
- EXP- 26.12.2021

After calibration Quality Control test for all level passed successfully.

Test results is coming OK	✓	Remarks	Done
---------------------------	---	---------	------

  
Service Engineer