



**EQAS CAPA
CMC VELLORE
BIOCHEMISTRY**

S. No.	Result Received Date	Cycle No.	Errors Incidental Description	Root cause Analysis	Corrective Measures Taken
1)	5/02/21	January 21	Value of Glucose not accepted Value of TSH not accepted	Random error Random error	Sample retested Retest value = 124.15 mg/dl Sample retested. Retest value = 1.94 μ IU/ml

AS
12/02/21
Incharge Clinical Biochemistry
People's Hospital
Bhanpur, Bhopal (M.P.)