



**EQAS CAPA
CMC VELLORE
BIOCHEMISTRY**

S. No.	Result Received Date	Cycle No.	Errors Incidental Description	Root cause Analysis	Corrective Measures Taken
1)	1/2/21	December 2020	Value of ALT not accepted	calibration required	calibration was done Sample retested Retest value = 168 IU/L

~~AS~~
4/1/12
Incharge Clinical Biochemistry
People's Hospital
Bhanpur, Bhopal (M.P.)