



**EQAS CAPA
CMC VELLORE
BIOCHEMISTRY**

S. No.	Result Received Date	Cycle No.	Errors Incidental Description	Root cause Analysis	Corrective Measures Taken
1)	5/02/21	January 21	Value of Glucose not accepted Value of TSH not accepted	Random error Random error	Sample retested Retest value = 124.15 mg/dL Sample retested. Retest value = 1.94 μ IU/ml

AS
12/02/21
Incharge Clinical Biochemistry
People's Hospital
Bhanpur, Bhopal (M.P.)



**PEOPLE'S
UNIVERSITY**

PEOPLE'S HOSPITAL

ASSOCIATED HOSPITAL OF PEOPLE'S COLLEGE OF MEDICAL
SCIENCE & RESEARCH CENTRE People's Campus Bhanpur ,
Bhopal (462037) M.P.India Phone: 0755-4005200,
E-mail: peoples@gmail.com

Registration No.: NH/031JUNE07



DEPARTMENT OF BIOCHEMISTRY

Patient Name	: Mr COBAS CPL PATHO	UHID	: PMH19214438
Age/Sex	: 2 Years 7 Months / Male	Referred By	: Dr A Self
IPD/OPD	: OPD	Ward	:
Reported On	: 12/02/2021 11:17 AM	Lab No	: 326543
Sample Collection	: 11/02/2021 04:00 PM	Ordered On	: 11/02/2021 03 29 PM

BIOCHEMISTRY

<u>TEST</u>	<u>RESULT</u>	<u>UNITS</u>	<u>REFERENCE RANGE</u>
BS			
Blood Sugar (R)	: 124.15	mg/dl	70-140


Barcode:
32654302



-- End of Report --

(Dr.Kavitarati Dharwadkar) (Dr.Ashish Jain) (Dr.Shrikant Chandrakar) (Dr.Archana Sharma)
(H.O.D.)

Technician Dharmendra Patel


Clinical Biochemist

Incharge Clinical Biochemistry
People's Hospital
Bhanpur, Bhopal (M.P.)