



**EQAS CAPA
CMC VELLORE
BIOCHEMISTRY**

S. No.	Result Received Date	Cycle No.	Errors Incidental Description	Root cause Analysis	Corrective Measures Taken
1)	1/2/21	December 2020	Value of ALT not accepted	calibration required	Calibration was done Sample retested Retest value = 168 IU/L

AS
4/1/12
Incharge Clinical Biochemistry
People's Hospital
Bhanpur, Bhopal (M.P.)



PEOPLE'S
UNIVERSITY

PEOPLE'S HOSPITAL

ASSOCIATED HOSPITAL OF PEOPLE'S COLLEGE OF MEDICAL
SCIENCE & RESEARCH CENTRE People's Campus Bhanpur,
Bhopal (462037) M.P. India Phone: 0755-4005200,
E-mail: peoples@gmail.com
Registration No.: NH/031JUNE07



DEPARTMENT OF BIOCHEMISTRY

Patient Name : Mr COBAS CPL PATHO UHID : PMH19214438
Age/Sex : 2 Years 6 Months / Male Referred By : Dr A Self
IPD/OPD : OPD Ward :
Reported On : 04/01/2021 12:50 PM Lab No : 318805
Sample Collection : 04/01/2021 12:40 PM Ordered On : 04/01/2021 12:37 PM

BIOCHEMISTRY

TEST	RESULT	UNITS	REFERENCE RANGE
SGPT	: 168	IU/L	5-40

Barcode:
31880502



-- End of Report --

Calibration of ALT was done.
EQAS sample retested.
Result obtained on retesting 168 IU/L.

(Dr. Kavitarati Dharwadkar) (Dr. Ashish Jain) (Dr. Shrikant Chandrakar) (Dr. Archana Sharina)
(H.O.D.)

Technician Diya Jadon

AS
Clinical Biochemist
Incharge Clinical Biochemistry
People's Hospital
Bhanpur, Bhopal (M.P.)

04-01-2021 11:38

Test	Abs.	Abs. Reagent	Kinetic blank limit	Initial abs.	Abs. main filter	Blank abs. limit	Remarks	Date
ALT-GPT	0.0012			1.0797		1.4000	Initial Blank Abs < Blank Abs limit	04-01-2021 08:37

Test	Type Name (Lot)	Kit Abs. No.	Theor. Conc.	Unit	Factor	Factor limits	Remarks	Date
ALT-GPT	SER BIOCALIBRATOR N (071XA)	0.0324	92	U/L	2948.264			04-01-2021 08:41
ALT-GPT	SER BIOCALIBRATOR N (071XA)	0.0343	92	U/L	2782.083			04-01-2021 11:38

~~AS~~
 4/01
 Incharge Clinical Biochemist
 People's Hospital
 Bhanpur, Bhopal (M.P.)