

## ELVIN RESEARCH CEN

Chhalera, Street No. 3, Sector-44, Noida. +91 8287612454 | +91 8287600574

krccalibration@gmail.com | info@kelvin.org.in

www.kelvin.org.in

## CALIBRATION CERTIFICATE

User Name: M/s

PEOPLE'S HOSPITAL

PEOPLE'S CAMPUS, BHANPUR, BHOPAL(M.P)-462037 (INDIA)

Page No.: 1 of 1

Unit Under Calibration	REFRIGERATOR (TEMPERATURE METER)		C
Certificate No.	L'DC POR CONTROL (TEMPERATURE METER)	Make / Model No.	HTC-1 /
	KRC/2020/GH/002-097	Range/ Size	2~8 °C
Service Request Date	12-Mar-2020		
Calibration Date	1-Jul-2020	Least Count	0.1 °C
Suggested Due Date		UUC Serial No.	
	30-Jun-2021	UUC I.D. No.	PH/CPL P/RFT/1
Date Of Issue	2-Jul-2020		
Discipline		Location	CENTRAL PATHOLOGY LAB (PATHOLOGY)
Discipline	MEDICAL	Visual Inspection	OK

Revelant Standard IEC 60601-2-21:2009 **Calibration Procedure** KRC/CPM/7.2-18-H-WI Performed At Site

**ENVIRONMENTAL CONDITIONS** 

The second secon			
Temperature	25 ± 4 °C	Humidity	$50 \pm 20 \% RH$

MASTER INSTRUMENT USED FOR CALIBRATION

Instrumnet Name	Make	Serial / I.D. No.	Traceability	ULR No.	Suggested Due Dat
4-wire PRT with Indicator	Fluke	33010023	SWASTIK	CC305219000000 182F	24-Nov-2020

## CALIBRATION RESULTS IN °C.

^	CALIBRATION RESULTS IN °C			
Ura	Master Average Reading (X)	UUC Average Reading (Y)	Error = Y-X	
E.	2.00	2.0	0.00	
10	3.03	3.0	-0.03	
	4.09	4.0	-0.09	
	5.11	5.0	-0.11	
	6.16	6.0	-0.16	
	8.33	8.0	-0.33	

**Expanded Uncertainty** 

Uncertainty of Measurement: The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2 such that the coverage probability corresponds to approximately 95%.

## Note:-

- 1. The calibration results reported in this certificate are valid at the time of and the stated condition of measurement.
- 2. The results reported relate only to the above calibrated item.
- 3. This report should not be reproduced except in full without our prior permission in writing.
- 4. Calibration certificate without signature are not valid.
- 5. Rdg. and UUC stands for reading and unit under calibration respectively.
- 6. This Certificate is in non nabl.



\*END OF REPORT\*