

Date:01/07/2020

CALIBRATION REPORT

Customer Name: FATIMA HOSPITAL

Model:Fully Automated Bio-Chemistry Analyzer EM-200

Serial No.:B2001646

Calibration done date: 01/07/2020

Next Calibration due Date: 01/07/2021

Lab In charge: Dr. Aditi Kumar

*This is to certify that above mentioned product has been verified of calibration
By
Checking Lamp Voltage 11.8V*

Checking gain of Photo meter for all wavelength values obtained are as under:

Filter	340	405	505	546	578	600	660	700
Gain	676	547	455	462	475	480	488	490

Checked calibration of probe on all locations.

*Checked temperature at Reagent tray at 8.0 degree (Range 4 to 11 degree) and
reaction tray 37.2 degree (Range 36.5 to 37.5 Degree)*

*Followed by calibration of all the three types of chemistries (picked up
CHOL, CREAT and SGPT) and QC run with satisfactory values of QC*

Calibration at site performed by Mr. Shivarchan Pandey



Authorized Signatory
(Mr. Shivarchan Pandey)
Application Specialist
For Transasia Bio-Medicals Ltd.