

LAB MONTHLY SUMMARY

Lab Name **DR B LAL CLINICAL LABORATORY PVT LTD** Lab No **13559**
 Month **December** Year **2020**
 Constituent Group **HbA1c**

Date of Result Entered : **16/12/2020**Date of Report Published : **31/12/2020**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	HbA1c	Other Immunoassay	Any Analyser	143	8.41	12.22	1.03	6 %	-2.34	0.17

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ± 1.0 to ± 2.0	Good.
Between ± 2.0 to ± 3.0	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

Page 1 of 1

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator
 Christian Medical College, Vellore

***** End of Report *****