



**PROFICIENCY TESTING REPORT**  
**ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME**  
 NABL accredited program as per ISO/IEC 17043:2010 standard  
 Organized By Department of Hematology, AIIMS, New Delhi-110029



*Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens*

EQAP CODE No. : 3650

Distribution No.: 151-J

Month/Year: December/2020

Instrument ID: ERBA ELITE 580 (K110519122060)

Name & Contact No. of PT Co-ordinator: Dr. Seema Tyagi (Prof.), Hematology, AIIMS, Delhi,  
 Tel: 9013085730 , E-Mail : accuracy2000@gmail.com

Date of issue &amp; status of the report: 11-01-2021[Final].

### CBC and Retic Assessment

Test Parameters	S.No.	Among Lab (Accuracy Testing)						Within Lab (Precision Testing)			
		Your Result 1	Your Result 2	Your Results Sum of 2 Value	Consensus result sum of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score	Yours Results Diff. of 2 Values	Consensus Result Diff. of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score
WBC x10 <sup>3</sup> /µl	1	4.37	4.35	8.72	8.4	0.0310	0.49	0.02	0.1	0.0110	-0.64
RBC x10 <sup>6</sup> /µl	1	5.17	5.13	10.3	10.21	0.0160	0.23	0.04	0.04	0.0350	0.00
Hb g/dl	1	9.5	9.4	18.9	19	0.0310	-0.17	0.1	0.1	0.0080	0.00
HCT%	1	29.6	29.4	59	60.2	0.1690	-0.31	0.2	0.3	0.0330	-0.22
MCV-fl	1	57.4	57.2	114.6	118.4	0.2120	-0.82	0.2	0.3	0.0270	-0.22
MCH-Pg	1	18.3	18.2	36.5	37.4	0.0680	-0.52	0.1	0.2	0.0120	-0.67
MCHC-g/dl	1	31.9	31.9	63.8	63.5	0.1590	0.09	0	0.3	0.0260	-0.67
Plt. x10 <sup>3</sup> /µl	1	218	212	430	417.5	2.09	0.25	6	8.5	0.65	-0.26
Retic %	2	13.2	12.2	25.4	17.4	0.41	0.85	1	0.6	0.05	0.77

### P.S . Assesment

YOUR REPORT			CONSENSUS REPORT		
DLC%	3	Nrbcs=01/100 WBC , Poly=04 L=06, E=00, Mono/Promono=04 , B1=86 P.M.=00, Mye=00, Meta=00, Other=00	Blasts: 70-80, Lymph: 5-15, Poly: 2-5, nRBC/Eo/Mono/Pro/My/Meta: 0-5		
RBC Morphology	3	Nc/Nc	Predominantly: Normocytic Normochromic, Moderate: Anisocytosis, Mild: Microcytic.		
Diagnosis	3	ACUTE LEUKEMIA	Acute Leukemia (Lymphoblastic).		

EQAP Code Distribution No.: 151- Month/Year:  
No.: 3650 J December/2020

Instrument ID: ERBA ELITE 580  
(K110519122060)

**COMBINED DATA VALUES OF TOTAL PARTICIPANTS**

Test parameters	S.No.	Total participants covered in the current dist.	Total No. responded	% of Labs with Z Score 0-2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
				Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
WBC x10 <sup>3</sup> /µl	1	350	224	82.59	80.8	3.13	4.02	13.84	14.29
RBC x10 <sup>6</sup> /µl	1	350	224	88.39	89.29	4.02	3.57	6.7	6.25
Hb g/dl	1	350	224	82.59	91.07	6.7	3.57	10.27	4.91
HCT%	1	350	224	89.29	89.73	5.8	2.23	4.02	7.14
MCV-fl	1	350	224	87.05	90.18	6.7	4.02	5.36	4.91
MCH-Pg	1	350	224	86.61	85.27	7.14	4.46	5.36	9.38
MCHC-g/dl	1	350	224	88.39	87.05	8.48	6.25	2.23	5.8
Plt. x10 <sup>3</sup> /µl	1	350	224	93.3	93.3	3.57	3.13	2.23	3.13
ReticCount%	2	350	199	93.97	81.91	4.52	16.08	1.51	4.02
PS Assessment	3	350	209	Acceptable:81.9%,Warning Signal:8.6%,Unacceptable :9.5%					

**\*Comments:**

- 1). **Among Lab (EQA) : Results acceptable.**
- 2). **Within Lab (IQA) : Precision acceptable.**

**Note-1: EQA** (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.

**IQA** ( Internal Quality Assurance) : Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

**Note-2:** Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values - Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (IQA)= (Your Result Difference of two values - Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR = 0.7413 x IQR

**Note-3:** Z score 0 to ±2: Acceptable, Z score ±2 to ±3 :Warning Signal, Z score > ±3 : Unacceptable [As per ISO/IEC 13528:2015 standard]

**Note-4:** Z score value between "0 to ±2" are texted in green colour. Z score value between "±2 to ±3" are texted in orange colour. Z score value > ±3 are texted in red colour.

**Note-5:** Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value (0.3\*SDPA). To pass the stability test, average difference in measurement values of first and last day sample ( $\bar{x}-\bar{y}$ ) should be smaller than the check value (0.3\*SDPA).

**Note-6:** ISHTM-AIIMS-EQAP does not subcontract any task of its scheme

**Note-7:** Participants are free to use methods/analyzer of their own choice.

**Note-8:** Proficiency testing (PT ) samples are sent quarterly to each participant.

**Note-9:** All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme's website [www.ishtmaiimseqap.com](http://www.ishtmaiimseqap.com).

Report authorized by,



Dr. Seema Tyagi (Prof.)  
PT Co-ordinator: ISHTM-AIIMS-EQAP  
Department of Hematology, AIIMS, New Delhi

-----End Of Report-----