

ALP-AMP (U/L) BA 200 FULLY AUTOMATIC ANALYZER

Class	Name	Type	No.	Abs.	Result	Unit	Factor	Date
Blank	ALP-AMP			-0.0096		U/L		
			1	-0.0096		U/L		
Calibrator	BDPL CAL	SER		0.0529	207	U/L	3310.4550	
			1	0.0529	207	U/L		
Control	BIOSYSTEMS L1098	SER		0.0287	127	U/L		
			1	0.0287	127	U/L		
Control	BIOSYSTEMS L2088	SER		0.0878	322	U/L		
			1	0.0878	322	U/L		

ALT-GPT (U/L) BA 200 FULLY AUTOMATIC ANALYZER

Class	Name	Type	No.	Abs.	Result	Unit	Factor	Date
Blank	ALT-GPT			-0.0001		U/L		
			1	-0.0001		U/L		
Calibrator	BDPL CAL	SER		0.0360	88	U/L	2425.0700	
			1	0.0360	88	U/L		
Control	BIOSYSTEMS L1098	SER		0.0168	41	U/L		
			1	0.0168	41	U/L		
Control	BIOSYSTEMS L2088	SER		0.0583	142	U/L		
			1	0.0583	142	U/L		

ASO (IU/mL)

Class	Name	Type	No.	Abs.	Result	Unit	Factor	Date
Blank	ASO			0.4153		IU/mL		
			1	0.4153		IU/mL		
Calibrator	ASO CAL	SER		0.6463	249.00	IU/mL	1077.7810	
			1	0.6463	249.00	IU/mL		

AST-GOT (U/L) BA 200 FULLY AUTOMATIC ANALYZER

Class	Name	Type	No.	Abs.	Result	Unit	Factor	Date
Blank	AST-GOT			0.0010		U/L		
			1	0.0010		U/L		
Calibrator	BDPL CAL	SER		0.0614	125	U/L	2069.4130	
			1	0.0614	125	U/L		
Control	BIOSYSTEMS L1098	SER		0.0244	48	U/L		
			1	0.0244	48	U/L		
Control	BIOSYSTEMS L2088	SER		0.0885	181	U/L		
			1	0.0885	181	U/L		

BILI DIRECT DPD (mg/dL) BA 200 FULLY AUTOMATIC ANALYZER

Class	Name	Type	No.	Abs.	Result	Unit	Factor	Date
Blank	BILI DIRECT DPD			0.0017		mg/dL		
			1	0.0017		mg/dL		
Calibrator	BDPL CAL	SER		0.0264	0.79	mg/dL	31.9466	
			1	0.0264	0.79	mg/dL		
Control	BIOSYSTEMS L1098	SER		0.0264	0.79	mg/dL		
			1	0.0264	0.79	mg/dL		
Control	BIOSYSTEMS L2088	SER		0.0355	1.08	mg/dL		
			1	0.0355	1.08	mg/dL		

BILI TOTAL DPD (mg/dL) BA 200 FULLY AUTOMATIC ANALYZER

Class	Name	Type	No.	Abs.	Result	Unit	Factor	Date
Blank	BILI TOTAL DPD			0.0014		mg/dL		
			1	0.0014		mg/dL		
Calibrator	BDPL CAL	SER		0.0592	3.58	mg/dL	61.9837	
			1	0.0592	3.58	mg/dL		
Control	BIOSYSTEMS L1098	SER		0.0356	2.11	mg/dL		
			1	0.0356	2.11	mg/dL		
Control	BIOSYSTEMS L2088	SER		0.0970	5.92	mg/dL		
			1	0.0970	5.92	mg/dL		

CHOL HDL DIRECT (mg/dL) BA 200 FULLY AUTOMATIC ANALYZER

Class	Name	Type	No.	Abs.	Result	Unit	Factor	Date
Blank	CHOL HDL DIRECT			0.0066		mg/dL		
			1	0.0066		mg/dL		
Calibrator	HDL/LDL	SER		0.1418	50.90	mg/dL	376.5029	
			1	0.1418	50.90	mg/dL		

CHOL LDL DIRECT (mg/dL) BA 200 FULLY AUTOMATIC ANALYZER

Class	Name	Type	No.	Abs.	Result	Unit	Factor	Date
Blank	CHOL LDL DIRECT			0.0017		mg/dL		
			1	0.0017		mg/dL		
Calibrator	HDL/LDL	SER		0.1706	98.60	mg/dL	583.7032	
			1	0.1706	98.60	mg/dL		

CHOLESTEROL (mg/dL) BA 200 FULLY AUTOMATIC ANALYZER
--

Class	Name	Type	No.	Abs.	Result	Unit	Factor	Date
Blank	CHOLESTEROL			0.0538		mg/dL		
			1	0.0538		mg/dL		
Calibrator	BDPL CAL	SER		0.3720	203	mg/dL	638.0403	
			1	0.3720	203	mg/dL		
Control	BIOSYSTEMS L1098	SER		0.3029	159	mg/dL		
			1	0.3029	159	mg/dL		
Control	BIOSYSTEMS L2088	SER		0.4538	255	mg/dL		
			1	0.4538	255	mg/dL		

CK (U/L)

Class	Name	Type	No.	Abs.	Result	Unit	Factor	Date
Blank	CK			-0.0003		U/L		
			1	-0.0003		U/L		
Calibrator	BDPL CAL	SER		0.0811	376	U/L	4624.1860	
			1	0.0811	376	U/L		
Control	BIOSYSTEMS L1098	SER		0.0456	212	U/L		
			1	0.0456	212	U/L		
Control	BIOSYSTEMS L2088	SER		0.1238	574	U/L		
			1	0.1238	574	U/L		

CREATININE (mg/dL) BA 200 FULLY AUTOMATIC ANALYZER

Class	Name	Type	No.	Abs.	Result	Unit	Factor	Date
Blank	CREATININE			0.0021		mg/dL		
			1	0.0021		mg/dL		
Calibrator	BDPL CAL	SER		0.1094	3.72	mg/dL	34.6621	
			1	0.1094	3.72	mg/dL		
Control	BIOSYSTEMS L1098	SER		0.0599	1.64	mg/dL		
			1	0.0599	1.64	mg/dL		
Control	BIOSYSTEMS L2088	SER		0.1520	4.83	mg/dL		
			1	0.1520	4.83	mg/dL		

GLUCOSE (mg/dL) BA 200 FULLY AUTOMATIC ANALYZER
--

Class	Name	Type	No.	Abs.	Result	Unit	Factor	Date
Blank	GLUCOSE			0.0300		mg/dL		
			1	0.0300		mg/dL		
Calibrator	BDPL CAL	SER		0.6298	179	mg/dL	298.4333	
			1	0.6298	179	mg/dL		
Control	BIOSYSTEMS L1098	SER		0.3263	88	mg/dL		
			1	0.3263	88	mg/dL		
Control	BIOSYSTEMS L2088	SER		0.9152	264	mg/dL		
			1	0.9152	264	mg/dL		

PROTEIN TOTAL (g/dL) BA 200 FULLY AUTOMATIC ANALYZER

Class	Name	Type	No.	Abs.	Result	Unit	Factor	Date
Blank	PROTEIN TOTAL			-0.0822		g/dL		
			1	-0.0822		g/dL		
Calibrator	BDPL CAL	SER		0.1028	6.10	g/dL	32.9793	
			1	0.1028	6.10	g/dL		
Control	BIOSYSTEMS L1098	SER		0.0705	5.04	g/dL		
			1	0.0705	5.04	g/dL		
Control	BIOSYSTEMS L2088	SER		0.1696	8.30	g/dL		
			1	0.1696	8.30	g/dL		

UREA-BUN-UV (mg/dL) BA 200 FULLY AUTOMATIC ANALYZER
--

Class	Name	Type	No.	Abs.	Result	Unit	Factor	Date
Blank	UREA-BUN-UV			0.0112		mg/dL		
			1	0.0112		mg/dL		
Calibrator	BDPL CAL	SER		0.1281	79	mg/dL	673.3076	
			1	0.1281	79	mg/dL		
Control	BIOSYSTEMS L1098	SER		0.0971	58	mg/dL		
			1	0.0971	58	mg/dL		
Control	BIOSYSTEMS L2088	SER		0.2071	132	mg/dL		
			1	0.2071	132	mg/dL		

URIC ACID (mg/dL) BA 200 FULLY AUTOMATIC ANALYZER

Class	Name	Type	No.	Abs.	Result	Unit	Factor	Date
Blank	URIC ACID			0.0488		mg/dL		
			1	0.0488		mg/dL		
Calibrator	BDPL CAL	SER		0.2543	7.95	mg/dL	38.6971	
			1	0.2543	7.95	mg/dL		
Control	BIOSYSTEMS L1098	SER		0.1967	5.72	mg/dL		
			1	0.1967	5.72	mg/dL		
Control	BIOSYSTEMS L2088	SER		0.3083	10.04	mg/dL		
			1	0.3083	10.04	mg/dL		