

LAB MONTHLY SUMMARY



Lab Name **DISTRICT MALE HOSPITAL, FATEHPUR** Lab No **14773**
 Month **May** Year **2021**
 Constituent Group **Chemistry I**



Date of Result Entered : 17/05/2021

Date of Report Published : 05/06/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE	GOD-POD	Any Analyser (Automation / Semi Automation)	657	212.15	9.39	19.91	187.41 mg/dl	-1.24	1.55
2	UREA	Urease UV / GLDH	Any Analyser (Automation / Semi Automation)	545	76.44	11.22	8.58	79.7 mg/dl	0.38	0.73
3	CREATININE	Jaffes Kinetic - Alkaline picrate	Any Analyser (Automation / Semi Automation)	444	3.12	12.76	0.40	3.5 mg/dl	0.95	0.04
4	T.BILIRUBIN	Others (DPD, Vanadate Oxidation)	Any Analyser (Automation / Semi Automation)	130	2.57	13.82	0.36	2.1 mg/dl	-1.32	0.06
5	T-PROTEIN	Biuret - Colorimetric	Any Analyser (Automation / Semi Automation)	602	5.29	10.14	0.54	5.57 g/dl	0.52	0.04
6	ALBUMIN	BCG - colorimetric	Any Analyser (Automation / Semi Automation)	606	3.20	9.49	0.30	3.7 g/dl	1.64	0.02
7	CALCIUM	Arsenazo III	Any Analyser (Automation / Semi Automation)	458	9.91	8.25	0.82	10.4 mg/dl	0.60	0.08
8	URIC ACID	Enzymatic / Uricase Colorimetric	Any Analyser (Automation / Semi Automation)	495	4.77	16.15	0.77	6.39 mg/dl	2.1	0.07
9	CHOLESTEROL	CHOD-PAP	Any Analyser (Automation / Semi Automation)	616	110.22	10.68	11.78	119.47 mg/dl	0.79	0.95
10	ALP	PNP DEA kinetic	Any Analyser (Automation / Semi Automation)	211	567.05	19.49	110.50	543.56 U/L	-0.21	15.21

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ±1.0 to ±2.0	Good.
Between ±2.0 to ±3.0	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator
 Christian Medical College, Vellore

***** End of Report *****