

LAB MONTHLY SUMMARY



Lab Name **CHITTARANJAN NATIONAL CANCER INSTITUTE** Lab No **14009**
 Month **January** Year **2021**
 Constituent Group **Chemistry II**



Date of Result Entered : 11/01/2021

Date of Report Published : 05/02/2021

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	HEXOKINASE II	Beckman AU480/680/5800/DXC700AU	17	120.51	5.52	6.65	118 mg/dl	-0.38	3.23
2	UREA II	Urease UV / GLDH II	Beckman AU480/680/5800/DXC700AU	27	32.07	5.99	1.92	33 mg/dl	0.48	0.74
3	CREATININE II	Jaffes End point II	Any Analyser (Automation / Semi Automation)	532	1.32	12.99	0.17	1.4 mg/dl	0.47	0.01
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Beckman AU480/680/5800/DXC700AU	26	2.17	6.72	0.15	2.1 mg/dl	-0.48	0.06
5	T-PROTEIN II	Biuret - Colorimetric II	Beckman AU480/680/5800/DXC700AU	22	5.18	3.30	0.17	5.4 g/dl	1.29	0.07
6	ALBUMIN II	BCG - Colorimetric II	Beckman AU480/680/5800/DXC700AU	24	3.19	5.86	0.19	3.2 g/dl	0.05	0.08
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Beckman AU480/680/5800/DXC700AU	23	5.20	6.79	0.35	5.3 mg/dl	0.28	0.15
8	CHOLESTROL II	CHOD-PAP II	Beckman AU480/680/5800/DXC700AU	25	113.04	3.13	3.53	103 mg/dl	-2.84	1.41
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Beckman AU480/680/5800/DXC700AU	26	137.41	7.14	9.81	124 mg/dl	-1.37	3.85

SDI Range	Interpretation
Within -1.0 to +1.0	Excellent.
Between ±1.0 to ±2.0	Good.
Between ±2.0 to ±3.0	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

Contact details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Dr. Pamela Christudoss
 CMC EQAS Co-Ordinator
 Christian Medical College,
 Vellore

***** End of Report *****