

CUSTOMER / BRANCH: M/s. Fatima Laboratory , Fatima Hospital, Gorakhpur

DISCRIPTION OF MODEL: FREEZER

MODEL: RPF-224 SR.NO: ZDBP-04788

CERTIFICATE NO: REMI/0182/2020-21

CALIBRATION DATE: 08/02/2021

CALIBRATION VALIDITY: 07/08/2021

**DETAILS OF EQUIPMENT USED FOR CALIBRATION**

Description	Certified by	Certificate No.	Calibration Date	Validity
Digital Temperature Indicator	Classic Instrumentation Pvt. Ltd.	CI/2K19-2K20/05/TH005	09/07/2020	08/07/2021

**OBSERVATIONS**

Sr. No.	Instrument Temp. Displayed (°C)	Indicator Temp. (°C)	Temperature Deviation	Calibration Date	Remark
1.	-40°C	-40.4°C	0.4°C	08/02/2021	O.K

- Remark:
1. Calibration Traceable to national standards.
  2. Overall Deviation is within specified accuracy.
  3. Result are related only to the item calibrated.
  4. Maximum tolerance permissible is  $\pm 1^{\circ}\text{C}$  &  $\pm 5^{\circ}\text{C}$  whichever is applicable & RPM 1.5% - 2% FSD. &  $\pm 5\text{g.}$  to set weight.

Calibrated by: Mr. Vineet Kumar

**SALES & SERVICE BRANCHES :**

Ahmedabad, Bengaluru, Kolkata, Chennai, Hyderabad, Nagpur, New Delhi, Kochi, Kanpur, Indore, Goa