

Calibration Certificate

Certificate No.:	CCTPL/28-01/25	Field:	Mechanical
SRF No.:	CCTPL/0101/01	Date of Receipt:	03/02/2021
SRF Date:	03/02/2021	Date of Calibration:	04/02/2021
Company Name & Address	M/s. Fatima Hospital	Recommended Due Date.:	03/02/2022
	Mother Taresa Road, Padri Bazar, Gorakhpur. Pin - 273014	As per agreed by customer	
		Certificate Issue Date:	04/02/2021

DUC* Details

Instrument Name:	DTC With Sensor (Incubator)	Instrument Sr. No.:	----
Make:	Ambassador	Instrument Id No.:	-----
Model No.:	----	Location:	Lab
Range:	0 - 99.9 °C	DUC Condition:	OK
Readability:	0.1 °C	Calibration Performed at:	At Site

Standard Equipments Used(Traceable to National Standard)

Sr. No	Instrument Name	Cert. No	Cal. By	Calibration Due Date
1	RTD with Indicator	CCTPL/R/RTD/01	CCTPL, Noida	17/01/2021

Environment Condition	Temperature	Humidity	Reference Standard	Calibration Procedure No.
	25±5°C	55±10%RH	DKD-R-5-7	CCTPL/TH/CP/05S


Calibration Result

Sr. No.	Nominal Temperature (°C)	Standard Reading (°C)	DUC* Reading (°C)	Deviation (°C)	Expanded Uncertainty ±(°C)
1	37	37.125	37	-0.125	1.3

Note : DUC* : Device Under Calibration

Uncertainty of measurement at approx 95% Confidence Level with coverage factor k = 2.



Calibrated By : 
Priyank Sharma
Calibration Engineer

Approved By : 
Sandeep Sharma
Director

- Note.**
- 1.This Certificate only to the particular instrument submitted for calibration.
 - 2.This Calibration Result reported in this certificate valid at the time of and under the stated condition of measurement.
 - 3.This particular certificate can not be reproduce except in full, without prior permission of CCTPL.

--- End of Certificate ---

Page 1 of 1

