


CALIBRATION CERTIFICATE

Customer Name	Pace Hospitals		
Department	Laboratory		
Address	Beside Avasa Hotel, Near Hitech City Metro Station, Hyderabad		
	Date of Calibration	01/Apr/2021	
Instrument	BC 5140	Next Calibration Date	30/Sep/2021
Serial No	TP-99000471	Calibrator Lot Number	PLUS0321
	Calibrator Expiry date	2021-04-05	

	Parameter	Target	Range	Mean of Measured Values	New Factor	Old Factor
Calibration Values	WBC x 10 ⁹ /L	8.93	8.73 - 9.13	8.93	100	99.8
	RBC x 10 ¹² /L	4.49	4.41 - 4.57	4.49	100	99.9
	HGB g/L	13.5	13.3 - 13.7	13.5	100	100
	MCV fl	92.0	90.0 - 94.0	92.0	100	100.1
	PLT x 10 ⁹ /L	260	248 - 272	260	100	97.6

This is to certify that the BC5140..... *H E M - 5 Part*Analyzer Sr. No: TP-99000471..... *9-Jan-2020*installed at Pace Hospitals, Hyderabad has been calibrated using the *CBC-SD CAL* standard reference material..... *C-BC-CAL-PLUS* Calibrator. Traceability of..... *R.S.D. shetye - CBC* Calibrator material from the manufacturer is attached along with this certificate.

Bharath Bhattula



Engineer- Customer Service



AGAPPE DIAGNOSTICS LTD.

ISO 9001:2015 | EN ISO 13485:2016 CERTIFIED COMPANY | CIN : U24239MH1998PLC115413

CORPORATE OFFICE / REAGENT PLANT
 Agappe Hills, Pattimattom (PO), Dist. Ernakulam, Kerala - 683 562, India.
 Tel: +91 484 286 7000 | Email: agappe@agappe.in

EQUIPMENT PLANT
 X/588-CB, Block No. 32, KJNFRA Small Industrial Park,
 Nellad, Cochin, Kerala, India - 686 721. Tel: +91 484 276 7477.

KOLKATA OFFICE
 406, Merlin Matrix, Plot No-10, Block-DN, Sector V,
 Salt Lake City, Kolkata - 700 091.
 Tel: +91 33-4003 0451 | Email: kolkattaoffice@agappe.in

MUMBAI [REGISTERED OFFICE]
 401 & 402, 4th Floor, Jaisingh Business Centre, 110

DELHI OFFICE

BANGALORE OFFICE