


|  |  |  |
|--|--|--|
| <b>MEASURE TECHNO LAB</b><br><b>2, B.T. ROAD (JAYANTI CINEMA COMPLEX)</b><br><b>BARRACKPORE, KOLKATA - 700120, W.B.</b>            | Form No. - MTL/22/2006                                 | <br><small>NABL ACCREDITED LABORATORY<br/>Certificate No. CC - 2545</small> |
|  | <b>CALIBRATION CERTIFICATE OF<br/>CENTRIFUGE</b>       |  |
| Phone : 033 - 2215 - 0032, 2215 - 9687, 8100875519,<br>Mobile: 9831190974 , LAB:- 8100143376,<br>E-mail: measuretechno@yahoo.co.in | CALIBRATION CERTIFICATE NO.:MTL / CNCI / R01 / 06 - 21 | ULR - CC254521000009808F<br>Page: 1 of 1   |

1.0 Service Request No.: MTL / 08A / 06 / 21 - 22

1.1 Issued to: M/s. Chittaranjan National Cancer Institute,  
Street No. 299, DJ Block (Newtown),  
Action Area - I, Newtown, West Bengal - 700156.

|  |                         |                     |                |          |
|--|-------------------------|---------------------|----------------|----------|
| 1.2 Description & Identification of item to be Calibrated: | a) Name:                | Centrifuge          | b) Code No.:   | N.S.     |
|  | c) Sl. No.:             | ZECN - 10352        | d) Make:       | Remi     |
|  | e) Model / Type:        | R - 8C              | f) Range:      | 5250 rpm |
|  | g) Sensor:              | N.A.                | h) Resolution: | 10 rpm   |
|  | i) End User:            | Biochemistry        | j) Accuracy:   | N.S.     |
|  | k) Calibration done at: | √On Site / In House |                |          |

|                               |          |  |          |
|-------------------------------|----------|--|----------|
| 1.3 Date of receipt of item : | 08-06-21 | 1.4 Physical Condition of DUC :            | OK       |
| 1.5 Date of calibration :     | 08-06-21 | 1.6 Recommended date of next calibration : | 08-06-22 |
| 1.7 Date of Issue :           | 10-06-21 |  |          |

1.8 Environmental Conditions During Calibration: **Temperature:** 25 °C ± 3 °C  
**Humidity:** 50 % RH ± 10 % RH

1.9 Method of Calibration: **SOP / RPM / CF / 01**

2.0 Traceability :

- a) Standards used for calibration are traceable to National standards through NABL Accredited Laboratory.  
 b) The following standards / Equipment have been used.  
 i) Digital Tachometer Cal. Certificate No. JRPM - CCTR - A&S - 2020 - 0027 (JRPM, Chennai) (Cal. Date: 11/07/20, Due Date: 10/07/21)

2.1 Result :

Mechanical Calibration

| Sl. No. | Parameter/ Range | Nominal Value on DUC rpm | Std. Value Average of Five Readings rpm | Error rpm | Measurement Expanded Uncertainty ± rpm |
|---------|------------------|--------------------------|---|-----------|--|
| 1.      |                  | 1000                     | 1001.2                                  | -1.2      | 1.28                                   |
| 2.      | Rotation         | 2000                     | 2001.5                                  | -1.5      |  |
| 3.      | 5250 rpm         | 3000                     | 3002.8                                  | -2.8      |  |
| 4.      |                  | 4000                     | 4003.1                                  | -3.1      |  |
| 5.      |                  | 5000                     | 5004.9                                  | -4.9      |  |

- Remarks: i) This result has an expanded uncertainty with a coverage factor k=2 at approximately 95% confidence level.  
 ii) The calibration certificate issued for this instrument is to be used for scientific or industrial purposes only.  
 iii) Error = DUC Reading - Standard Reading.


DUC - Device Under Calibration  
Std. - Standard

N.S. - Not Specified  
rpm - Rotation Per Minute

N.A. - Not Applicable

| Opinions and Interpretations |                            |
|------------------------------|----------------------------|
| Calibrated                   | √ Accepted / Valid for use |
| Limited Use                  | Rejected / Out of use      |

Testing Engineer  
 Calibrated by:  
**Measure Techno Lab**  
 DK Singh  
 Kolkata  
 Testing Engineer  
 10/06/21

Checked/Approved by:  
  
**Quality & Technical Manager**  
 S. Pandey