


MEASURE TECHNO LAB 2, B.T. ROAD (JAYANTI CINEMA COMPLEX) BARRACKPORE, KOLKATA - 700120, W.B.	Form No. - MTL/22/2006	 NABL ACCREDITED LABORATORY Certificate No. CC - 2343
	CALIBRATION CERTIFICATE OF MICRO PIPETTE	
Phone : 033 - 2215 - 0032, 2215 - 9687, 8100875519, Mobile: 9831190974 , LAB:- 8100143376, E-mail: measuretechno@yahoo.co.in	CALIBRATION CERTIFICATE NO.: MTL / CNCI / R05 / 06 - 21	
ULR - CC254521000009812F		Page: 1 of 1

1.0 Service Request No.: MTL / 08A / 06 / 21 - 22

1.1 Issued to: M/s. Chittaranjan National Cancer Institute,
Street No. 299, DJ Block (Newtown),
Action Area - I, Newtown, West Bengal - 700156.

1.2 Description & Identification of item to be Calibrated:

a) Name:	Micro Pipette	b) Code No.:	MP - 02
c) Sl. No.:	N.S.	d) Make:	Micropet
e) Model / Type:	N.S.	f) Range:	5 µl to 50 µl
g) Sensor:	N.A.	h) Resolution:	1 µl
i) End User:	Biochemistry	j) Accuracy:	N.S.
k) Calibration done at:	On Site / √ In House		

1.3 Date of receipt of item : 08-06-21

1.4 Physical Condition of DUC : OK

1.5 Date of calibration : 08-06-21

1.6 Recommended date of next calibration : 08-06-22

1.7 Date of Issue : 10-06-21

1.8 Environmental Conditions During Calibration:

Temperature:	23 °C ± 2 °C
Humidity:	50 % RH ± 10 % RH
Pressure:	1004.8 mbar

1.9 Method of Calibration: SOP / MASS / 02 (As Per ISO : 8655 - 6 : 2002)

2.0 Traceability :

a) Standards used for calibration are traceable to National standards through NABL Accredited Laboratory.

b) The following standards / Equipment have been used.

- i) 1 mg to 200 g Weight Box (E1) Cal. Certificate No. WI / May / 19 / 010 (WEIGH INDIA, New Delhi) (Cal. Date: 17/05/19, Due Date: 17/05/22)
- ii) Data Acquisition Switch Unit Cal. Certificate No. JRPM - CCTR - ET - 2020 - 0795 (JRPM, Chennai) (Cal. Date: 01/09/20, Due Date: 31/08/21)
- iii) RTD (PT - 100) Cal. Certificate No. TL / 020 / 891.2.1 (TEMPESENS, Udaipur) (Cal. Date: 27/10/20, Due Date: 26/10/21)

2.1 Result :

Mechanical Calibration

Sl. No.	Parameter/ Range	Nominal Value µl	Mass of Water mg	Volume of Water at 20 °C µl	Error µl	Measurement Expanded Uncertainty ± µl
1.	Volume	5	4.9738	4.9937	-0.0063	0.039
2.	5 µl to 50 µl	25	24.8876	24.9872	-0.0128	0.39
3.		50	49.7857	49.9848	-0.0152	0.39

Remarks: i) Cubical Expansion co-efficient of pipette material taken as $10^{-5} \mu\text{l} / ^\circ\text{K}$.

ii) This result has an expanded uncertainty with a coverage factor $k=2$ at approximately 95% confidence level.

iii) The calibration certificate issued for this instrument is to be used for scientific or industrial purposes only.

DUC - Device Under Calibration

N.S. - Not Specified

N.A. - Not Applicable

Opinions and Interpretations			
Calibrated	√	Accepted / Valid for use	
Limited Use		Rejected / Out of use	

Calibrated by:
Testing Engineer
S. Malapatra
Testing Engineer
Kolkata
 10/06/21

Checked / Approved by:
KOLKATA
10/06/21
Quality & Technical Manager
S. Pandey