



Status



Samples



Results



QC



Reagents



Diagnostics



Options



✓-Docs



Conditions



DIAGNOSTICS - Luminometer Performance Tests

IRS Aim: 1993276

LRS Aim: 2002676

Reference Photon Values

Reference Analog Values

	Reference Photon Values		IRS		LRS		Reference Analog Values		LRS	
	IRS	LRS	Dark	%Bias	%Bias	IRS	LRS	Dark	%Bias	
1	1983039	2026856	8	0.51	1.21	1972198	2037000	42000	1.71	
2	1981530	2004964	8	0.59	0.11	1970966	2021000	48000	0.91	
3	1985317	1987810	4	0.40	0.74	1982699	2005000	48000	0.12	
4	1983480	1975518	16	0.49	1.36	1965719	2005000	42000	0.12	
5	1986359	1963784	12	0.35	1.94	1983536	1988000	44000	0.73	
6	1985744	1957254	8	0.38	2.27	1970407	1986000	40000	0.83	
7	1984815	1954766	8	0.42	2.39	1967364	1985000	50000	0.88	
8	1989483	1949094	4	0.19	2.68	1984479	1981000	48000	1.08	
9	1988522	1945738	12	0.24	2.84	1971242	1975000	52000	1.38	
10	1987249	1944404	4	0.30	2.91	1976231	1969000	52000	1.68	
11	1988252	1942134	4	0.25	3.02	1970244	1976000	44000	1.33	

Touch 'Start' to begin test.



Return



Start



Stop



Print



Help