

3 QC result

Material Used : UC control

UC Control -L		UC Control -L Assay value		Status
Parameter	Result	Upper limit	Lower Limit	
Urobilinogen(URO)	Normal	Normal		Pass
Blood(Hemoglobin)	-			Pass
Bilirubin(BIL)	-			Pass
Ketones*(KET)	-			Pass
Glucose(GLU)	-			Pass
Protein(PRO)	-			Pass
pH(pH)	5.0	5.0	6.0	Pass
Nitrite(NIT)	-			Fail
Leukocytes(LEU)	-			Pass
Creatinine(CRE)	NA	10.0	50.0	Pass
Albumin(ALB)	NA			Fail
Specific Gravity(S.G)	1.0			Fail

**RESULT REVIEW** CIHRSR

15/09/20 17:12 Ver.1 PAPER FEED MENU

ERROR:00BB 121

RACK:E003-01 ID:\*\*\*\*\*

C.00000001 Strip 9A

09/09/20 18:48

URO normal mg/dL

BLD - mg/dL

BIL - mg/dL

KET - mg/dL

GLU - mg/dL

PRO - mg/dL

pH 5.0

NIT -

LEU - c/uL

S.G 1.009

COLOR L YELLOW

CLOUD -

CLEAR

▲ ▲ ▲ ▼ ▼

EDIT ID SEARCH SELECTION START OUTPUT VIEW NORMAL