

5. BACKGROUND CHECK

PARAMETER	RESULT	Range
WBC-N	0.0	$0.1 \times 10^3/\mu\text{L}$ or less
WBC-D	0.0	$0.1 \times 10^3/\mu\text{L}$ or less
RBC	0.0	$0.02 \times 10^6/\mu\text{L}$ or less
HGB	0.0	0.1g/dl or less
PLT	0	$10 \times 10^6/\mu\text{L}$ or less

Technical Service Department
Transasia Bio-Medicals Ltd